



An impedimetric biosensor based on electrophoretically assembled ZnO nanorods and carboxylated graphene nanoflakes on an indium tin oxide electrode for detection of the DNA of *Escherichia coli* O157:H7

Nandita Jaiswal¹ · Chandra Mouli Pandey² · Shipra Solanki³ · Ida Tiwari¹ · Bansi Dhar Malhotra³

Received: 23 May 2019 / Accepted: 11 October 2019
© Springer-Verlag GmbH Austria, part of Springer Nature 2019

Abstract

Aminopropyltrimethoxysilane (APTMS)-functionalized zinc oxide (ZnO) nanorods and carboxylated graphene nanoflakes (c-GNF) were used in a composite that was electrophoretically deposited on an indium tin oxide (ITO) coated glass substrate. The modified ITO electrodes were characterized using various microscopic and spectroscopic techniques which confirm the deposition of the APTMS-ZnO/c-GNF composite. The electrodes have been used for the covalent immobilization of an *Escherichia coli* O157:H7 (*E. coli*)-specific DNA prob. Impedimetric studies revealed that the gene sensor displays linear response in a wide range of target DNA concentration (10^{-16} M to 10^{-6} M) with a detection limit of 0.1 fM. The studies on the cross-reactivity to other water-borne pathogens show that the bioelectrode is highly specific.

Keywords Biosensor · *Escherichia coli* · Zinc oxide · Graphene nanoflakes · Electrophoretic deposition

Introduction

Nanostructured materials have gathered interest as functional materials in various aspects of designing biomedical devices [1–5]. Nanostructured metal oxides (NMOs) have been widely used in biosensing applications due to their various characteristics, such as large surface to volume ratio, high electrocatalytic activity, their small band gap that enhances their electron conductivity and the adsorption ability [6] which helps in

increasing the loading of biomolecules per unit mass of particles. The employment of NMOs may open the possibility of new signal transduction due to their sub-micrometer dimension and can be useful for in vivo analysis. The NMOs reported to have their potential application in the field of biosensor are of zinc, iron, cerium, tin, zirconium, titanium and magnesium [7]. Among these zinc oxide (ZnO), has been widely used for the development of various sensing platforms. ZnO has gained much interest due to its novel characteristics such as biocompatibility, catalytic efficiency, non-toxicity, as well as chemical and electrical stability [8–11]. ZnO shows various nanostructures of different morphologies such as nanowires, nanoflakes, nanorods, nanobelts, nanoneedles and nanocombs [12–16].

Though NMOs have shown excellent properties as materials for biosensing fabrication, but there are some limitations such as low conductivity as compared to carbon materials, low surface area and high resistance. So, it will be logical to develop a composite material containing both carbon and transition metal oxide as the electrode material, combining the merits and mitigating the shortcomings of both the individual components [17, 18]. The use of graphene in nanostructured carbon-metal oxides composites not only provides physical support to metal oxides but also serve as channels for charge transport due to its high surface area and the robust conductive

Nandita Jaiswal and Chandra Mouli Pandey contributed equally to this work.

Electronic supplementary material The online version of this article (<https://doi.org/10.1007/s00604-019-3921-8>) contains supplementary material, which is available to authorized users.

✉ Ida Tiwari
idatiwari@bhu.ac.in

¹ Department of Chemistry(Centre of Advanced Study), Institute of Science, Banaras Hindu University, Varanasi 221005, India

² Department of Applied Chemistry, Delhi Technological University, Delhi 110042, India

³ Department of Biotechnology, Delhi Technological University, Delhi 110042, India

structure. In graphene-metal oxide composite, the graphene provides the benefit of high electron conductivity while electro-activities of metal oxides contribute to high specific capacitance. The performance of biosensors based on such materials is governed by the interface between the nanostructured composite material and biomolecule [19]. The nano bio-interface formation and its properties depend upon nature of the nanostructured material and also other parameters related to effective surface area, roughness, porosity, conductance, functional groups etc. The biomolecules are said to bind the metal nanocomposites generally by physical adsorption or chemical binding [20] and the activity of the biomolecule can be retained by establishing a biocompatible microenvironment on the biointerface [21, 22]. Thus the use of nanostructured composite material would decide the sensitivity, selectivity, detection limit, cost and shelf-life of the biosensor.

The enter-hemorrhagic stereotype of bacterium *Escherichia coli* is *E. coli* O157:H7 and its infection can lead to kidney failure and hemorrhagic diarrhea [23]. Various methods such as culture and colony counting are available for the detection of the pathogen. The traditional methods require an excess of time and the results are also ambiguous [24]. Some methods involve enrichment of culture by growing it for a specific time and then separated and identified using polymerase chain reaction or immunological methods of identifying genes. These methods are quite complicated and costly in terms of requirement of labeled antibodies and their pre-separation or pre-enrichment. Thus, there is an urgent requirement for a rapid, sensitive and cost-effective technique to overcome these problems. An alternative technique to the methods mentioned above is the use of biosensors for pathogenic bacterial genome detection [25, 26].

The biosensors based on nanostructured composite materials for nucleic acid detection have aroused much more interest due to their efficiency of converting the hybridization event into an analytical signal. Wang et al. reported a biosensor based on ZnO nanowires, multiwalled carbon nanotubes (MWCNT) and gold nanoparticles composite utilizing thiolated single-stranded DNA probe to immobilize it on the gold nanoparticle surface. The detection limit of this DNA sensor was found to be 0.35 pM using $[\text{Ru}(\text{NH}_3)_6]^{3+}$ as a redox indicator [27]. Sol-gel derived nanostructured ZrO_2 film for the detection of 16 s rRNA coding region for *E. coli* has been reported with a sensitive determination in the range of 10^{-6} to 10^6 pM of complementary DNA [28]. Yang et al. reported a composite based on nano ZrO_2 , Chitosan and MWCNT for the development of DNA sensor on glassy carbon electrode. The composite provided good charge transport characteristics with high surface area favoring the increase in loading capacity and enhanced selectivity with a detection limit of 75 pM [29]. For point of care analysis, an ideal detection method should be such that it may detect the disease directly from food and water samples with minimal

requirement of sample preparation. However, the drawbacks with nucleic acid based biosensor are that the extraction of DNA from culture samples and processing involve laborious pre-treatment steps. Thus efforts are being made to develop affinity based biosensors (using antibody-antigen interactions) and aptasensors to allow the detection of *E. coli* O157:H7 cells directly from food and water samples but till date they are not very applicable and have not been fully developed into technology. Hence, it becomes imperative to design such materials which can overcome these drawbacks.

The present work reports studies conducted to exploit the properties of ZnO nanorods by combining it with a new class of graphene material known as carboxylated graphene nanoflakes (c-GNF) [30]. The ZnO nanorods were firstly modified with aminopropyltrimethoxysilane (APTMS) to introduce NH_2 groups on the surface of ZnO nanorods. Then the electrostatic interaction of this APTMS functionalized ZnO nanorods (APTMS-ZnO) with the carboxylic acid groups on c-GNF was done for the formation of APTMS functionalized ZnO nanorods-carboxylated graphene nanoflakes composite (APTMS-ZnO/c-GNF) composite. Further, the composite was electrophoretically deposited using magnesium (II) electrolyte which provided overall positive charge to the colloid suspension for the fabrication of impedimetric biosensor for *E. coli* DNA detection.

Chemicals and materials required

Zinc(II) nitrate hexahydrate (www.alfa.com), aminopropyltrimethoxysilane (www.sigmaaldrich.com), glutaraldehyde (www.spectrochem.co.in) and sodium hydroxide pellets (www.sdfine.com). The specific sequence of *E. coli* probe (17 mer.) has been identified from the 16 s rRNA coding region of the *E. coli* genome. The complementary, non-complementary and one-base mismatch target sequences were procured from Sigma Aldrich (www.sigmaaldrich.com). Sequences of oligonucleotide probes used are:

Probe I- probe DNA (pDNA): amino-5'-GGT CCG CTT GCT CTC GC-3'.

Probe II- complementary DNA (cDNA): 5'-GCG AGA GCA AGC GGA CC-3'.

Probe III- non-complementary DNA: 5'-CTA GTC GTA TAG TAG GC-3'.

Probe IV- one-base mismatch DNA: 5'-GCG AGA GAA AGC GGA CC-3'.

The solution of oligonucleotide was made in Tris-EDTA buffer (1 M Tris-HCl, 0.5 M EDTA, pH 8.0) (www.merck.com) and was stored at -20°C prior to use.

Apparatus and measurements

X-ray diffraction (XRD) analysis was performed using D8 Advance/Discover Bruker, Germany, Diffractometer with Cu K α . The UV-vis spectra was recorded using Hitachi U-3900 spectrophotometer while the transmission electron microscopy (TEM) and energy dispersive X-ray analysis (EDX) were carried out using TECNAI 12 G2FEI microscope, operating at 120 kV instrument. FT-IR spectroscopic study was done using PerkinElmer spectrometer (Spectrum two, UK), the scanning electron microscopy (SEM) images were recorded using a JEOLJSM-6700F field-emitting scanning electron microscope (FESEM, 15 kV) while the contact angle was measured using contact angle meter (Data physics OCA15EC). Electrochemical characterization were carried out on an Autolab potentiostat/galvanostat (www.metrohm-autolab.com) using a three-electrode system with ITO as working electrode, Ag/AgCl as a reference electrode, and platinum foil as a counter electrode in phosphate buffer solution (PBS of 100 mM concentration with pH 7.0 in 0.9% NaCl).

Experimental section

Synthesis of APTMS functionalized ZnO nanorods

The synthesis of ZnO nanorods are discussed in section 1 of supp. Material [31]. For functionalization of the ZnO nanorods, APTMS was used. For amino functionalization of ZnO, 0.2 g ZnO powder was ultrasonicated in 20 mL toluene followed by addition of 400 μ L of APTMS under nitrogen atmosphere in a round bottom flask and refluxed at 110 °C for 6 h to get APTMS functionalized ZnO nanorods (APTMS-ZnO) [32, 33].

Synthesis of APTMS functionalized ZnO nanorods-carboxylated graphene nanoflakes composite

The c-GNF were synthesized as reported in the literature [30]. A dispersion of c-GNF was made using DI water (0.5 mgmL⁻¹) via ultrasonication for 2 h. Further 10 mg of APTMS-ZnO nanorods were added to the above dispersion and ultrasonicated for 2 h followed by subsequent stirring (4 h). Then the suspension was centrifuged at 10,000 rpm for 1 h to obtain APTMS-ZnO nanorods-c-GNF composite (APTMS-ZnO/c-GNF). Finally, the obtained material was washed with distilled water and then dried at 70 °C under vacuum (Scheme 1a).

Electrophoretic deposition of APTMS-ZnO/c-GNF

Prior to deposition, the ITO coated glass electrode were hydrolyzed by immersing in H₂O₂/NH₄OH/H₂O solution (1:1:5, v/v) for about 30 min at 80 °C. For the deposition, APTMS-ZnO/c-GNF (5 mg) were dispersed in 10 mL acetonitrile. The electrophoretic deposition was carried using the DC source by applying a voltage of 120 V for 2 min. 10⁻⁴ M of Mg(NO₃)₂·6H₂O was added to the colloidal suspension (APTMS-ZnO/c-GNF in acetonitrile) acting as an electrolyte (for creating a charge on the colloids).


Immobilization and hybridization of probe DNA on APTMS-ZnO/c-GNF/ITO electrode

The coupling of amine terminal of probe DNA (pDNA) and APTMS-ZnO/c-GNF/ITO electrode was done by immobilizing 20 μ L of pDNA on the modified electrode using glutaraldehyde as cross linker (0.1%, v/v; 4 h). The fabricated pDNA/APTMS-ZnO/c-GNF/ITO bioelectrode was employed for the detection of *E. coli* complementary DNA (cDNA). The different concentrations of pDNA and cDNA were prepared in Tris-EDTA buffer (1 M Tris-HCl, 0.5 M EDTA) of pH 8.0 from the stock solution of 1 $\times$ 10⁻⁶ M concentration. The pDNA/APTMS-ZnO/c-GNF/ITO was subjected to different concentrations of cDNA for 30 min and the corresponding Rct value was measured using electrochemical impedance spectroscopy. The different steps involved in the fabrication of APTMS-ZnO/c-GNF/ITO and pDNA/APTMS-ZnO/c-GNF/ITO has been depicted in scheme 1b.

Extraction and pretreatment of DNA from bacterial clinical samples

The extraction of DNA was carried out using the panel of strains comprising *E. coli*, *Salmonella Typhimurium*, *Neisseria meningitides* and *Klebsiella pneumonia*. The DNA was extracted by suspending the bacterial colonies and vortexing and pouring them into 100 mL of MilliQ water. The suspension was further boiled for 10 min and centrifuged at 10,000 rpm for 5 min and equal volume (100 mL) of 24:1 (v/v) chloroform-iso-amyl alcohol was added to the boiled and separated suspension. This solution was again centrifuged at 12,000 rpm for 10 min, which resulted in the formation of an aqueous layer containing DNA. This aqueous layer was carefully separated out and kept at -20 °C prior to use. For denatured single-stranded DNA, the DNA samples extracted from the above-mentioned method were heated on water bath (95 °C) for 5 min and were immediately chilled in ice. These aliquots of samples were subjected to sonication for 15 min at 120 V to break the long DNA strands into smaller fragments.

Scheme 1 Schematic representation of (a) synthesis of APTMS-ZnO/c-GNF and (b) subsequent fabrication of nucleic acid biosensor for *E. coli* DNA detection using APTMS-ZnO/c-GNF/ITO electrode.



Results and discussion

Morphological characterization of the deposited materials

The morphology of the electrodeposited APTMS-ZnO/c-GNF on ITO electrode was studied using SEM (Fig. 1). The SEM image of APTMS-ZnO/c-GNF/ITO shows dendritic structure that might be due to the assembly of APTMS-ZnO nanorods with c-GNF on applying potential during the electrodeposition (Fig. 1a and b at different magnification). Even the rods of the ZnO have not lost its morphology and dendritic structure (Fig. 1c) would lead to a greater surface area of the APTMS-ZnO/c-GNF/ITO which will provide an excellent

platform for pDNA attachment. While in the case of electrodeposition of APTMS-ZnO no such assembly is observed and smoothsheet-like morphology can be seen for APTMS-ZnO/ITO (Fig. 1d).

Electrochemical characterization

Electrochemical impedance spectroscopy

The electrochemical properties of the deposited material are evaluated using electrochemical impedance spectroscopic (EIS) studies since it is a powerful tool for observing the charge transfer processes at the electrode-electrolyte interface [25]. The impedimetric spectra are seen in Fig. 2a called as the




Fig. 1 (a), (b), (c) SEM images of APTMS-ZnO/c-GNF/ITO electrode at different magnifications and (d) SEM image of APTMS-ZnO/ITO electrode

Nyquist plot which consists of two parts one semicircular part at high frequency and the other linear plot at a lower frequency. The semicircular high-frequency part signifies the electron transfer mediated process. The linear lower frequency part corresponds to the diffusion-limited process. The impedimetric spectrum of the fabricated biosensor was fitted using a modified Randles equivalent circuit. This circuit is made up of different elements such as electrolyte resistance (R_s), Warburg impedance (Z_W), double layer capacitance (CPE) and the electron transfer resistance (R_{ct}). The R_s is the resistance which is observed between the working and the reference electrode. Z_W is the element associated with the process at the electrode-electrolyte interface due to diffusion of redox probe ions. CPE is related to the double layer capacitance. The numerical value of R_{ct} is helpful in deriving the electrochemical response of the fabricated biosensor. The measurements of the impedance were made to carry out at the open circuit voltage and the test frequency was in the range of 100 kHz to 0.1 Hz with amplitude of ± 5 mV [25].

The EIS technique was adopted to investigate the electrochemical behavior of APTMS-ZnO/c-GNF/ITO electrode, pDNA/APTMS-ZnO/c-GNF/ITO electrode and pDNA/APTMS-ZnO/c-GNF/ITO electrode incubated with *E. coli* cDNA by comparing their R_{ct} values (Fig. 2a). The semicircular region of the impedance curve reflects the numerical value of R_{ct} , derived by fitting a Randles and Ershler model circuit. Also, the presence of a functional group on the composite material APTMS-ZnO/c-GNF might chemisorb the complementary target of the oligonucleotide. To avoid the false positive signal due to chemisorption, there is a need to optimize the concentration of pDNA. The optimized concentration of pDNA was found to be 10^{-6} M, which

showed the leveling of the R_{ct} value (optimization data not shown).

Cyclic voltammetry

The cyclic voltammetry studies were used to investigate the electrocatalytic behavior of the fabricated electrodes towards the redox behavior of $[\text{Fe}(\text{CN})_6]^{3-/4-}$ [25]. The outcome of the cyclic voltammetry as the function of scan rate (ν) for APTMS-ZnO/c-GNF/ITO and pDNA/APTMS-ZnO/c-GNF/ITO can be seen in Fig. 2b and c. The anodic and cathodic peak corresponds to oxidation and reduction of $[\text{Fe}(\text{CN})_6]^{3-/4-}$ redox couple respectively. The anodic and cathodic peak currents were found to vary linearly with ν in the range of 10 to 300 mVs^{-1} indicating the electrochemical reaction of the redox pair at the working electrode to be diffusion-limited (Fig. S7 A and B).

Electrochemical response study

The electrochemical response of the DNA hybridization was studied using the EIS technique [34]. Upon immobilization of pDNA on APTMS-ZnO/c-GNF/ITO an increase in the R_{ct} value (754 Ω) was observed because of the repulsion between the $[\text{Fe}(\text{CN})_6]^{3-/4-}$ electrolytic redox couple and the phosphate backbone of the pDNA (Fig. 3a curve (i)). There was a further increase in R_{ct} value after the hybridization of complementary DNA (889 Ω) due to an increase in the number of phosphate backbone in the double-stranded DNA (Fig. 3a curve (ii-xii)). It can be inferred that with a change in the negative charge and conformation transition, the interfacial properties changes and this becomes the basis of sensitive determination of cDNA of *E. coli*. The signal used for the measurement was the difference in the R_{ct} value of pDNA immobilized electrode and the electrode after hybridization with cDNA $\{\Delta R_{ct} = R_{ct}(\text{cDNA}) - R_{ct}(\text{pDNA})\}$. There was a linear relationship between the analytical signal (ΔR_{ct}) and the log of concentration of cDNA target ranging from 10^{-16} to 10^{-6} M (Fig. 3b) and follows the following regression eq. (1):

$$\Delta R_{ct} = 3840.02 + 235.14 [\log (\text{concentration of cDNA})] \quad (1)$$

with linear regression of 0.9952. The detection limit of the fabricated pDNA/APTMS-ZnO/c-GNF/ITO based biosensor is 0.1 fM. As the APTMS-ZnO/c-GNF contain ample functional groups which provide a conducive platform for the immobilization of pDNA, leading to higher loading of pDNA on to the APTMS-ZnO/c-GNF/ITO electrode, resulting in increased hybridization efficiency. From the data obtained it can be easily concluded that compared with other previously




Fig. 2 (a) Nyquist diagram (Z_i vs. Z_r) for the Faradic impedance in the frequency range from 100 kHz to 0.1 Hz for (i) APTMS-ZnO/c-GNF/ITO electrode, (ii) pDNA/APTMS-ZnO/c-GNF/ITO bioelectrode, and (iii) cDNA on pDNA/APTMS-ZnO/c-GNF/ITO bioelectrode., (b) Cyclic

voltammogram of APTMS-ZnO/c-GNF/ITO electrode and (c) pDNA/APTMS-ZnO/c-GNF/ITO electrode (in 5 mM $[\text{Fe}(\text{CN})_6]^{3-/4-}$, PBS, 100 mM, pH 7.0, 0.9% NaCl)

reported methods which have their inherent limitations as discussed in preceding section 1 and other works reported based on hybridization this genosensor shows a wider detection range, better sensitivity and low detection limit (cf. Table S3 and S4).

Selectivity study


The pDNA/APTMS-ZnO/c-GNF/ITO bioelectrode was found to be selective towards different target DNA sequences such as non-complementary, one base mismatch and complementary sequences. The selectivity was studied using EIS technique (Fig. S9). There was only slight increase in the R_{ct} value after incubating the bioelectrode with non-complementary DNA sequence (844 Ω) (Fig. S9 curve (ii)) as compared to pDNA (curve (i)). This is very obvious since the non-complementary DNA base sequences do not match with the pDNA base sequence and no hybridization would lead to a negligible change in the charge distribution of electrode surface and therefore very less change in R_{ct} value. When the pDNA/APTMS-ZnO/c-GNF/ITO bioelectrode

was incubated with one-base mismatch DNA sequence (curve iii), there was a slight increase in the R_{ct} value in comparison to pDNA/APTMS-ZnO/c-GNF/ITO bioelectrode which can be clearly seen by the semicircular diameter of the curve (i) and curve (iii) of Fig. S9. This might be due to partial hybridization of pDNA with bases of one-base mismatch DNA as a result of which there is increase in negative charge at the electrode surface. The negatively charged electrode surface now applied a blocking effect i.e. repulsion towards the redox ions. While on exposure of pDNA/APTMS-ZnO/c-GNF/ITO bioelectrode with cDNA there is complete hybridization of bases takes place leading to increase in negative charge and higher blocking effect towards the redox ions (curve iv) as can be ascertained from high R_{ct} value.

Specificity study

The specificity of pDNA/APTMS-ZnO/c-GNF/ITO bioelectrode was checked towards the isolated DNA from the panel strains of *E. coli* and various other water-borne pathogens such as *Salmonella Typhimurium*, *Neisseria*

Fig. 3 (a) EIS data for detection of cDNA hybridized on pDNA/APTMS-ZnO/c-GNF/ITO bioelectrode, (b) the calibration plot for detection of cDNA on pDNA/APTMS-ZnO/c-GNF/ITO bioelectrode (in 5 mM $[\text{Fe}(\text{CN})_6]^{3-/-4-}$, PBS, 100 mM, pH 7.0, 0.9% NaCl)



meningitides and *Klebsiella pneumonia* using EIS study (Fig. S10). There was a marked increase in R_{ct} value when the bioelectrode was incubated with clinical *E. coli* strains indicating the complete hybridization. Whereas, negligible change was recorded when the electrode was incubated with other pathogenic strains *S. Typhimurium*, *N. meningitides* and *K. pneumonia*. While the strains of other bacteria such as *S. Typhimurium* do not affect the hybridization of the bioelectrode with the strains of *E. coli* revealing the specificity of the fabricated biosensor.

The reusability of the fabricated electrode was evaluated by immersing the electrode in a buffer solution (pH 8.0) containing Tris buffer (10 mM) and EDTA (1 mM) at 100 °C for 5 min, followed by cooling in an ice bath for about 25 min. This process completely removes the cDNA via thermal denaturation. The R_{ct} value of the nucleic acid sensor was found to decrease after each regeneration process (5 cycles) with a total loss of 10.42% of the original R_{ct} value. This reduction in signal may be due to surface fouling during the regeneration process indicating that the biosensor can be used for at least 5 times (data not shown).

Conclusion

In summary, the work emphasizes the synthesis of APTMS functionalized ZnO-carboxylated graphene nanoflakes composite using wet chemical route and its electrophoretic deposition on ITO supported electrode at 120 V. This fabricated platform was immobilized with probe DNA sequences for the selective determination of *E. coli* O157:H7 DNA using impedimetric technique. The proposed *E. coli* O157:H7 DNA sensor retained its activity for 5 repetitive uses with a loss of approximately 10% activity. The lowest concentration of *E. coli* O157:H7 DNA detection using the fabricated biosensor was found to be 0.1 fM. The sensing platform can be interestingly used for the detection of other water borne pathogenic DNA sequences. Besides all this, efforts are being made

to develop affinity based biosensors (using antibody–antigen interactions) to allow the detection of *E. coli* O157:H7 cells directly from water samples.

Acknowledgements N.J. and S.S. are thankful to CSIR, India for the award of SRF and RA, respectively. C.M.P. acknowledges the Department of Science and Technology, New Delhi, India for DST-INSPIRE Faculty Award (DST/INSPIRE/04/2015/000932). We thank Dr. Gajjala Sumana (CSIR-National Physical Laboratory) for interesting discussions. N.J. and I.T. are sincerely thankful to Prof. O.N. Srivastava (Dept. of Physics, I.Sc. B.H.U) and Prof. Rajiv Prakash (IIT B.H.U) for providing TEM and SEM facilities respectively. B.D.M. thanks Science & Engineering Research Board (Govt. of India) for the award of a Distinguished Fellowship (SB/S9/YSCP/SERB-DF/2018).

Compliance with ethical standards

Conflict of interest The authors declare no conflict of interest.

References

- Pandey CM, Dewan S, Chawla S, Yadav BK, Sumana G, Malhotra BD (2016) Controlled deposition of functionalized silica-coated zinc oxide nano-assemblies at the air/water interface for blood cancer detection. *Anal Chim Acta* 937:29–38
- Zhang P, Wang S (2014) Designing fractal nanostructured biointerfaces for biomedical applications. *Chem phys chem* 15: 1550–1561
- Xu Q, Wang L, Lei J, Deng S, Ju H (2013) Platinum nanodendrite functionalized graphene nanosheets as a non-enzymatic label for electrochemical immunosensing. *J Mater Chem B* 1:5347–5352
- Wang J (2005) Nanomaterial-based electrochemical biosensors. *Analyst* 130:421–426
- Pan C, Zhu J (2009) The syntheses, properties and applications of Si, ZnO, metal, and heterojunction nanowires. *J Mater Chem* 19: 869–884
- Rahman MM, Ahammad AJS, Jin JH, Ahn SJ, Lee JJ (2010) A comprehensive review of glucose biosensors based on nanostructured metal-oxides. *Sensors* 10:4855–4886
- Solanki PR, Kaushik A, Agrawal VV, Malhotra BD (2011) nano-structured metal oxide-based biosensors. *NPG Asia Mater* 3:17–24 (2011)
- Chai G, Lupan O, Chow L, Heinrich H (2009) Crossed zinc oxide nanorods for ultraviolet radiation detection. *Sens Actuators A: Physical* 150:184–187

9. Mehrabian M, Azimirad R, Mirabbaszadeh K, Afarideh H, Davoudian M (2011) UV detecting properties of hydrothermally synthesized ZnO nanorods. *Physica E Low dimens Syst Nanostruct* 43:1141–1145
10. Ali SMU, Nur O, Willander M, Danielsson B (2010) A fast and sensitive potentiometric glucose microsensor based on glucose oxidase coated ZnO nanowires grown on a thin silver wire. *Sensors Actuators B Chem* 145:869–874
11. Fulati A, Ali SMU, Asif MH, Alvi NH, Willander M, Brännmark C, Stralfors P, Börjesson SI, Elinder F, Danielsson B (2010) An intracellular glucose biosensor based on nanoflake ZnO. *Sensors Actuators B Chem* 150:673–680
12. Mohammed AM, Ibraheem IJ, Obaid AS, Bououdina M (2017) Nanostructured ZnO-based biosensor: DNA immobilization and hybridization. *Sens Biosensing Res* 15:46–52
13. Ahmad R, Tripathy N, Jang NK, Khang G, Hahn YB (2015) Fabrication of highly sensitive uric acid biosensor based on directly grown ZnO nanosheets on electrode surface. *Sens. Actuators B Chem* 206:146–151
14. Tak M, Gupta V, Tomar M (2014) Flower-like ZnO nanostructure based electrochemical DNA biosensor for bacterial meningitis detection. *Biosens Bioelectron* 59:200–207
15. Chen X, Yan H, Shi Z, Feng Y, Li J, Lin Q, Wang X, Sun W (2017) A novel biosensor based on electro-co-deposition of sodium alginate-Fe₃O₄-graphene composite on the carbon ionic liquid electrode for the direct electrochemistry and electrocatalysis of myoglobin. *Polym Bull* 74:75–90
16. Salih E, Mekawy M, Hassan RYA, El-Sherbiny IM (2016) Synthesis, characterization and electrochemical-sensor applications of zinc oxide/graphene oxide nanocomposite. *J Nanostructure Chem* 6:137–144
17. Tiwari JN, Tiwari RN, Kim KS (2012) Zero-dimensional, one-dimensional, two-dimensional and three-dimensional nanostructured materials for advanced electrochemical energy devices. *Prog Mater Sci* 57:724–803
18. Pokropivny VV, Skorokhod VV (2007) Classification of nanostructures by dimensionality and concept of surface forms engineering in nanomaterial science. *Mater Sci Eng C* 27:990–993
19. Nel AE, Mädler L, Velegol D, Xia T, Hoek EMV, Somasundaran P, Klaessig F, Castranova V, Thompson M (2009) Understanding biophysicochemical interactions at the nano–bio interface. *Nat Mater* 8:543–557
20. Loo AH, Bonanni A, Pumera M (2013) Thrombin aptasensing with inherently electroactive graphene oxide nanoplatelets as labels. *Nanoscale* 5:4758–4762
21. Chang H, Wang Y, Li J (2011) Electrochemical DNA sensors: from nanoconstruction to biosensing. *Curr Org Chem* 15:506–517
22. Chen A, Chatterjee S (2013) Nanomaterials based electrochemical sensors for biomedical applications. *Chem Soc Rev* 42:5425–5438
23. Scallan E, Hoekstra RM, Angulo FJ, Tauxe RV, Widdowson MA, Roy SL, Jones JL, Griffin PM (2011) Foodborne illness acquired in the United States-major pathogens. *Emerg Infect Dis* 17:7–15
24. Tiwari I, Singh M, Pandey CM, Sumana G (2015) Electrochemical detection of a pathogenic *Escherichia coli* specific DNA sequence based on a graphene oxide–chitosan composite decorated with nickel ferrite nanoparticles. *RSC Adv* 5:67115–67124
25. Jaiswal N, Pandey CM, Soni A, Tiwari I, Rosillo-Lopez M, Salzmänn CG, Malhotra BD, Sumana G (2018) Electrochemical genosensor based on carboxylated graphene for detection of water-borne pathogen. *Sensors Actuators B Chem* 275:312–321
26. Xu M, Wang R, Li Y (2017) Electrochemical biosensors for rapid detection of *Escherichia coli* O157:H7. *Talanta* 162:511–522
27. Wang J, Li S, Zhang Y (2010) A sensitive DNA biosensor fabricated from gold nanoparticles, carbon nanotubes, and zinc oxide nanowires on a glassy carbon electrode. *Electrochim Acta* 55:4436–4440
28. Solanki PR, Kaushik A, Chavhan PM, Maheshwari SN, Malhotra BD (2009) Nanostructured zirconium oxide based genosensor for *Escherichia coli* detection. *Electrochem Commun* 11:2272–2277
29. Yang Y, Wang Z, Yang M, Li J, Zheng F, Shen G, Yu R (2007) Electrical detection of deoxyribonucleic acid hybridization based on carbon-nanotubes/nano zirconium dioxide/chitosan-modified electrodes. *Anal Chim Acta* 584:268–274
30. Rosillo-Lopez M, Lee TJ, Bella M, Harta M, Salzmänn CG (2015) Formation and chemistry of carboxylic anhydrides at the graphene edge. *RSC Adv* 5:104198–104202
31. Wu C, Qiao X, Chen J, Wang H, Tan F, Li S (2008) A novel chemical route to prepare ZnO nanoparticles. *Mater Lett* 60:1828–1832
32. Zhu J, He J (2012) Facile synthesis of Graphene-wrapped honeycomb MnO₂ Nanospheres and their application in Supercapacitors. *Appl Mater Interfaces* 4:1770–1776
33. Ivers-Tiffée E, Weber A, Schichlein H (2010) Electrochemical impedance spectroscopy. *Handbook of Fuel Cells*, John Wiley Sons Ltd Chichester
34. Khan SB, Faisal M, Rahman MM, Jamal A (2011) Low-temperature growth of ZnO nanoparticles: Photocatalyst and acetone sensor. *Talanta* 85:943–949

Publisher's note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.