

Accepted Manuscript

Highly sensitive detection of NT-proBNP by molecular motor

Jie Zhang, Yao-Gen Shu

PII: S2352-3042(16)30051-4

DOI: [10.1016/j.gendis.2016.11.002](https://doi.org/10.1016/j.gendis.2016.11.002)

Reference: GENDIS 114

To appear in: *Genes & Diseases*

Received Date: 20 August 2016

Revised Date: 29 October 2016

Accepted Date: 11 November 2016

Please cite this article as: Zhang J, Shu YG, Highly sensitive detection of NT-proBNP by molecular motor, *Genes & Diseases* (2016), doi: 10.1016/j.gendis.2016.11.002.

This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.



Highly sensitive detection of NT-proBNP by molecular motor

Jie Zhang¹, Yao-Gen Shu^{2*}

¹Beijing Entry-Exit Inspection and Quarantine Bureau, Beijing, 100026

²Institute of Theoretical Physics, Chinese Academy of Sciences, Beijing, 100190

*To whom correspondence should be addressed. E-mail: shuyg@itp.ac.cn

Abstract

FoF₁-ATPase is an active rotary motor, and generates three-ATP for each rotation. At saturated substrate concentration, the motor can achieve about 10³ r.p.m, which means one motor can generate about 10⁵ ATP molecules during 30 minutes. Here, we constituted a novel nanodevice with a molecular rotary motor and a “battery”, FoF₁-ATPase and chromatophore, and presented a novel method of sandwich type rotary biosensor based on ϵ subunit with one target-to-one motor, in which one target corresponds 10⁵ ATP molecules as detection signals during 30 minutes. The target such as NT-proBNP detection demonstrated that this novel nanodevice has potential to be developed into an ultrasensitive biosensor to detect low expressed targets.

Keywords:

molecular motor; FoF₁-ATPase; highly sensitive detection.

1. Introduction

FoF₁-ATPase is the ubiquitous enzyme that uses the transmembrane electrochemical potential to synthesize ATP in bacteria, chloroplasts and mitochondria. The holoenzyme can be divided into two rotary motors, Fo and F₁. F₁ motor consists of a crown type “stator” ($\alpha_3\beta_3$) and a eccentric “rotor” (γ), while Fo motor consists of a “stator” (a subunit) embedded in membrane and a ring channels “rotor” (c_n). The two “stators” are fixed by b_2 and δ subunits, while the two “rotors” are mechanically coupled by ϵ subunit. The membrane embedded Fo unit converts the proton motive force (p.m.f) into mechanical rotation of the “rotor”, thereby causing cyclic conformational change of $\alpha_3\beta_3$ crown (“stator”) in F₁ and driving three ATP molecules synthesis for each rotation at nearly 100% efficiency^[1-7].

In vitro, however, FoF₁-ATPase must be reconstituted in polymersome and coupled with Bacteriochlorophyll to maintain its ATP synthesis, where the Bacteriochlorophyll converted the light energy into transmembrane p.m.f^[8]. The polymersome with Bacteriochlorophyll was named chromatophore. The combination of FoF₁-ATPase and chromatophore is a sophisticated nanomachine, in which chromatophore function as a “battery” to power the rotary motor FoF₁-ATPase, as well as the “battery” can remotely be recharged by light. Furthermore, the combination of FoF₁-ATPase and chromatophore can conveniently be prepared by the phototrophic bacterium, in which the cells were disrupted by sonication on ice. Each chromatophore vesicle of 100 nm diameter contains on average one FoF₁-ATPase^[9].

Additionally, kinetics simulation^[10] and nanoporous membrane experiment^[11] have demonstrated that the motor can achieve about 10³ r.p.m at saturated substrate concentration and the chromatophore can processively power the motor for more than one hour once the “battery” was recharged by light^[12-14], which means one motor can generate about 105 ATP molecules during 30 minutes^[15].




Figure 1: Sandwich type biosensor with one target-to-one motor. The red target, NT-proBNP has two sites, the one is employed to be fixed to the coverslip by antibody1 and avidin; while another site is used to be linked to ϵ subunit of motor by antibody2, avidin and antibody of ϵ complex. This immunoassay shows that the number of targets equals to that of motors.

On the other hand, NT-proBNP is the amino-terminal fragment of the prohormone of BNP (brain natriuretic peptide), which is secreted mainly by atrial or ventricular myocardial cells, essentially in response to ventricular wall stress. The biological actions of BNP include diuresis, vasodilatation, reduction of systemic and regional (renal and cardiac) sympathetic tone, myocardial relaxation, anti-hypertrophic and anti-fibrotic effects and cytoprotection of cardiomyocytes. In clinical practice, BNP and NT-proBNP have been introduced because of their stability in biological specimens and their utility in the management of suspected or established heart failure (diagnosis, prognosis and guided medical treatment). That is, NT-proBNP is the biomarker of cardiac impairment^[16]. However, the current sensitivity of NT-proBNP detection is 220pg/ml^[17]. There is an urge to have a high sensitive detection of NT-proBNP due to its low expression in early cardiac impairment. Here, we constituted a nanodevice with the combination of FoF₁-ATPase and chromatophore to detect the NT-proBNP, a biomarker of early diagnoses of heart failure.




Figure 2: The relation between the relative measured value and NT-proBNP concentration. The light density emitted from luciferase-luciferin at different NT-pro-BNP concentration is normalized into the relative value by that of buffer. The light density was determined by the ATP concentration and measured by luciferase/luciferin ATP detection kit. The measured value increases monotonically with the NT-proBNP concentration. The limiting sensitivity achieved 15 pg/ml.

2. Results

As Fig.1 shows, when target NT-proBNP was captured between antibody1 and antibody2, the motor embedded in chromatophore can be linked to the coverslip surface, while the free motors can be washed away. The number of captured motors is equal to that of NT-proBNP because of there is only one ϵ subunit in F_0F_1 -ATPase motor. The amount of ATP generated by F_0F_1 -ATPase, thus, is determined by the number of motors and time of synthesis based on the same storage energy in chromatophore. Here, the motor function as an amplifier in which the number of target NT-proBNP is amplified to 10^5 times ATP molecules during 30 minutes. The detector is able to be sensitive to 15 pg/ml of NT-proBNP as shown in Fig.2. Therefore, the active nanodevice has a potential to be developed into a dynamic biosensor.

3. Materials and Methods

3.1. Cell lines and reagents

Thermomicrobium roseum wa0073 (ATCC27502) was purchased from ATCC (USA). The luciferase/luciferin ATP detection kits were purchased from Promega Corporation (USA). ADP, (+)-biotin N-hydroxysuccinimide ester and NeutrAvidin were purchased from Sigma-Aldrich (St.

Louis, USA). The microplate luminometer was a Centro XS3 LB 960 (Germany).

3.2. Preparation of chromatophore containing the *F_oF₁*-ATPase

The chromatophore containing the *F_oF₁*-ATPase was isolated and purified according to our previous published protocols^[14, 18]. *Thermomicrobium roseum* was cultured at 60 °C for 24 h, and the cells were collected by centrifugation at 4,000 rpm for 20 minutes. The pellets were resuspended in 20 ml buffer A (pH 6.0), containing 20 mM Tricine-NaOH (pH 6.0), 2 mM MgCl₂, 100 mM NaCl, and 10% glycerin (v/v), and sonicated for 3 minutes. The lysate was centrifuged at 8000 rpm with R20A2 rotor for 30 minutes at 4 °C, and the supernatant was collected and centrifuged at 40,000 rpm for 90 minutes at 4 °C. The precipitate containing the chromatophores was resuspended in buffer B (pH 8.0), with 20 mM Tricine-NaOH (pH 8.0), 2 mM MgCl₂, 100 mM NaCl, and 10% glycerin (v/v), and stored at -80 °C for further use.

3.3. Preparation of monoclonal antibodies of ϵ -subunit

The ϵ -subunit was expressed and purified as in Ref.[19]. The ϵ -subunit monoclonal antibodies were prepared according to the method for monoclonal antibody production procedure in Ref.[20], purified by precipitation with 33% (NH₄)₂SO₄ at 4 °C for 12 h, and the IgG parts were separated using Sephadex G-200 and stored at -20 °C before use.

3.4. NT-proBNP detection

The NT-proBNP (from the Dapbiotech. company, Beijing) was incubated with antibody 24C5 for 1 h at 37 °C, followed by three times washings, then 1 μ l Ab-BNP2 antibody-conjugated chromatophore was added and incubated for 30 minutes at 37 °C. After three times washings, the ATP synthesis reaction was initiated and reacted for 60 minutes at 37 °C, then the ATP concentration was assayed using luciferase/luciferin ATP detection kits.

3.5. ATP synthesis assay

The ATP synthesis activity of *F_oF₁*-ATPase within the chromatophores was determined using the luciferin-luciferase method. 30 μ l luciferase/luciferin working solution was added into ATP synthesis well. The chemiluminescence signals displayed in microplate luminometer were recorded immediately. The YS101 type high sensitive detector of chemiluminescence was manufactured by Yishang Innovation Technology Co., Ltd. The relative light density emitted from luciferase-luciferin at different NT-pro-BNP concentration is shown in Fig.2. in which the light density of buffer is normalized. The results indicate that the measured value increases monotonically with the NT-proBNP concentration, and the sensitivity of detection was lower than 15 pg/ml.

4. Conclusions

We have developed a novel nanodevice constituted with a rotary motor and a “battery”, *F_oF₁*-ATPase and chromatophore. The former can processively rotate at about 10³ r.p.m for more than one hour once the latter was recharged by light. If the nanodevice is captured by a target such as NT-proBNP and processively rotate for 30 minutes, the number of targets will be amplified by 10⁵ ATP molecules^[15]. The sensitivity of the detection was lower than 15 pg/ml, which is more sensitive than the current sensitivity 220 pg/ml^[17]. This method has potential to be developed into an ultrasensitive biosensor to detect low expressed targets such as NT-proBNP.

Acknowledgements

This work is supported by the National Basic Research Program of China (973 Program) under grant No. 2013CB932804, the National Natural Science Foundation of China under Grant No. 11574329 and 11322543, the Science and Technology Planning Project of General Administration of Quality Supervision, Inspection and Quarantine (AQSIQ) of PR China under Grant No. 2015IK011, and AQSIQ industrial public service scientific research project of the Ministry of Science and Technology of P.R. China under Grant No. 201410049. The YS101 type high sensitive detector of chemiluminescence was manufactured by Yishang Innovation Technology Co., Ltd.

Reference

- [1] Boyer PD. The ATP synthase-A splendid molecular machine. *Annu. Rev. Biochem.* 1997; **66**: 717-749.
- [2] Noji H, Yasuda R, Yoshida M, *et.al.* Direct observation of the rotation of F₁-ATPase. *Nature* 1997; **386**: 299-302.
- [3] Yasuda Y, Noji H, Kinosita Jr K, *et.al.* F₁-ATPase Is a Highly Efficient Molecular Motor that Rotates with Discrete 120° Steps. *Cell* 1998; **93**: 1117-1124.
- [4] Abrahams JP, Leslie AGW, Lutter R, *et.al.* Structure at 2.8Å resolution of F₁-ATPase from bovine heart mitochondria. *Nature* 1994; **370**: 621- 628.
- [5] Diez M, Zimmermann B, Börsch M, *et.al.* Proton-powered subunit rotation in single membrane-bound FoF₁-ATP synthase. *Nat. Struct. Mol. Biol.* 2004; **11**: 135-141.
- [6] Toyabe S, Muneyuki E. Single molecule thermodynamics of ATP synthesis by F₁-ATPase. *New J. Phys.* 2015; **17**: 015008.
- [7] Shu YG, Yue JC, Ou-Yang ZC. FoF₁-ATPase, rotary motor and biosensor. *Nanoscale* 2010; **2**: 1284-1293.
- [8] Choi HJ, Montemagno CD. Artificial Organelle: ATP synthesis from cellular mimetic polymersomes. *Nano. Lett.* 2005; **5**: 2538-2542.
- [9] Feniouk BA, Cherepanov DA, Voskoboynikova NE, *et.al.* Chromatophore Vesicles of Rhodobacter capsulatus contain on average one FoF₁-ATP synthase each. *Biophys. J.* 2002; **82**: 1115-1122.
- [10] Shu YG, Lai PY. Systematic kinetics study of FoF₁-ATPase: analytic results and comparison with experiments. *J. Phys. Chem. B* 2008; **112**: 13453-13459.
- [11] Dong H, Nie RX, Hou X *et.al.* Assembly of FoF₁-ATPase into solid state nanoporous membrane. *Chem. Commun.* 2011; **47**: 3102-3104.
- [12] Zhang YH, Wang J, Cui YB, *et.al.* Rotary torque produced by proton motive force in FoF₁-ATP motor. *Biochem. Biophys. Res. Commun.* 2005; **331**: 370-374.
- [13] Deng ZT, Zhang Y, Yue JC, *et.al.* Green and orange CdTe quantum dots as effective pH-sensitive fluorescent probes for dual simultaneous and independent detection of viruses. *J. Phys. Chem. B* 2007; **111**: 12024-12031.
- [14] Cheng J, Zhang XA, Shu YG, *et.al.* FoF₁-ATPase activity regulated by external links on β subunits. *Biochem. Biophys. Res. Commun.* 2010; **391**: 182-186.
- [15] Shu YG, Ou-Yang ZC. microRNA detection with an active nanodevice FoF₁-ATPase. In *Proceedings of the 9th International Joint Conference on Biomedical Engineering Systems and Technologies (BIOSTEC 2016)* - Volume 1: *BIODEVICES*, pages 166-169. ISBN: 978-989-758-170-0
- [16] Hunt PJ, Richards AM, Nicholls MG, *et.al.* Immunoreactive aminoterminal pro-brain natriuretic peptide (NT-proBNP): a new marker of cardiac impairment. *Clinical Endocrinology* 1997; **47**: 287-297.
- [17] Sabatasso S, Vaucher P, Augsburger M, *et.al.* Sensitivity and specificity of NT-proBNP to detect heart failure at post mortem examination. *Int. J. Legal Med.* 2011; **125**: 849-856.
- [18] Shu YG, Ou-Yang ZC. FoF₁-ATPase stator regulation studied with a resonance model. In *Proceedings of the International Conference on Biomedical Electronics and Devices*, 2012, 132-137.

doi=10.5220/0003753401320137.

[19] Su T, Cui YB, Zhang XA, *et.al.* Constructing a novel Nanodevice powered by δ -free FoF₁-ATPase. *Biochem. Biophys. Res. Commun.* 2006; **350**: 1013-1018.

[20] Hanly WC, Artwohl JE, Bennett BT. Review of Polyclonal Antibody Production Procedures in Mammals and Poultry. *ILAR J.* 1995; **37**: 93-118.