

Advancements in electrochemical biosensing for respiratory virus detection: A review



Zhi Zhao ^{a, b, 1}, Changfu Huang ^{a, b, 1}, Ziyu Huang ^{a, b, 1}, Fengjuan Lin ^b, Qinlin He ^b,
Dan Tao ^{a, b}, Nicole Jaffrezic-Renault ^{c, *}, Zhenzhong Guo ^{a, **}

^a Hubei Province Key Laboratory of Occupational Hazard Identification and Control, Wuhan University of Science and Technology, Wuhan 430065, PR China

^b School of Public Health, Medical College, Wuhan University of Science and Technology, Wuhan 430065, PR China

^c University of Lyon, Institute of Analytical Sciences, UMR-CNRS 5280, 5, La Doua Street, Villeurbanne 69100, France

ARTICLE INFO

Article history:

Available online 12 March 2021

Keywords:

Electrochemical
Biosensors
Respiratory viruses
Virus detection
COVID-19
SARS-CoV-2

ABSTRACT

Respiratory viruses are real menace for human health which result in devastating epidemic disease. Consequently, it is in urgent need of identifying and quantifying virus with a rapid, sensitive and precise approach. The study of electrochemical biosensors for respiratory virus detection has become one of the most rapidly developing scientific fields. Recent developments in electrochemical biosensors concerning respiratory virus detection are comprehensively reviewed in this paper. This review is structured along common detecting objects of respiratory viruses, electrochemical biosensors, electrochemical biosensors for respiratory virus detection and future challenges. The electrochemical biosensors for respiratory virus detection are introduced, including nucleic acids-based, immunosensors and other affinity biosensors. Lastly, for Coronavirus disease 2019 (COVID-19) diagnosis, the future challenges regarding developing electrochemical biosensor-based Point-of-Care Tests (POCTs) are summarized. This review is expected to provide a helpful guide for the researchers entering this interdisciplinary field and developing more novel electrochemical biosensors for respiratory virus detection.

© 2021 Elsevier B.V. All rights reserved.

1. Introduction

Respiratory viruses, well-known as influenza virus and coronavirus, usually result in viral respiratory infections through contact as well as airborne transmission [1]. The infected individuals generally present fever, dry cough, fatigue, sputum production and loss of smell, such acute respiratory virus illnesses symptoms. Though sounds like a mild cold, acute respiratory disease caused by respiratory viruses have brought death and pandemics over the past years [2,3]. Only Respiratory Syncytial Viral (RSV) could lead to 14,000 deaths among adults older than 65 years every year in the US [4]. Currently, Severe Acute Respiratory Syndrome coronavirus 2 (SARS-CoV-2) is the responsible culprits of the Coronavirus disease 2019 (COVID-19) pandemic. According to the data collected from the World Health Organization (WHO), there are totally over

6284,000 cases and 1465,000 deaths in 220 countries, areas or territories by 2 December 2020 [5]. The prevention and control have been taken depending on the features of the spread of the respiratory viruses, such as wearing the P2/N95 masks to prevent airborne spread, cleaning contaminated surfaces to avoid risky contact [6]. However, owing to the non-specific and comprehensive symptoms among the respiratory viruses and the silent transmission from positive asymptomatic, early accurate diagnosis and isolation of patients remain to be crucial for controlling the pandemic resulted by the respiratory viruses [7]. Thus, respiratory virus detection would be particularly decisive.

Conventional methods for respiratory virus detection are mostly based on lab-based techniques. From initial virus cultures, morphological observation, and serological tests to subsequent reverse transcription–polymerase chain reaction (RT–PCR) [8], isothermal amplification techniques [9], immunochromatography (IC) [10], enzyme-linked immunosorbent assay (ELISA) or an immunofluorescence assay (IFA) [11] and classical diagnostic methods have helped physicians to distinguish the causative agents with accuracy. Although, in clinical practice, cumbersome sample-preparation, high cost, professional operators and time-consuming

* Corresponding author.

** Corresponding author.

E-mail addresses: nicole.jaffrezic@univ-lyon1.fr (N. Jaffrezic-Renault), zhongbujueqi@hotmail.com (Z. Guo).

¹ These authors contributed equally to this work.

Abbreviations

ACE2	Angiotensin-converting enzyme II	mAb	Monoclonal antibodies
ALP	Alkaline phosphatase	MCH	6-mercapto-1-hexanol
AP	Auxiliary probe	MERS-CoV	Middle East Respiratory Syndrome coronavirus
APP	4-amino phenyl phosphate	MNPs	Magnetic nanoparticles
Au NPs	Gold nanoparticles	N	Nucleocapsid
BDD	Boron-doped diamond	NA	Neuraminidase
bi-FMNs	bifunctional fluorescence magnetic nanospheres	ORF	Open reading frame
CNTs	Carbon nanotubes	pAb	Polyclonal antibodies
COVID-19	Coronavirus disease 2019	p-AP	P-aminophenol
CP	Capture probe	p-APP	P-aminophenyl phosphate monohydrate.
CV	Cyclic voltammetry	PCR	Polymerase chain reaction
DPV	Differential pulse voltammetry	PDMS	Polydimethylsiloxane
dsDNA	Double strand DNA	PNA	Peanut agglutinin
E	Envelope	POCTs	Point-of-Care Tests
EDOT	3,4-ethylenedioxythiophene	RBD	Receptor binding domain
EIS	Electrochemical impedance spectroscopy	RGO	Reduced graphene oxide
ELISA	Enzyme-linked immuno sorbent assay	RSV	Respiratory Syncytial Viral
Fab	Fragment-antigen binding	RT-PCR	Reverse transcription-polymerase chain reaction
FDA	Food and Drug Administration	S	Spike
GO	Graphene oxide	SAM	Self-assembled monolayer
HA	Hemagglutinin	SARS-CoV	Severe Acute Respiratory Syndrome coronavirus
HAU	Hemagglutination unit	SARS-CoV-2	Severe Acute Respiratory Syndrome coronavirus 2
HEPES	4-(2-hydroxyethyl)-1-piperazineethanesulfonic acid	scFv	Single-chain Fv fragments
IC	Immunochromatography	SELEX	Systematic Evolution of Ligands by Exponential Enrichment
IFA	immunofluorescence assay	SERS	Surface Enhanced Raman Scattering
IgG	Immunoglobulin G	SPCE	Screen-printed carbon electrode
IgM	Immunoglobulin M	SPEs	Screen-printed electrodes
ITO	Indium tin oxide	SPR	Surface Plasmon Resonance
K _D	Dissociation constants	ssDNA	Single-stranded DNA
LFIA	Lateral Flow Immunoassay	SWV	Square wave voltammetry
LOD	Limit of detection	upE	RNA upstream of the E gene
LP	Label probe	UTR	Untranslated Regions
M	Membrane	WHO	World Health Organization
M2	Matrix protein 2		

equally become the drawbacks of most classical lab-based techniques [12]. There is still a demand to exploit rapid, simple, cheap assays with precision on respiratory virus detection. Biosensors, cooperating the bio-recognition elements with the sensor system, are capable of recognizing the targets with high sensitivity and selectivity [13]. Biosensors have arisen in numerous areas, including environment monitoring, food safety, drug control, disease diagnosis and so on [14]. Among them, many optical based techniques are proposed for virus detection such as Surface Plasmon Resonance (SPR) [15], Lateral Flow Immunoassay (LFIA) [16], Surface Enhanced Raman Scattering (SERS) [17].

Electrochemical biosensors have aroused burgeoning attention because of intrinsic strengths: simplicity, rapid response, flexibility, miniaturized instrumentation, excellent sensitivity and low cost [18], which have been emerging alternative tools for the quantitative or semi-quantitative analyzing respiratory viruses. Excellent reviews are accessible in the literature about the state-of-art of electrochemical biosensors for pathogen detection: Anusha et al. [19] highlighted various types of electrochemical biosensing techniques and the role of biorecognition molecules in sensing of dengue virus; Kaushik et al. [20] discussed the recent developments in developing intelligent sensing strategies to monitor Zika virus; Rasouli et al. [21] gathered the advancements in electrochemical DNA biosensors for the detection of human papillomavirus virus. However, these reviews are all restricted to include only a kind of virus, which lack of the summary of electrochemical

detection methods for a class of viruses. Furthermore, there are other excellent reviews that present the current state of biosensors for respiratory virus detection: Ribeiro et al. [22] covered important advancements in the biosensor field in terms of most current respiratory viruses, presenting the development in the assembly of the devices and figures of advantages. Samson et al. [23] present all the novel types of biosensors that could be used for the rapid detection of COVID-19. Ruiz de Eguilaz et al. [24] reported on virus and antibody detection using electrochemical methods, focusing on recent key innovations which drive the progress of portable, high performance point-of-care technologies. Nevertheless, few articles cover and focus on both electrochemical biosensor background and respiratory virus detection, or their key aspects are a kind of special field. For example, Nelson et al. [25] provided a brief overview of currently available Point-of-Care Tests (POCTs) for the diagnosis of emerging and new respiratory viruses along with their merits and limitations, and discussed recently published methods and techniques with a potential use in future POCTs. Therefore, our review article aims to fill the blank by combining essential background information about electrochemical biosensors with the rapidly moving advancements of electrochemical biosensors for respiratory virus detection.

Hence, we reviewed the recent advances in electrochemical biosensors for respiratory virus detection. In this review, common detecting objects of respiratory viruses, electrochemical biosensors, electrochemical biosensors for respiratory virus

detection and future challenges are discussed successively. When exploring the methods for testing a new virus, it is often worthy of reviewing the already existing methods for other congeneric virus in comparison. Therefore, it is anticipated that this review regarding respiratory viruses will provide a complete guide to develop novel COVID-19 diagnosis assays with prominent accuracy and sensitivity, thereby performing appropriate antiviral therapies for patients.

2. Common detecting objects of respiratory viruses

For respiratory virus detection, the whole virus, their structural proteins, gene sequences and antibodies could be the targets. Here we will give a comprehensive discussion regarding common detecting objects of representative respiratory viruses: influenza virus, MERS-CoV and SARS-CoV-2.

2.1. Whole virus and their structural proteins

2.1.1. Influenza virus

Basically, the whole influenza virus and the structural proteins, including M1 protein, hemagglutinin (HA) and neuraminidase (NA) all can serve as antigens for influenza virus detection. The type of influenza virus: A, B and C are classified according to the encoding proteins: matrix protein M1 and viral nucleoproteins. M1 protein is the only essential viral component for virus-like particles formation and suitable for all serotypes of influenza virus [26]. Besides, the virus can combine with the host cells through the contacts of HA and NA. There has been 18 HA and 11 NA variants so far owing to their high variety. The subtype of influenza virus is usually decided by the properties of HA and NA [27].

2.1.2. The Middle East Respiratory Syndrome corona virus (MERS-CoV)

Belonging to coronavirus, MERS-CoV owns four structural proteins: spike (S) protein, envelope (E) protein, membrane (M) protein and nucleocapsid (N) protein. The S protein is involved in the binding process between the virus and the host cell surface receptors. The E protein is the smallest protein in the major structural proteins, mediating virus assembly and budding. The M protein is able to decide the shape of the virus envelope. The N protein is the only protein binding to the RNA genome [28]. Among these, the S protein is the most-frequently used antigens because of its significant role in the attachment of the virus to the host cells. However, there are still few published articles about the detection of the whole MERS-CoV virus.

2.1.3. SARS-CoV-2

The whole SARS-CoV-2 and their four structural proteins: S, E, M and N could be used as targets for SARS-CoV-2 detection. M and E protein are essential proteins when occurring viral assembly, while S and N proteins are the most significant biomarkers in terms of COVID-19 early diagnosis. The S protein can mediate the fusion of the virus and the host cell membrane, making the virus more easily enter the host cells [29]. Besides, the highly immunogenic S protein could promote producing neutralizing antibodies as well as T-cell responses in the SARS-CoV-2 patients [30]. Moreover, the S1 subunit of the S protein exhibits the receptor binding domain (RBD) with strong binding affinity for the host angiotensin-converting enzyme II (ACE2) receptor on the human cells [31]. Therefore, the RBD protein of SARS-CoV-2 could also be selected as the targets.

2.2. Gene sequences derived from viruses

2.2.1. Influenza virus

The origins of the derived gene are generally classified into two groups: (i) deriving from the biomarkers of the influenza virus. The most frequently-used RNA transcripts and DNA oligonucleotides when diagnosing influenza virus are the HA gene of them. (ii) sequences of DNA derived from influenza virus then amplified by polymerase chain reaction (PCR). Although some electrochemical biosensors are able to detect gene sequences in the pure samples, there is still distance before their application to real samples owing to the high background responses from matrix effects [32]. Therefore, researchers begin to detect the amplified products from PCR to solve the problems from real samples. Nevertheless, for amplified products, the efficiency will decrease when the targets and probe hybridize because of interference factors [33]. In fact, it is the ideal that the electrochemical biosensors do not rely on the PCR technique or less, which may increase workload. Unluckily, the electrochemical biosensors independent on PCR are chiefly suitable for abundant DNA targets. The low-abundance DNA analytes even if not depending on PCR, still involve quantitative real-time PCR [34].

2.2.2. MERS-CoV

The genome of MERS-CoV includes 30,119 nucleotides and 11 open reading frames (ORF). The first open reading frames (ORF 1a and 1b) at the 5'-Untranslated Regions (UTR) (278 nucleotides) have become essential detecting objects in the MERS-CoV specie identification, which are predicted to encode nonstructural proteins [35]. The genes downstream to ORF1ab encode for structural proteins and accessory proteins (Fig. 1). The RNA upstream of the E gene (upE) has also been recommended by WHO for MERS-CoV detection [36]. Besides, with a sensitivity of ≤ 10 copies/reaction, identifying the MERS-CoV N gene is an alternative method complementing upE and ORF 1a approaches, recommended by the US Food and Drug Administration (FDA) [37].

2.2.3. SARS-CoV-2

Similar to MERS-CoV, the 5'-terminal genome ORF1a/b encode two large polyproteins, the other ORFs on the genome encode four main structural proteins and accessory proteins. ORF 1a, ORF 1b, non-structural RNA-dependent RNA polymerase, S gene, N gene of Severe Acute Respiratory Syndrome coronavirus (SARS-CoV) are the preferred targets for nucleic acid tests [38]. Owing to the 79% similarity of the whole-genome between SARS-CoV and SARS-CoV-2, unique primers or guide RNAs are required for distinguishing SARS-CoV-2 with no cross-reactivity for SARS-CoV [39]. To avoid the "false negative" result, multiple gene sequences are usually detected simultaneously in the COVID-19 diagnosis.

2.3. Antibodies

2.3.1. Influenza virus

It is well known that the immunoglobulin M (IgM) presents in patients' blood after 3–6 days, and immunoglobulin G (IgG) presents after 8 days [40]. Moreover, the specific antibodies of structural protein are also alternatives for influenza virus detection, such as the antibodies of HA and NA. The vaccines could induce the increase of virus-specific antibodies about virus invasion [41].

2.3.2. MERS-CoV

Generally, antibodies to proteins S, 3a, N, and 9b could be detected in the serum samples of convalescent-phase patients [42]. Anti-S and anti-N are detectable until week 30, and anti-N appears earlier than anti-S, so anti-S may be preferable with convalescent sera comparatively [43]. Whereas, for early diagnosis of diseases




Fig. 1. The genome of SARS-CoV, MERS-CoV and SARS-CoV-2, all of which consist of conserved replicase domain (ORF 1ab) (blue). The structural genes (green) S, E, M and N encode the structural proteins: spike (S) protein, envelope (E) protein, membrane (M) protein and nucleocapsid (N) protein, respectively. Different coronaviruses have different accessory genes (orange). Reproduced with permission from Ref. [35].

related to respiratory viruses, detecting relative specific antibodies is not appropriate, which may be useful for treating convalescent patients [44].

2.3.3. SARS-CoV-2

For SARS-CoV-2 infection, IgG against N protein is detectable as early as 4 days after infection. Zhang et al. [45] have confirmed that IgG and IgM could be detected by enzyme-linked immuno sorbent assay (ELISA) in the serum of the patients after 5 days of infection. After SARS-CoV infection, it has been proved that the sensitivity of N-based IgG ELISA (94.7%) is significantly higher than that of S-based IgG ELISA (58.9%) [46], but there is still no report to present the sensitivity of SARS-CoV-2 IgG/IgM.

3. Electrochemical biosensors

The development of chemical and biosensors is one of the most active fields in current analysis and research. Biosensors are small devices including bio-recognition elements and signal transducers, which can be used for the direct detection of objects in samples [47]. Electrochemical sensors, using electrodes as energy exchangers, are the important branch of biosensors. Electrochemical sensors occupy an important position in current biosensors, widely applied in the clinical, industrial, environmental and agricultural analysis [48,49]. Therefore, we discussed the electrochemical biosensors utilizing the framework upon the working principles, merits and defects of electrochemical biosensors, electrochemical transduction, bio-recognition elements and nanomaterials. The components and principle as to electrochemical biosensors used for the detection of the respiratory viruses are displayed in the Fig. 2.

3.1. Working principles, merits and defects

The biosensor is an analytical system composed of three essential parts: the bio-recognition element, the transducer and signal output [50]. The diagnosis molecules process could be summarized as: the targets firstly are recognized by the specific bio-recognition elements via amounts of interaction like the covalent bond or non-covalent bond; then the changes could be felt by the transducer and further translated into the digital detector; finally, the digital signals are output by the digital device such as computers and phones [51]. Particularly, the transducers of the electrochemical biosensors are a variety of electrodes, such as glassy carbon electrodes, gold electrodes, screen-printed electrodes

(SPEs) and carbon paste electrodes. The electrodes in the electrochemical biosensors provide the platform for kinds of modification, which aim at improving the property of analytical system: sensitivity, selectivity, stability, reproducibility and so on [52]. Thus, the well-designed electrochemical biosensors exhibit abundant advantages: low-cost, quick-response, simple, high sensitivity with the help of electrode fabrication and the bio-recognition element design [53].

On the one hand, compared with other transduction processes, that of the electrochemical biosensors could be completed at the electrochemical workstation at least, which reduce the cost of test greatly. This is because the electrochemical detection is based on the result of direct electronic signals, like amperometric, voltammetric and impedimetric changes. Therefore, the detecting process could be over in a short time [54]. Moreover, the electrochemical biosensors are capable of realizing label-free detection without the incorporation with any label, making POCTs possible [55]. In addition, the high sensitivity of electrochemical biosensors could be ensured by applying the bio-recognition elements with high specificity and affinity or decorating the electrodes with special materials with excellent electronic performance [56]. Over the past few years, the electrochemical biosensors have gained numerous progresses in the analytical field owing to the advantages, especially in the diagnosis of the pathogens, offering a kind of new possibility for healthcare. The electrochemical biosensors have been utilized to monitor the virus particles during virus outbreaks in epidemic areas.

On the other hand, even if most of electrochemical biosensors are successfully tested in buffered solutions or diluted real samples spiked with targets, matrix effects always influence the analytical performance of the biosensors in practice. Therefore, the stability and accuracy of electrochemical biosensors remain to be the biggest limitations, especially after repeated usages and long storage. Besides, owing to some interaction between the biorecognition elements and targets is irreversible, thus these electrochemical biosensors could only be used once, increasing the cost of testing.

3.2. Electrochemical transduction

There have been a variety of electrochemical biosensors fabricated for respiratory virus detection, the most commonly used electrochemical techniques are chronoamperometry, cyclic voltammetry (CV), differential pulse voltammetry (DPV), and square wave voltammetry (SWV) and the electrochemical impedance




Fig. 2. Schematic description of components and principle for electrochemical biosensors used in detection of respiratory viruses. (ssDNA: Single-stranded DNA.)

spectroscopy (EIS) whose principles are described in Ref. [57]. Voltammetric biosensors (CV, DPV and SWV etc.) have been widely implemented for respiratory virus detection owing to their fast response, less sample, simple preparation and excellent reproducibility. However, on account of the requirement of the extra electroactive species, its application is limited in some degree for respiratory virus detection [58]. EIS technique is attractive for biomedical and biological fields in accordance with the ability of revealing the weak interaction between different species. Moreover, the EIS is the only research method for studying the interactions between biolayers, which have active effect on the designing rapid, stable, sensitive and portable electrochemical biosensors for respiratory virus detection.

3.3. Bio-recognition elements

Bio-recognition element is the key component of the electrochemical biosensors. Only when the recognition of the targets is guaranteed, the later steps can start. Bio-recognition element in the electrochemical biosensors could be divided into biocatalytic and biocomplexing. Biocatalytic elements, such as enzymes, cells and tissues, are based on the catalytic reactions for recognizing targets. For example, enzymes are involved in various chemical sensing applications, which are primarily served as signal labels in the respiratory virus detection. Enzymes are usually introduced during the secondary binding process. Biocomplexing elements are the most-frequently used bio-recognition elements in the respiratory virus detection, which rely on the interaction of targets with macromolecules or organized molecular assemblies. Antibodies, aptamer and peptide are common bio-recognition elements in the respiratory virus detection. Some researchers also used imprinted polymers as bio-recognition elements in the electrochemical biosensors.

3.4. Nanomaterials

The modification of the working electrode is very important in the fabrication process of the electrochemical biosensors, resulting in the link between analytes in the bulk solution and sensing

interface. The affinity of the biosensors is usually improved by modifying with bio-recognition elements, and the sensitivity of the biosensors is often enhanced by realizing signal amplification through the addition of nanomaterials. The common nanomaterials and their properties utilized in electrochemical biosensing are briefly introduced as followed:

- (i) Gold-based nanomaterials. Metallic nanoparticles, owning unique optical/electrical properties, especially gold nanoparticles (Au NPs) have been served as stable immobilizer for bio-recognition elements without distorting their bioactivity, meanwhile facilitating excellent electron transfer between the targets and sensing interface. Both various functional groups ($-SH$, $-NH_2$, $-CN$) and amine or thiol linkers could coordinate Au NPs attachment forming multi-layered bionanocomposite-film on the interface [59].
- (ii) Carbon-based nanomaterials. Graphene oxide (GO), reduced graphene oxide (RGO) and carbon nanotubes (CNTs) are used under other circumstance in designing biosensors with high sensitivity. The main advantage of the carbon-based nanomaterials is increasing the electron transfer rates. Additionally, by chemically functionalizing the surface architecture, both the electrical conductivity and the surface area could be enhanced and result in the improvement of the sensitivity of the biosensors [60].
- (iii) Magnetic nanoparticles (MNPs). Their handling and the large variation of surface allow them to be employed as coating support for further modification, and its high surface energy and large surface area allow electrons transfer more efficiently at the same time. Moreover, owing to being controllable by external magnet, when attached with labels and bio-recognition elements simultaneously, the MNPs are able to realize the reproducible magnetic virus separation and further signal amplification in the real clinical samples [61].

4. Electrochemical biosensors for respiratory virus detection

According to the type of bio-recognition element, we divided the electrochemical biosensors for respiratory virus detection into

three groups: nucleic acid-based, immunosensors and other affinity biosensors. Their advantages and limitations when applied for respiratory virus detection are summarized in Table 1. Next, we would review the recent electrochemical biosensors for respiratory virus detection in terms of the classification.

4.1. Nucleic acids-based

Electrochemical biosensors based on nucleic acids as recognition element generally used DNA or RNA. The DNA or RNA sequences are usually immobilized on the sensing interface. Owing to the specific binding between probes and targets, the formation on the electrode, like double strand DNA (dsDNA), could trigger the properties change of the electrode surface, which can be detected via electrochemical techniques. The electrochemical signals are generally from the electron transfer of redox-active probe with the electrode, and the common redox-active probes are $[\text{Fe}(\text{CN})_6]^{3-/4-}$ and $[\text{Ru}(\text{NH}_3)_6]^{3+}$ complexes [62]. Nucleic acid-based electrochemical biosensors own various merits: high specificity, stability, possibilities for miniaturization, which are very attractive for the fabrication of biosensors [63]. The nucleic acid-based electrochemical biosensors for respiratory virus detection are summarized in Table 2.

Single-stranded (ss) DNA, hairpin DNA, peptide nucleic acid, and locked nucleic acid are the probe often used in the electrochemical biosensors [73]. The most common probe in kinds of nucleic acid-based electrochemical biosensors is ssDNA. Specially, aptamer, a kind of ssDNA with high affinity and selectivity toward targets has been widely utilized in ssDNA-based electrochemical biosensors. The aptamer is selected from Systematic Evolution of Ligands by Exponential Enrichment (SELEX), which could combine with targets via interaction like hydrogen bonds, van der Waals forces [74]. The targets of aptamer can be proteins, nucleic acid or chemical substances. Comparing with antibodies, the aptamer is size-smaller, low-cost, more stable, easy-to-produce and of lower immunogenicity, which has considerable potential for developing novel electrochemical biosensors with high specificity [75].

i) Detection of proteins or whole virus

Bhardwaj et al. [45] selected an ssDNA aptamer against stem region of HA protein of influenza A virus by five rounds of SELEX. Simultaneously, mini-HA protein and whole H1N1 virus could be recognized by this aptamer. The dissociation constants (K_D) of the developed aptamer are higher than the average K_D of the influenza virus antibodies, which means the affinity of aptamers is superior to relative antibodies. The specific aptamer was adsorbed on the

working areas of the ITO/glass strips previously functionalized by a polyethylenimine solution, the final aptasensor achieved a H1N1 virus limit of detection (LOD) of 3.7 plaque-forming units (PFU) per mL. More importantly, six strains of H1N1 influenza A viruses could be identified by the aptamer-modified electrode, indicating the possibility of the rapid subtyping of H1N1 and diagnostic applications. Apart from single aptamer as recognition element, the most usual detecting assay in aptamer-based biosensor is the aptamer-target-antibody sandwich method. The dual recognition pattern greatly improves the accuracy and selectivity of the detection process, decreasing the LOD of biosensors. Diba et al. [67] fabricated an amperometric bio-affinity electrochemical sensor for avian influenza virus proteins detection with aptamer modified Au NPs decorated on carbon chips. The electrochemical signals were from the reaction between alkaline phosphatase (ALP) and 4-amino phenyl phosphate (APP). The current generated from the Au NPs-aptamer/H5N1/anti-H5N1-ALP sandwich complex with the enzyme substrate increased with the concentration of H5N1. Differential pulse voltammetry was used for detection with a linear dynamic range of 100 fM–10 pM. The 100 fM LOD of the aptamer-antibody sandwich platform compares favorably with commercial antibody ELISA kits. The proposed biosensor has been used in the detecting H5N1 protein for diluted human serum samples.

However, to immobilize the aptamer, the fabrications of the electrode often involve labeling and anchoring operation, which required complex steps. In order to solve the problem, Lee et al. [65] introduced a multi-functional probe which consists of recognition part, signal producing part and combining part. It was immobilized on the porous Au NPs modified electrode for avian influenza virus detection. The recognition part was based on the specific aptamer of HA protein. The DNA 3 way-junction probe could realize three steps: recognizing, immobilizing and generating without additional process and loss of functionality. Besides, the multifunctional DNA probe could also insert redox probe, functional groups and other aptamers. The multi-functional probe-based electrochemical biosensor showed the LOD of HA protein at 1 pM in 4-(2-hydroxyethyl)-1-piperazineethanesulfonic acid (HEPES) solution and 1 pM in diluted-chicken serum, respectively. Although the proposed biosensor didn't own the lowest LOD, the redox probe labeling step and signal amplification step were both reduced compared to previous works.

ii) Detection of PCR ssDNA products

Alafeef et al. [76] reported using antisense oligonucleotides directed electrochemical biosensor chip for realizing the digital diagnosis. The sensing chip was based on the paper-based

Table 1
Advantages and limitations of common bio-recognition elements applied for respiratory virus detection.

Type of electrochemical biosensors	bio-recognition elements	Advantages	limitations
Nucleic acids-based	ss-DNA	Detection of ssDNA PCR products, easy to produce and more stable	Limited for gene sequence detection, strict hybridization conditions and expensive
	Aptamer	Size-smaller, low-cost, more stable, easy-to-produce and of lower immunogenicity	The strict hybridization conditions, long-term SELEX process and sometimes need complex steps
Immunosensors	Monoclonal antibodies (mAb)	More specific than pAb, avoiding the cross reaction	Expensive, instable and complexity-to-synthesis
	Polyclonal antibodies (pAb)	Less expensive, more epitopes and mass-productive	Instable and easily appearing cross reaction
	Antibody single-chain Fv fragments (scFv)	Highly customizable, low variability and smaller size compared with whole antibody	Slow synthesis, lower affinities compared with whole antibodies and can't be produced for small molecules
Others affinity biosensors	Fetuin A	Low-cost, selective and lower limit of detection	Limited to influenza virus
	Peptides	Easily being designed and prepared	Less specific compared with aptamers and antibodies
	Glycans	Storing more code information	Limited to a few viruses, the affinities need to be proved further

Table 2
Nucleic acid-based Electrochemical biosensors for respiratory virus detection.

Type	Virus	Recognition element	Linear range	LOD	Electrochemical method	Ref.
Detection of proteins or whole virus	H1N1	aptamer	10^1 PFU mL ⁻¹ – 10^4 PFU mL ⁻¹	3.7 PFU mL ⁻¹	DPV	[45]
	H5N1	DNA probe	1 pM – 100 nM	HEPES buffer: 1 pM chicken serum: 1 pM	CV	[64]
	AIV	anti-AIV NP aptamer	2 nM – 2 μ M	1.13 nM	CV	[65]
	H7N9	DNA tetrahedral probe and ssDNA	1 pM – 100 nM	100 fM	amperometry	[66]
	H5N1	aptamer	100 fM – 10 pM	100 fM	DPV	[67]
	H1N1	aptamer against inactivated intact H1N1	/	0.3 ng mL ⁻¹	EIS	[68]
Detection of PCR ssDNA products	H5N1	thiolated ssDNA probe	/	RNA transcripts: 10 pM DNA oligonucleotides: 1 pM	SWV	[69]
	H5N1	ssDNA probe	1–10 pM	1.39 pM	SWV	[70]
	Influenza A	DNA probe	1.0 fM – 1.0 nM	84 aM	DPV	[71]
	H1N1	HA gene specific ssDNA probe	0.1–400 ng in 6 μ L	0.004 ng in 6 μ L	EIS	[72]

electrochemical sensor chip modified with Au NPs. The highly specific antisense oligonucleotides towards viral N gene were served as bio-recognition element, yielding a nucleic-acid-testing device with a readout presented by a hand-held reader. The samples collected from Vero cells infected with SARS-CoV-2 virus and clinical specimens have been tested for the device, whose incubation time was less than 5 min, with a sensitivity of 231 (copies μ L⁻¹)⁻¹ and LOD of 6.9 copies μ L⁻¹ without further amplification. For most nucleic acid-based electrochemical biosensors, the nucleic acid-probes are generally immobilized on the sensing interface through the attachment between points. The density of the recognition elements couldn't be ensured to be homogeneous, resulting in the additional process to block the unspecific adsorption [77], the DNA nanotechnology has been as the solution to solve the problem. The DNA with different structures is designed to control the recognition, such as DNA tetrahedra. The three vertices of the DNA tetrahedra are usually modified with thiol groups, the DNA tetrahedra will attach to the electrode surface via Au–S bond thus one signal probe could be immobilized on one DNA tetrahedra with the fourth vertex [61]. Comparing to the conventional point-tethered signal probe, the signal anchored by DNA tetrahedra present 5000-fold greater affinity [78]. Essentially, because of the high mechanical rigidity of the DNA tetrahedra, the signal probes will keep an upright orientation on the electrode surface even without the help of 6-mercapto-1-hexanol (MCH). Latest advances have also extended the applications of DNA tetrahedra in nucleic acid-based electrochemical biosensors. Dong et al. [66] developed a DNA tetrahedra-based electrochemical biosensor for H7N9 virus ssDNA detection, the amperometric signals were recorded from the interaction between the avidin-horseradish peroxidase attached to bio-ssDNA (biotin-labeled ssDNA) and 3,3',5,5'-tetramethylbenzidine substrate. Before testing, H7N9 virus cDNA was employed to conducting asymmetric PCR for obtain H7N9 virus ssDNA targets, the dependence degree study of the developed biosensors on PCR is also proceeded, the results showed ssDNA products from only one cycle of asymmetric PCR could be identified by the proposed sensor platform. The detection limit of the biosensor for asymmetric PCR ssDNA products was determined to be 97 fM. The asymmetric PCR ssDNA products and PCR-free samples both could be distinguished from zero samples by DNA biosensor. It is also the first time that the DNA tetrahedra-based electrochemical biosensor was proposed to be tested in the clinical samples, which potentially verified the practicability of DNA tetrahedra probe (Fig. 3).

Zhao et al. [79] firstly proposed supersandwich-type electrochemical biosensor regarding SARS-CoV-2 from COVID-19 patients by a smartphone (Fig. 4). The supersandwich-type electrochemical

biosensor included: capture probe (CP), auxiliary probe (AP), label probe (LP), and target sequence. The 5'- and 3'-terminals of target sequence are complementary to CP and LP, respectively. The 5'- and 3'-regions of AP have complementary sequences with two LP regions. The detection was based on using CP and LP, AP and LP to hybridize frequently for producing long concatemers, resulting in high sensitivity. Besides, p-sulfocalix [8] arene functionalized graphene was utilized to enrich toluidine blue, which was an approach of facilitating of LP with signal probes for selectivity enhancement. The detectable ratios (85.5% and 46.2%) were rather higher than those that were obtained using RT-PCR (56.5% and 7.7%) according to the testing for 88 RNA extracts from 25 SARS-CoV-2-confirmed patients and eight recovery patients.

Totally, for nucleic acids hybridization assays, electrochemical biosensors based on nucleic acids probe is the first choice, and aptasensors are suitable for both nucleic acids and other small molecules. The affinity of the probe depending by the sequences selection of the probe mostly decides the specificity of the electrochemical biosensors. The conditions of the hybridization such as the buffer composition and temperatures are also the influencing factors. Therefore, the design of electrochemical biosensors based on nucleic acids probe are comparatively strict. Besides, when the sensitivity of the biosensor is insufficient, it is common to use tagged hairpin probes or hybridized tapered sequences as amplification steps, which may add additional experimental steps.

4.2. Immunosensors

Antibodies are the bio-recognition elements of the electrochemical immunosensors. Antibodies are a series of serum proteins produced by B-lymphocytes and plasma cells, which could recognize and bind the targets (antigens). The antibody contains two antibody fragment-antigen binding (Fab) that are held by the key hinge disulfide bridges. The disulfide-termed Fab fragments are named Fab' which allow the binding with the sensing interface via the covalent bond [80]. Antibodies are the workhorse in commercial and lab bioanalytical assays due to their high specificity, extreme affinity and great sensitivity, showing interesting applications for detecting virus, proteins, and cancer cells [81]. The antibodies could be obtained by amounts of methods, natural or recombinant, as monoclonal or as polyclonal. Nevertheless, comparing to the nucleic acid elements, the defects of the antibodies are high-cost, instability, complexity-to-synthesis, the affinity of which may be affected by adding the signal tags, and can't be used for small molecules, drugs and metal ions [82].




Fig. 3. A DNA tetrahedral nanostructure-based electrochemical biosensor was developed to detect avian influenza A (H7N9) virus through recognizing a fragment of the hemagglutinin gene sequence. Biotin-labeled (bio)-ssDNA was the bio-recognition element toward targets, which also could combine with avidin-horseradish peroxidase (HRP) probes through biotin-avidin interaction. The DNA hybridization hence was transformed into the redox reaction of TMB (enhanced K-blue substrate) and H_2O_2 . Reproduced with permission from Ref. [66].

The most prominent antibodies in respiratory virus detection are monoclonal antibodies (mAb), polyclonal antibodies (pAb) and antibody single-chain Fv fragments (scFv). mAb are more specific than pAb because mAb could only combine with single epitope hence avoiding the cross reaction, and the pAb are produced towards various epitopes on a single antigen [83]. While the pAb are less expensive and mass-productive providing the widespread application in biosensors construction. The scFv fragments include one light chain and one heavy chain with a molecular weight of 30 kDa, with smaller size compared with whole antibody and low variability, the scFv fragments are brilliant for antigen capture [84]. The merits of the antibody-antigen reaction are high specificity, reversible binding between surface chemical groups, suitable ratio and concentration and staged reaction. The special properties of antibody-antigen reaction make antibody-based electrochemical biosensors being one of the most versatile and available detection tools for respiratory virus. The antibody-based electrochemical biosensors for respiratory virus detection are summarized in Table 3.

According to different antibodies application, usually mimicking ELISA, the antibody-based electrochemical biosensors consist the following patterns: standard (non-competitive), competitive direct, competitive indirect and sandwich. Generally, the specific antibodies are immobilized on the transducer surface in the respiratory virus detection, hence standard (non-competitive) and sandwich are the most used antibody formats in the respiratory virus detection with electrochemical biosensors. Competitive direct and competitive indirect are less commonly used, because respiratory viruses are usually small-sized, and difficult to be attached on the electrode surface. Furthermore, depending on if the labels are used, the antibody-based electrochemical biosensors could be divided into label-free immunosensors and label-based immunosensors.

4.2.1. Label-free Immunosensors

Standard is the representative antibody format in label-free immunosensors. The virus particles are captured by the antibodies modified electrode, generating the properties change of sensing interface. The signals could be detected directly with the electrochemical workstation. Label-free electrochemical biosensors are the fastest and simplest with high selectivity and non-cross-reactivity, widely used in the rapid and stable monitoring of respiratory viruses. EIS is the most commonly used electrochemical techniques in the label-free immunosensors, the change from

before and after binding to the targets are directly transferred into the change of the interfacial impedance or the change in charge transfer resistance to electroactive probe dissolved in electrolyte. Nidzworski et al. [91] employed the boron-doped diamond (BDD) electrode functionalized with polyclonal anti-M1 antibodies for influenza virus detection. The BDD electrode was dealt with 4-aminobenzoic acid for forming self-assembled monolayer (SAM), then anti-M1 antibodies could be immobilized on the SAM. Hence, the M1 protein was captured onto the BDD electrode, of which changed the impedance spectra. The electrochemical biosensor has a LOD of 1 fg mL^{-1} M1 protein in saliva buffer within 5 min, per sample which corresponds to 5–10 virus particles. Besides, the assay has been verified by applying into different strains of influenza A virus. Meanwhile, as label-free electrochemical biosensors need more simple sensing protocol, they have been integrated with portable devices. Singh et al. [88] reported a novel label-free RGO-modified electrochemical immunosensor, cooperated with a microfluidic platform for influenza A H1N1 virus detection (Fig. 5). The three microelectrodes were fabricated on the glass substrate, then modified with RGO and mAb, and encapsulated with a polydimethylsiloxane (PDMS) microchannel finally. The amino groups on antibodies could form the direct linkage with amounts of carboxyl groups on RGO surface in absence of linker or spacer. Moreover, the large surface area of RGO presents lots of defects and electroactive sites, hence improving the sensitivity. The microfluidic label-free immunosensor presented excellent linear range of 1 to 10^4 PFU mL^{-1} and improved LOD (0.5 PFU mL^{-1}), exhibiting the potential of being handheld multianalyte sensing devices for clinical diagnosis. Label-free methods do not integrate any amplification step which could limit their sensitivity.

4.2.2. Label-based immunosensors

Sandwich is the common antibody format in the respiratory virus detection with label-based immunosensors. The detected antigen is sandwiched between two antibodies, one of which are attached on the transducer surface, called capture antibody. The other one is the detection antibody, which is usually labeled with enzyme, nanomaterials or biotin, it can directly measure the amount of antigen. The dual-recognition consolidate the specificity of the biosensors and own better label availabilities. The pAb and mAb are the most frequently used antibody combination in the sandwich-based immunosensors. Owing to the capture antibody will be attached to the electrode, the multi-site binding of antibody and antigen is restricted, so if pAb served as capture antibody, the




Fig. 4. Schematic representation of SARS-CoV-2 detection using the electrochemical biosensor. (A) Prepare of premix A and B; (B) Process of electrochemical detection using a smartphone. Reproduced with permission from Ref. [79].

advantage of high affinity cannot be exerted well. In addition, some pAb may occupy the epitope of mAb, resulting in less binding amount of detection antibody. Therefore, the mAb is often as capture antibody, and pAb is as detection antibody. For instance, Wu et al. [99] according to the ELISA designed the ultrasensitive electrochemical biosensors for H7N9 virus counting. The microelectrode array was modified with Au NPs and mAb, the MNPs decorated with Fe_3O_4 nanoparticles and quantum dots were incubated with pAb and ALP, forming bifunctional fluorescence magnetic nanospheres (bi-FMNs). The fact that pAb could conjugated with modified MNPs was supported by the color change of the fluorescence. Firstly, a single virus could be separated from the complex samples by one bi-FMNs at most, which is controllable by

the proportion of bi-FMNs to virus concentration. Then abundant complexes were transferred into the electrolyte, captured by the mAb modified microelectrode assay. Because the ALP on the bi-FMNs can catalyze the dephosphorylation of p-aminophenyl phosphate monohydrate (p-APP) to produce p-aminophenol (p-AP), hence inducing the reduction from Ag^+ to Ag^0 on the sensing interface. The changes from the Ag deposition could be recorded by linear sweep voltammetry. Finally, signals are counted as “0” or “1” depending on digital analysis, the virus concentrations could be estimated through the probability of “0”. The LOD of the label-based immunosensor was 7.8 fg mL^{-1} , which was 1–3 orders of magnitude more sensitive than previous research. Not all sandwich-based immunosensors use pAb and mAb as receptors,

Table 3
Antibody-based electrochemical biosensors for respiratory virus detection.

Type	Label	Virus	Recognition element	Linear range	LOD	Assay time	Electrochemical method	Ref
Label-free	/	H1N1, H3N2	Anti-M1 antibody	/	50 fg mL ⁻¹	0.1 h	EIS	[85]
	/	MERS-CoV, HCoV	Anti-recombinant spike protein S1 antibody	MERS-CoV: 1.0 pg mL ⁻¹ HCoV: 0.4 pg mL ⁻¹	MERS-CoV: 0.001 –100 ng mL ⁻¹ HCoV: 0.01 –10,000 ng mL ⁻¹	20 min	SWV	[86]
	/	H5N1	scFv against HA H5	The short fragment: 0.6 pg mL ⁻¹ The long fragment: 0.9 pg mL ⁻¹	The short fragment: 4.0 –20 pg mL ⁻¹ The long fragment: 1.0 –8 pg mL ⁻¹	/	SWV	[87]
	/	H1N1	mAb	0.5 PFU mL ⁻¹	1 – 10 ⁴ PFU mL ⁻¹	/	Chronoamperometry	[88]
	/	AIV H7	H7-mAb and H7-pAb	1.6 pg mL ⁻¹	1.6 pg mL ⁻¹ – 16 ng mL ⁻¹	/	LSV	[89]
	/	H1N1	Goat anti-influenza A antibody	113 PFU mL ⁻¹	10 – 10 ⁴ PFU mL ⁻¹	30 min	DPV	[90]
	/	Influenza virus	Anti-M1 antibody	1 fg mL ⁻¹ in saliva buffer	/	5 min	EIS	[91]
	/	H5N1, H1N1	mAb against the HA proteins	H5N1: 9.4 pM H1N1: 8.3 pM	25–500 pM	1 min	Chronoamperometry	[92]
	/	H1N1	Anti-H1N1 antibody	Phosphate-buffered saline: 26.04 PFU mL ⁻¹ diluted saliva: 33.11 PFU mL ⁻¹	10–10 ⁴ PFU mL ⁻¹	/	EIS	[93]
Label-based	MNP	H9N2	Anti-M2 antibody	8–128 HAU	8 HAU	160 s	Chronoamperometry	[94]
	HRP	H1N1, H5N1 and H7N9	Anti-H1N1, H5N1 and H7N9 antibodies	1 pg mL ⁻¹ – 10 ng mL ⁻¹	1 pg mL ⁻¹	/	Amperometry	[95]
	MNP	H5N1	Anti-H5N1 antibody	0.0025–0.16 HAU	0.0022 HAU in 6 µL	/	CV	[96]
	HRP	H1N1	Anti-influenza A HA antibody	/	5 PFU mL ⁻¹ for saliva samples	6 min	EIS	[97]
	MNP	H7N9	mAb and biotinylated antibody	0.011 ng mL ⁻¹	0.02–50 ng mL ⁻¹	1.5 h	LSV	[98]
	Fluorescence MNP	H7N9	mAb and pAb	7.8 fg mL ⁻¹	0.01–1.5 pg mL ⁻¹	/	LSV	[99]
	MNP	H7N9	mAb and rabbit derived pAb	6.8 pg mL ⁻¹	0.01–20 ng mL ⁻¹	/	LSV	[100]

other bio-recognition elements are also suitable for sandwich format. Sayhi et al. [94] employed anti-Matrix protein 2 (M2) antibody attached to MNPs and fetuin modified with Au NPs for electrochemical detection of H9N2 virus, the sandwich conformation was finally separated from real samples by applying a permanent magnetic field (Fig. 6). After the treatment in acid solution, the sandwich conformation was destroyed, the MNPs were removed by magnet. Because Au NPs can catalyze the hydrogen ions reduction in acidic medium under an appropriate potential, the Au NPs were deposited on the electrode and generated current signals, which was also proportional to the virus titer. The proposed immunosensor displayed the linear relationship between the virus titer in range 8–128 hemagglutination unit (HAU) and cathodic current, with LOD of less than 16 HAU titer. Although the LOD is higher than already published immunosensors, the approach with short detection time leaves out pretreatment steps and overcomes the difficulty of the virus separation from the bulk phase. Generally speaking, the sandwich-based immunosensors are of high sensitivity, high specificity, whose antigen without prior purification. Undeniably, the label-based detection procedures are time-consuming and an antigen must have at least two antibody binding sites.

4.3. Other affinity biosensors

Except for nucleic acid, antibodies, there have been other kinds of bio-recognition element presented in the electrochemical biosensors for respiratory virus detection: fetuin A, peptides and glycan.

4.3.1. Fetuin A

Fetuin A is a kind of glycoprotein derived from fetal calf serum, every fetuin A has terminal 12–14 sialic acid residues. Fetuin A is diffusely cooperated with peanut agglutinin (PNA) lectin [101]. Owing to the fact that fetuin A could combine with different influenza virus via NA protein, it could serve as bio-recognition element in influenza A detection with lower cost and high selectivity. For example, Anik et al. [102] developed an electrochemical biosensor based on graphene-Au hybrid nanocomposite for recognizing influenza A. The biosensor utilized fetuin A as bio-recognition element: Firstly, the fetuin A was immobilized onto the electrode surface for NA protein capture, and PNA specific binding sites would display after the interaction, then washed the NA protein on the SPEs, because the sugars from fetuin A have been masked by NA protein, the PNA lectin hence could bond to the N-acetylgalactosamine galactose-(Gal β1-3GalNAc). The resistance changes on the electrode surface were recorded by electrochemical impedance spectroscopy. The biosensor has a linear range between 10⁻⁸ U mL⁻¹ and 10⁻¹ U mL⁻¹, which has been applied into H9N2 detection in real samples. Besides, the biosensor's LOD of 10⁻⁸ U mL⁻¹ is lower than LOD values of ELISA assays relying on NA activity or antibody-antigen interaction.

4.3.2. Peptides

Easily being designed and prepared, peptides are theoretically favorable for antigens and drugs measurement. Previous studies showed the pentapeptide Ala-Arg-Leu-Pro-Arg is available to combine with the binding sites of all kinds of HA protein [103]. Surely, the corresponding N-stearoyl derivatives and carbosilane-




Fig. 5. Schematic illustration of the microfluidics-integrated electrochemical immunosensing chip coated with RGO, followed by antibody immobilization using EDC/NHS coupling for the detection of influenza virus H1N1. Reproduced with permission from Ref. [88].

based dendrimers could inhibit the activity of seasonal H1N1 and H3N2 except for H1 and H3 HAs [104]. Therefore, Matsubara et al. [105] modified the BDD electrode with a sialyloligosaccharide receptor-mimic peptide, the density of the peptide and dendrimer generation terminated on the electrode could affect the probability that the respiratory virus were captured by the functionalized electrode. Electrochemical impedance spectroscopy was used for the virus identification according to the resistance variation. The proposed electrochemical biosensor could isolate the avian virus particles from H5N3, H7N1 and H9N2, presenting the satisfactory specificity and practicability. Faced with the antigenic drift and new subtypes of the respiratory, the designed peptide dendrimer has great potential as antibodies candidates. Besides, Tara Bahadur et al. [106] developed an electrochemical biosensor toward influenza virus particles based on the selection of electrosensitive peptide ligand *in vitro* (Fig. 7). The electrochemically sensitive 3,4-ethylenedioxythiophene (EDOT) moiety was modified with a peptide ligand then worked as electro-polymerization monomers. In the scheme, the real samples were mixed with the solution including the peptide ligand-EDOT monomers. The presentation of the virus particles would influence the electro-polymerization of the peptide ligand-EDOT monomers on the electrode surface, consequently affecting the efficiency of the electron transfer between the redox molecules and the electrode. The LOD of the detection system was found to be $12.5 \mu\text{g mL}^{-1}$, which is 2.5-fold more sensitive than the dot blot immune assay or conventional rapid diagnosis test. The “turn-on system”: the current increases when there is influenza virus doesn’t need negative control measurement for practical application.

4.3.3. Glycans

Glycans are a kind of complicated carbohydrates which usually form the dense sugar layer on the numerous cell surface. The cell-cell recognition and host-pathogen interactions are both realized through the glycan coat [107]. For instance, in the influenza A virus infection process, HA protein interact with host glycans terminated in sialic acid firstly. Compared with DNA and proteins, the glycans could store more code information as there are over 10 million glycan molecules on the cell surface [108]. The function of glycan bio-recognition has been applied into the development of the diagnosis approaches and vaccines design. Hushegyi et al. [109] utilized glycans as natural viral receptors in the impedimetric biosensor design for inactivated, but intact influenza virus H3N2 detection. The gold electrode surface functionalized with thiols bearing oligoethylene glycol moieties formed a mixed SAM layer (self-assembled monolayer) for glycan immobilization. The biosensor could detect at least 13 virus particles in $1 \mu\text{L}$ real samples, revealing a LOD of 5 aM. It was the lowest LOD for influenza virus detection compared with published glycan-based electrochemical biosensors at that time. However, the application of glycans is limited to a few respiratory viruses, and the affinities of glycans need to be proved further.

5. Future challenges

The global health crisis of the COVID-19 pandemic defines the greatest challenge the world is faced with at the present time, with the most important focus being the sensitivity and specificity enhancement, to which current innovations should pay attention




Fig. 6. (A) Schematic illustration of the strategy used to develop the gold nanoparticle-based chronoamperometric magneto-immunosensor for influenza virus detection. The influenza virus could be recognized by anti-Matrix protein 2 (M2) antibody modified magnetic nanomaterials (MNP) and fetuin decorated Au NPs. (B) Chronoamperometric curves obtained without influenza virus (Allantoic fluid) and with 8; 16; 32; 64 and 128 hemagglutinin Units (HAU) of the virus (upper panel). (C) Diagrams (lower panel) correspond to the response of the magneto immunoassay to various influenza virus titers ranging from 8 HAU to 128 HAU (blue) and to various concentration of non-infected allantoic fluid in 1 M HCl solution (red). SPCE: Screen-printed carbon electrode. Reproduced with permission from Ref. [94].




Fig. 7. Strategy for detection of influenza virus using an electrosensitive peptide ligand. Reproduced with permission from Ref. [106].

for early detection of COVID-19 disease or future pandemic strains. Simple, low cost, easy to operate and fast-response electrochemical biosensors exactly meet the potential to be integrated into POCTs for COVID-19 diagnosis. Although efforts have been put to design electrochemical biosensors for COVID-19 diagnosis, few portable electrochemical biosensors were produced. There still exists numerous challenges to move from the bench to their use in POCTs.

5.1. The sample preparation

A large number of interferers, such as proteins, antibodies, DNA, cells, etc. in various complex samples can disrupt the detection process of the targets. The sample pretreatment requirement before analysis to exclude the influence of matrix effects is a main impact factor for specificity and sensitivity improvement. So how to isolate the viruses from the real samples is the key step during the sample preparation. The viruses usually only occupy a small volume of the whole volume, so there is always a small possibility for virus to be captured by the receptor on the transducer when the whole volume is very small. Obviously, the viruses couldn't be concentrated without any preparation. At present, the use of magnetic nanoparticles and selecting most perfect bio-recognition elements are the two main approaches for solving the problem. For example, the specific receptors are coupled with MNPs to capture and separate the targets from complex sample; the association constant of the antibody should be maximized during the antibody selection.

5.2. The immobilization of the bio-recognition elements

The immobilization process of the bio-recognition elements is vital to reduce mistakes and errors during virus detection. Currently, the key recognition interaction in many electrochemical biosensors is often irreversible, hence the initial properties couldn't be restored after every detection, the biosensor part should be disposable, which is the rule for medical consumables. Moreover, during the modification process, the affinity of bio-recognition

elements is related to the immobilization process, and the efficiency of the immobilization can influence that of detection. How to ensure that the receptor distribution on the electrode surface is uniform and roughly the same between the same batch, without affecting the efficiency of receptor recognition is the question for most portable electrochemical biosensors to be considered.

5.3. The miniaturization of the system

Basically, a whole research process of the available electrochemical biosensor-based POCTs includes: optimization of the operation condition, integrating a sensing chip with micro-/nano-electronics, and interfacing of the sensing platform with a wireless device, transforming into on-site analytical devices, big data analytics and result output [79]. The development of the whole smart sensing system is a multidisciplinary project and need public-private participation. POCTs aim to be carried out close to the patients, the sample volumes, reagent use, transducer and power all need to be miniaturized without reducing the current density and transfer characteristics, and the whole system needs to be wearable and wireless. Most of the electrochemical biosensors own excellent properties and could easily be miniaturized and then should be associated with the whole system.

5.4. The reproducibility and stability

To guarantee the accuracy of the POCTs, the reproductivity and stability of electrochemical biosensors should be improved dramatically. In a whole fabrication process of the electrochemical biosensor, there are many influencing factors: environmental conditions, operating procedures, performance of the instrument. Among these, the most difficult to be automatized is the manual steps for the preparation of the biosensor. Besides, the stability of the POCTs is also supposed to be excellent, because the storage conditions are often difficult to achieve at the laboratory level, before its use.

5.5. Environment-friendly and the cost

The environment-friendly and cheap POCTs are often the last hurdle before a biosensor is implemented for POCTs. With the development of material science, numerous nanomaterials have been introduced into the electrochemical biosensors. The potential health impacts and environmental pollution from the widespread usage of the nanomaterials could not be ignored. Besides, the cost of the POCTs should be affordable for primary medical institutions. Therefore, the materials used in the equipment manufacturing process should be as low-cost as possible meanwhile without affecting the performance. Now, paper-based microfluidic devices are relatively environmentally friendly and low-cost, therefore having been the most frequently used substrate platform. Carbon-based nanomaterials are also the excellent green alternative with less pollution.

It is evident that great effort is still required to overtake above challenges in the portable electrochemical biosensor design for SARS-CoV-2 POCTs detection, but we still believe that with the increasing trend in multidisciplinary integration, the ideal POCTs for COVID-19 diagnosis will be produced just around the corner.

6. Conclusion

Overall, we have presented the common detecting targets of the respiratory viruses, key parts of electrochemical biosensors design and discussed different bio-recognition element-based electrochemical biosensors. Future challenges in electrochemical biosensors for respiratory virus determination, especially for application in POCTs are discussed. In every section, several examples were explained, and all the analytical performance of recent developments are gathered in tables with their detection limits. We believe that the advancements from core technologies at multiple-disciplines areas will offer great potential of a next generation of highly specific, sensitive, selective, and reliable electrochemical biosensors for respiratory virus detection. More urgently, the developed electrochemical biosensors could make for better surveillance and control of SARS-CoV-2 infection in populations.

Declaration of competing interest

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper.

Acknowledgments

This work was supported by the key research and development project of Hubei province, China (NO. 2020BCB022), High-level foreign expert introduction program, China (No. G20200227010). CNRS is acknowledged for support through IRN MAREES # 0876.

References

- [1] A. Parihar, P. Ranjan, S.K. Sanghi, A.K. Srivastava, R. Khan, Point-of-Care biosensor-based diagnosis of COVID-19 holds promise to combat current and future pandemics, *ACS Appl. Bio Mater.* (2020). <https://doi.org/10.1021/acsbm.0c01083>.
- [2] N. Zhang, L. Wang, X. Deng, R. Liang, M. Su, C. He, L. Hu, Y. Su, J. Ren, F. Yu, L. Du, S. Jiang, Recent advances in the detection of respiratory virus infection in humans, *J. Med. Virol.* 92 (2020) 408–417. <https://doi.org/10.1002/jmv.25674>.
- [3] M. Khoury, J. Cuenca, F.F. Cruz, F.E. Figueroa, P.R.M. Rocco, D.J. Weiss, Current status of cell-based therapies for respiratory virus infections: applicability to COVID-19, *Eur. Respir. J.* 55 (2020). <https://doi.org/10.1183/13993003.00858-2020>.
- [4] M.M. Azar, L. Landry, Detection of influenza A and B viruses and respiratory, *J. Clin. Microbiol.* 1988 (2018) 1–13.
- [5] <https://www.who.int/emergencies/diseases/novel-coronavirus-2019>.
- [6] M.G. Ison, H.H. Hirsch, Community-acquired respiratory viruses in transplant patients: diversity, impact, unmet clinical needs, *Clin. Microbiol. Rev.* 32 (2019) 1–33. <https://doi.org/10.1128/CMR.00042-19>.
- [7] J.M. Heraud, N.H. Razanajatovo, C. Viboud, Global circulation of respiratory viruses: from local observations to global predictions, *The Lancet Glob. Health* 7 (2019) e982–e983. [https://doi.org/10.1016/S2214-109X\(19\)30277-3](https://doi.org/10.1016/S2214-109X(19)30277-3).
- [8] Y. Pan, D. Zhang, P. Yang, L.L.M. Poon, Q. Wang, Viral load of SARS-CoV-2 in clinical samples, *Lancet Infect. Dis.* 20 (2020) 411–412. [https://doi.org/10.1016/S1473-3099\(20\)30113-4](https://doi.org/10.1016/S1473-3099(20)30113-4).
- [9] J. Broughton, X. Deng, G. Yu, C. Fasching, J. Singh, J. Streithorst, A. Granados, A. Sotomayor-Gonzalez, K. Zorn, A. Gopez, E. Hsu, W. Gu, S. Miller, C.-Y. Pan, H. Guevara, D. Wadford, J. Chen, C. Chiu, Rapid detection of 2019 novel coronavirus SARS-CoV-2 using a CRISPR-based DETECTR lateral flow assay, *MedRxiv: Preprint Server Health Science* (2020) 1–28. <https://doi.org/10.1101/2020.03.06.20032334>.
- [10] A. Sakurai, K. Takayama, N. Nomura, N. Kajiura, M. Okamatsu, N. Yamamoto, T. Tamura, J. Yamada, M. Hashimoto, Y. Sakoda, Y. Suda, Y. Kobayashi, H. Kida, F. Shibasaki, Fluorescent immunochromatography for rapid and sensitive typing of seasonal influenza viruses, *PLoS One* 10 (2015). <https://doi.org/10.1371/journal.pone.0116715>.
- [11] H. Tani, M. Kimura, L. Tan, Y. Yoshida, T. Ozawa, H. Kishi, S. Fukushi, M. Saijo, K. Sano, T. Suzuki, H. Kawasui, A. Ueno, Y. Miyajima, Y. Fukui, I. Sakamaki, Y. Yamamoto, Y. Morinaga, Evaluation of SARS-CoV-2 neutralizing antibodies using a vesicular stomatitis virus possessing SARS-CoV-2 spike protein, *Virology* 18 (2021) 16. <https://doi.org/10.1186/s12985-021-01490-7>.
- [12] Q.X. Long, B.Z. Liu, H.J. Deng, G.C. Wu, K. Deng, Y.K. Chen, P. Liao, J.F. Qiu, Y. Lin, X.F. Cai, D.Q. Wang, Y. Hu, J.H. Ren, N. Tang, Y.Y. Xu, L.H. Yu, Z. Mo, F. Gong, X.L. Zhang, W.G. Tian, L. Hu, X.X. Zhang, J.L. Xiang, H.X. Du, H.W. Liu, C.H. Lang, X.H. Luo, S.B. Wu, X.P. Cui, Z. Zhou, M.M. Zhu, J. Wang, C.J. Xue, X.F. Li, L. Wang, Z.J. Li, K. Wang, C.C. Niu, Q.J. Yang, X.J. Tang, Y. Zhang, X.M. Liu, J.J. Li, D.C. Zhang, F. Zhang, P. Liu, J. Yuan, Q. Li, J.L. Hu, J. Chen, A.L. Huang, Antibody responses to SARS-CoV-2 in patients with COVID-19, *Nat. Med.* 26 (2020) 845–848. <https://doi.org/10.1038/s41591-020-0897-1>.
- [13] E. Cesewski, B.N. Johnson, Electrochemical biosensors for pathogen detection, *Biosens. Bioelectron.* 159 (2020) 112214. <https://doi.org/10.1016/j.bios.2020.112214>.
- [14] I.S. Kucherenko, Y.v. Topolnikova, O.O. Soldatkin, Advances in the biosensors for lactate and pyruvate detection for medical applications: a review, *Trends Anal. Chem.* 110 (2019) 160–172. <https://doi.org/10.1016/j.trac.2018.11.004>.
- [15] T.J. Park, M.S. Hyun, H.J. Lee, et al., A self-assembled fusion protein-based surface plasmon resonance biosensor for rapid diagnosis of severe acute respiratory syndrome, *Talanta* 79 (2009) 295–301. <https://doi.org/10.1016/j.talanta.2009.03.051>.
- [16] M. Soler, M.C. Estevez, M. Cardenas-Rubio, A. Astua, L.M. Lechuga, How nanophotonic label-free biosensors can contribute to rapid and massive diagnostics of respiratory virus infections: COVID-19 case, *ACS Sens.* 5 (2020) 2663–2678. <https://doi.org/10.1021/acssens.0c01180>.
- [17] L. Zhan, S.J. Zhen, X.Y. Wan, P.F. Gao, C.Z. Huang, A sensitive surface-enhanced Raman scattering enzyme-catalyzed immunoassay of respiratory syncytial virus, *Talanta* 148 (2016) 308–312. <https://doi.org/10.1016/j.talanta.2015.10.081>.
- [18] F. Cui, H.S. Zhou, Diagnostic methods and potential portable biosensors for coronavirus disease 2019, *Biosens. Bioelectron.* 165 (2020) 112349. <https://doi.org/10.1016/j.bios.2020.112349>.
- [19] J.R. Anusha, B.C. Kim, K.H. Yu, C.J. Raj, Electrochemical biosensing of mosquito-borne viral disease, dengue: a review, *Biosens. Bioelectron.* 142 (2019) 111511. <https://doi.org/10.1016/j.bios.2019.111511>.
- [20] A. Kaushik, S. Tiwari, R.D. Jayant, A. Vashist, R. Nikkhah-Moshaie, N. El-Hage, M. Nair, Electrochemical biosensors for early stage Zika diagnostics, *Trends Biotechnol.* 35 (2017) 308–317. <https://doi.org/10.1016/j.tibtech.2016.10.001>.
- [21] E. Rasouli, Z. Shahnava, W.J. Basirun, M. Rezayi, A. Avan, M. Ghayour-Mobarhan, R. Khandanlou, M.R. Johan, Advancements in electrochemical DNA sensor for detection of human papilloma virus - a review, *Anal. Biochem.* 556 (2018) 136–144. <https://doi.org/10.1016/j.ab.2018.07.002>.
- [22] B.V. Ribeiro, T.A.R. Cordeiro, G.R. Oliveira e Freitas, L.F. Ferreira, D.L. Franco, Biosensors for the detection of respiratory viruses: a review, *Talanta Open* 2 (2020) 100007. <https://doi.org/10.1016/j.talo.2020.100007>.
- [23] R. Samson, G.R. Navale, M.S. Dhame, Biosensors: frontiers in rapid detection of COVID-19, *3 Biotech* 10 (2020) 385. <https://doi.org/10.1007/s13205-020-02369-0>.
- [24] M.R. de Eguilaz, L.R. Cumba, R.J. Forster, Electrochemical detection of viruses and antibodies: a mini review, *Electrochem. Commun.* 116 (2020) 106762. <https://doi.org/10.1016/j.elecom.2020.106762>.
- [25] P.P. Nelson, B.A. Rath, P.C. Fragkou, E. Antalis, S. Tsioudras, C. Skevaki, Current and future point-of-care tests for emerging and new respiratory viruses and future perspectives, *Front. Cell. Infect. Microbiol.* 10 (2020). <https://doi.org/10.3389/fcimb.2020.00181>.
- [26] Y. Li, R.M. Reeves, X. Wang, Q. Bassat, W.A. Brooks, C. Cohen, D.P. Moore, M. Nunes, B. Rath, H. Campbell, H. Nair, S. Acacio, W.J. Alonso, M. Antonio, G. Ayora Talavera, D. Badarch, V.L. Baillie, G. Barrera-Badillo, G. Bigogo, S. Broor, D. Bruden, P. Buchy, P. Byass, J. Chipeta, W. Clara, D.A. Dang, C.C. de

- Freitas Lázaro Emadiato, M. de Jong, J.A. Díaz-Quinonez, L.A.H. Do, R.A. Fasce, L. Feng, M.J. Ferson, A. Gentile, B.D. Gessner, D. Goswami, S. Goyet, C.G. Grijalva, N. Halasa, O. Helffersee, D. Hessong, N. Homaira, J. Jara, K. Kahn, N. Khuri-Bulos, K.L. Kottloff, C.F. Lanata, O. Lopez, M.R. Lopez Bolaños, M.G. Lucero, F. Lucion, S.P. Lupisan, S.A. Madhi, O. Mekgoe, C. Moraleda, J. Moyes, K. Mulholland, P.K. Munywoki, F. Naby, T.H. Nguyen, M.P. Nicol, D.J. Nokes, D.E. Noyola, D. Onozuka, N. Palani, Y. Poovorawan, M. Rahman, K. Ramaekers, C. Romero, E.P. Schlaudecker, B. Schweiger, P. Seidenberg, E.A.F. Simoes, R. Singleton, S. Sistla, K. Sturm-Ramirez, N. Suntronwong, A. Sutanto, M.D. Tapia, S. Thamthitawat, I. Thongpan, G. Tillekeratne, Y.O. Tinoco, F.K. Treurnicht, C. Turner, P. Turner, R. van Doorn, M. van Ranst, B. Visseaux, S. Waicharoen, J. Wang, L.M. Yoshida, H.J. Zar, Global patterns in monthly activity of influenza virus, respiratory syncytial virus, parainfluenza virus, and metapneumovirus: a systematic analysis, *The Lancet Glob. Health* 7 (2019) e1031–e1045. [https://doi.org/10.1016/S2214-109X\(19\)30264-5](https://doi.org/10.1016/S2214-109X(19)30264-5).
- [27] E.C. Hutchinson, Influenza virus, *Trends Microbiol.* 26 (2018) 809–810. <https://doi.org/10.1016/j.tim.2018.05.013>.
- [28] Y.A. Malik, Properties of Coronavirus and SARS-CoV-2, (n.d).
- [29] B. Udugama, P. Kadhiresan, H.N. Kozlowski, A. Malekjahani, M. Osborne, V.Y.C. Li, H. Chen, S. Mubareka, J.B. Gubbay, W.C.W. Chan, Diagnosing COVID-19: the disease and tools for detection, *ACS Nano* 14 (2020) 3822–3835. <https://doi.org/10.1021/acsnano.0c02624>.
- [30] X. Yuan, C. Yang, Q. He, J. Chen, D. Yu, J. Li, S. Zhai, Z. Qin, K. Du, Z. Chu, P. Qin, Current and perspective diagnostic techniques for COVID-19, *ACS Infect. Dis.* 6 (2020) 1998–2016. <https://doi.org/10.1021/acscinfed.0c00365>.
- [31] P. Pokhrel, C. Hu, H. Mao, Detecting the coronavirus (CoVID-19), *ACS Sens.* 5 (2020) 2283–2297. <https://doi.org/10.1021/ACSENSORS.0C01153>.
- [32] N. Bhalla, Y. Pan, Z. Yang, A.F. Payam, Opportunities and challenges for biosensors and nanoscale Analytical tools for pandemics: COVID-19, *ACS Nano* 14 (2020) 7783–7807. <https://doi.org/10.1021/acsnano.0c04421>.
- [33] J.P. Tosar, G. Brañas, J. Laiz, Electrochemical DNA hybridization sensors applied to real and complex biological samples, *Biosens. Bioelectron.* 26 (2010) 1205–1217. <https://doi.org/10.1016/j.bios.2010.08.053>.
- [34] H. Duvensee, M. Mix, M. Stubbe, J. Gimsa, M. Adler, G.U. Flechsig, Electrochemical product detection of an asymmetric convective polymerase chain reaction, *Biosens. Bioelectron.* 25 (2009) 400–405. <https://doi.org/10.1016/j.bios.2009.07.025>.
- [35] Y. Yan, L. Chang, L. Wang, Laboratory testing of SARS-CoV, MERS-CoV, and SARS-CoV-2 (2019-nCoV): current status, challenges, and countermeasures, *Rev. Med. Virol.* 30 (2020). <https://doi.org/10.1002/rmv.2106>.
- [36] A. Zumla, S. Hui, S. Perlman, Middle East respiratory syndrome, *Lancet* 386 (2015) 995–1007. [https://doi.org/10.1016/S0140-6736\(15\)60454-8](https://doi.org/10.1016/S0140-6736(15)60454-8).
- [37] X. Lu, B. Whitaker, S.K.K. Sakthivel, S. Kamili, L.E. Rose, L. Lowe, E. Mohareb, E.M. Elassal, T. Al-sanouri, A. Haddadin, D.D. Erdman, Real-time reverse transcription-pcr assay panel for middle east respiratory syndrome coronavirus, *J. Clin. Microbiol.* 52 (2014) 67–75. <https://doi.org/10.1128/JCM.02533-13>.
- [38] V.M. Corman, O. Landt, M. Kaiser, R. Molenkamp, A. Meijer, D.K.W. Chu, T. Bleicker, S. Brünink, J. Schneider, M.L. Schmidt, D.G.J.C. Mulders, B.L. Haagmans, B. van der Veer, S. van den Brink, L. Wijsman, G. Goderski, J.L. Romette, J. Ellis, M. Zambon, M. Peiris, H. Goossens, C. Reusken, M.P.G. Koopmans, C. Drosten, Detection of 2019 novel coronavirus (2019-nCoV) by real-time RT-PCR, *Euro Surveill.* 25 (2020). <https://doi.org/10.2807/1560-7917.ES.2020.25.3.2000045>.
- [39] G.S. Park, K. Ku, S.H. Baek, S.J. Kim, S. il Kim, B.T. Kim, J.S. Maeng, Development of reverse transcription loop-mediated isothermal amplification assays targeting severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2), *J. Mol. Diagn.* 22 (2020) 729–735. <https://doi.org/10.1016/j.jmoldx.2020.03.006>.
- [40] S. Patris, M. Vandeput, J.M. Kauffmann, Antibodies as target for affinity biosensors, *Trac. Trends Anal. Chem.* 79 (2016) 239–246. <https://doi.org/10.1016/j.trac.2015.12.005>.
- [41] S.K. Arya, P. Kongsuphol, C.C. Wong, L.J. Polla, M.K. Park, Label free biosensor for sensitive human influenza virus hemagglutinin specific antibody detection using coiled-coil peptide modified microelectrode array based platform, *Sensor. Actuator. B Chem.* 194 (2014) 127–133. <https://doi.org/10.1016/j.snb.2013.12.066>.
- [42] M.S. Ho, W.J. Chen, H.Y. Chen, et al., Neutralizing antibody response and SARS severity, *Emerg. Infect. Dis.* 11 (11) (2005) 1730–1737. <https://doi.org/10.3201/eid1111.040659>.
- [43] P.C.Y. Woo, S.K.P. Lau, B.H.L. Wong, H.W. Tsoi, A.M.Y. Fung, R.Y.T. Kao, K.H. Chan, J.S.M. Peiris, K.Y. Yuen, Differential sensitivities of severe acute respiratory syndrome (SARS) coronavirus spike polypeptide enzyme-linked immunosorbent assay (ELISA) and SARS coronavirus nucleocapsid protein ELISA for serodiagnosis of SARS coronavirus pneumonia, *J. Clin. Microbiol.* 43 (2005) 3054–3058. <https://doi.org/10.1128/JCM.43.7.3054-3058.2005>.
- [44] K.K.W. To, O.T.Y. Tsang, C.C.Y. Yip, K.H. Chan, T.C. Wu, J.M.C. Chan, W.S. Leung, T.S.H. Chik, C.Y.C. Choi, D.H. Kandamby, D.C. Lung, A.R. Tam, R.W.S. Poon, A.Y.F. Fung, I.F.N. Hung, V.C.C. Cheng, J.F.W. Chan, K.Y. Yuen, Consistent detection of 2019 novel coronavirus in saliva, *Clin. Infect. Dis.* 71 (2020) 841–843. <https://doi.org/10.1093/cid/cia149>.
- [45] J. Bhardwaj, N. Chaudhary, H. Kim, J. Jang, Subtyping of influenza A H1N1 virus using a label-free electrochemical biosensor based on the DNA aptamer targeting the stem region of HA protein, *Anal. Chim. Acta* 1064 (2019) 94–103. <https://doi.org/10.1016/j.aca.2019.03.005>.
- [46] X. Xie, Z. Zhong, W. Zhao, C. Zheng, F. Wang, J. Liu, Chest CT for typical coronavirus disease 2019 (COVID-19) pneumonia: relationship to negative RT-PCR testing, *Radiology* 296 (2020) E41–E45. <https://doi.org/10.1148/radiol.2020200343>.
- [47] J. Riu, B. Giussani, Electrochemical biosensors for the detection of pathogenic bacteria in food, *Trac. Trends Anal. Chem.* 126 (2020) 115863. <https://doi.org/10.1016/j.trac.2020.115863>.
- [48] R. Antiochia, Developments in biosensors for CoV detection and future trends, *Biosens. Bioelectron.* 173 (2021) 112777. <https://doi.org/10.1016/j.bios.2020.112777>.
- [49] S. Tanisellass, M.K.M. Arshad, S.C.B. Gopinath, Graphene-based electrochemical biosensors for monitoring noncommunicable disease biomarkers, *Biosens. Bioelectron.* 130 (2019) 276–292. <https://doi.org/10.1016/j.bios.2019.01.047>.
- [50] A. Valizadeh, N. Sohrabi, F. Badrzadeh, Electrochemical detection of HIV-1 by nanomaterials, artificial cells, Nanomed. *Biotechnol.* 45 (2017) 1467–1477. <https://doi.org/10.1080/21691401.2017.1282494>.
- [51] H. Zhang, B.L. Miller, Immunosensor-based label-free and multiplex detection of influenza viruses: state of the art, *Biosens. Bioelectron.* 141 (2019). <https://doi.org/10.1016/j.bios.2019.111476>.
- [52] Z.A. Ahovan, A. Hashemi, L.M. de Plano, M. Gholipourmalekabadi, A. Seifalian, Bacteriophage based biosensors: trends, outcomes and challenges, *Nanomaterials* 10 (2020). <https://doi.org/10.3390/nano10030501>.
- [53] T. Lee, J.H. Ahn, S.Y. Park, G.H. Kim, J. Kim, T.H. Kim, I. Nam, C. Park, M.H. Lee, Recent advances in AIV biosensors composed of nanobio hybrid material, *Micromachines* 9 (2018) 1–17. <https://doi.org/10.3390/mi9120651>.
- [54] J. Leva-Bueno, S.A. Peyman, P.A. Millner, A review on impedimetric immunosensors for pathogen and biomarker detection, *Med. Microbiol. Immunol.* 209 (2020) 343–362. <https://doi.org/10.1007/s00430-020-00668-0>.
- [55] L.C. Chen, E. Wang, C.S. Tai, Y.C. Chiu, C.W. Li, Y.R. Lin, T.H. Lee, C.W. Huang, J.C. Chen, W.L. Chen, Improving the reproducibility, accuracy, and stability of an electrochemical biosensor platform for point-of-care use, *Biosens. Bioelectron.* 155 (2020) 112111. <https://doi.org/10.1016/j.bios.2020.112111>.
- [56] E.O. Blair, D.K. Corrigan, A review of microfabricated electrochemical biosensors for DNA detection, *Biosens. Bioelectron.* 134 (2019) 57–67. <https://doi.org/10.1016/j.bios.2019.03.055>.
- [57] C.M.A. Brett, A.M. Oliveira Brett, *Electrochemistry: Principles, Methods and Applications*, Oxford Science Publications, 1993. ISBN 9780198553885.
- [58] T. Xiao, J. Huang, D. Wang, T. Meng, X. Yang, Au and Au-based nanomaterials: synthesis and recent progress in electrochemical sensor applications, *Talanta* 206 (2020) 120210. <https://doi.org/10.1016/j.talanta.2019.120210>.
- [59] L. Qian, S. Durairaj, S. Prins, A. Chen, Biosensors and Bioelectronics Nanomaterial-based electrochemical sensors and biosensors for the detection of pharmaceutical compounds, *Biosens. Bioelectron.* (2020) 112836. <https://doi.org/10.1016/j.bios.2020.112836>.
- [60] K. Wu, R. Saha, D. Su, V.D. Krishna, J. Liu, M.C.-J. Cheeran, J.-P. Wang, Magnetic-nanosensor-based virus and pathogen detection strategies before and during COVID-19, *ACS Appl. Nano Mater.* 3 (2020) 9560–9580. <https://doi.org/10.1021/acsnanm.0c02048>.
- [61] S. Li, T. Tian, T. Zhang, X. Cai, Y. Lin, Advances in biological applications of self-assembled DNA tetrahedral nanostructures, *Mater. Today* 24 (2019) 57–68. <https://doi.org/10.1016/j.mattod.2018.08.002>.
- [62] Y. Saylan, Ö. Erdem, S. Ünal, A. Denizli, An alternative medical diagnosis method: biosensors for virus detection, *Biosensors* 9 (2019). <https://doi.org/10.3390/bios9020065>.
- [63] H.L.L. Yu, A. Maslova, I.M. Hsing, Rational design of electrochemical DNA biosensors for point-of-care applications, *ChemElectroChem* 4 (2017) 795–805. <https://doi.org/10.1002/celc.201600756>.
- [64] T. Lee, S.Y. Park, H. Jang, G.H. Kim, Y. Lee, C. Park, M. Mohammadniaei, M.H. Lee, J. Min, Fabrication of electrochemical biosensor consisted of multi-functional DNA structure/porous au nanoparticle for avian influenza virus (H5N1) in chicken serum, *Mater. Sci. Eng. C* 99 (2019) 511–519. <https://doi.org/10.1016/j.msec.2019.02.001>.
- [65] I. Lee, S.E. Kim, J. Lee, D.H. Woo, S. Lee, H. Pyo, C.S. Song, J. Lee, A self-calibrating electrochemical aptasensing platform: correcting external interference errors for the reliable and stable detection of avian influenza viruses, *Biosens. Bioelectron.* 152 (2020) 112010. <https://doi.org/10.1016/j.bios.2020.112010>.
- [66] S. Dong, R. Zhao, J. Zhu, X. Lu, Y. Li, S. Qiu, L. Jia, X. Jiao, S. Song, C. Fan, R.Z. Hao, H. bin Song, Electrochemical DNA biosensor based on a tetrahedral nanostructure probe for the detection of avian influenza A (H7N9) virus, *ACS Appl. Mater. Interfaces* 7 (2015) 8834–8842. <https://doi.org/10.1021/acsami.5b01438>.
- [67] F.S. Diba, S. Kim, H.J. Lee, Amperometric bioaffinity sensing platform for avian influenza virus proteins with aptamer modified gold nanoparticles on carbon chips, *Biosens. Bioelectron.* 72 (2015) 355–361. <https://doi.org/10.1016/j.bios.2015.05.020>.
- [68] C. Bai, Z. Lu, H. Jiang, Z. Yang, X. Liu, H. Ding, H. Li, J. Dong, A. Huang, T. Fang, Y. Jiang, L. Zhu, X. Lou, S. Li, N. Shao, Aptamer selection and application in multivalent binding-based electrical impedance detection of inactivated H1N1 virus, *Biosens. Bioelectron.* 110 (2018) 162–167. <https://doi.org/10.1016/j.bios.2018.03.047>.
- [69] K. Malecka, A. Stachyra, A. Góra-Sochacka, A. Sirko, W. Zagórski-Ostoja, H. Radecka, J. Radecki, Electrochemical genosensor based on disc and screen printed gold electrodes for detection of specific DNA and RNA sequences

- derived from Avian Influenza Virus H5N1, *Sensor. Actuator. B Chem.* 224 (2016) 290–297. <https://doi.org/10.1016/j.snb.2015.10.044>.
- [70] K. Kurzatowska, A. Sirko, W. Zagórski-Ostojka, W. Dehaen, H. Radecka, J. Radecki, Electrochemical label-free and reagentless genosensor based on an ion barrier switch-off system for DNA sequence-specific detection of the avian influenza virus, *Anal. Chem.* 87 (2015) 9702–9709. <https://doi.org/10.1021/acs.analchem.5b01988>.
- [71] J. Mohammadi, A. Moattari, N. Sattarahmady, N. Pirbonyeh, H. Yadegari, H. Heli, Electrochemical biosensing of influenza A subtype genome based on meso/macroporous cobalt (II) oxide nanoflakes-applied to human samples, *Anal. Chim. Acta* 979 (2017) 51–57. <https://doi.org/10.1016/j.aca.2017.05.010>.
- [72] Ravina, H. Mohan, P.S. Gill, A. Kumar, Hemagglutinin gene based biosensor for early detection of swine flu (H1N1) infection in human, *Int. J. Biol. Macromol.* 130 (2019) 720–726. <https://doi.org/10.1016/j.ijbiomac.2019.02.149>.
- [73] J. Astill, E. Fraser, R. Dara, S. Sharif, Detecting and predicting emerging disease in poultry with the implementation of new technologies and big data: a focus on avian influenza virus, *Front. Veterin. Sci.* 5 (2018). <https://doi.org/10.3389/fvets.2018.00263>.
- [74] X. Fu, F. Peng, J. Lee, Q. Yang, F. Zhang, M. Xiong, G. Kong, H.M. Meng, G. Ke, X.B. Zhang, Aptamer-Functionalized DNA Nanostructures for Biological Applications, Springer International Publishing, 2020. https://doi.org/10.1007/978-1-4939-6911-1_25.
- [75] Y. Li, R. Wang, Aptasensors for detection of avian influenza virus H5N1, *Methods Mol. Biol.* 1572 (2017) 379–402. https://doi.org/10.1007/978-1-4939-6911-1_25.
- [76] M. Alafeef, K. Dighe, P. Moitra, D. Pan, Rapid, ultrasensitive, and quantitative detection of SARS-CoV-2 using antisense oligonucleotides directed electrochemical biosensor chip, *ACS Nano* 2019 (2020). <https://doi.org/10.1021/acsnano.0c06392>.
- [77] F. Nicolson, A. Ali, M.F. Kircher, S. Pal, DNA nanostructures and DNA-functionalized nanoparticles for cancer theranostics, *Adv. Sci.* (2020) 1–27. <https://doi.org/10.1002/advs.202001669>.
- [78] T. Zhou, R. Luo, Y. Li, J. Fan, Y. Hu, C. Tong, B. Liu, D. Li, Activity assay and intracellular imaging of APE1 assisted with tetrahedral DNA nanostructure modified-dnzyme and molecular beacon, *Sensor. Actuator. B Chem.* 317 (2020). <https://doi.org/10.1016/j.snb.2020.128203>.
- [79] H. Zhao, F. Liu, W. Xie, T.C. Zhou, J. Ouyang, L. Jin, H. Li, C.Y. Zhao, L. Zhang, J. Wei, Y.P. Zhang, C.P. Li, Ultrasensitive sandwich-type electrochemical sensor for SARS-CoV-2 from the infected COVID-19 patients using a smart-phone, *Sensor. Actuator. B Chem.* 327 (2021) 128899. <https://doi.org/10.1016/j.snb.2020.128899>.
- [80] V. Crivianu-Gaita, M. Thompson, Aptamers, antibody scFv, and antibody Fab' fragments: an overview and comparison of three of the most versatile biosensor biorecognition elements, *Biosens. Bioelectron.* 85 (2016) 32–45. <https://doi.org/10.1016/j.bios.2016.04.091>.
- [81] W. Chen, Y. Yuan, X. Jiang, Antibody and antibody fragments for cancer immunotherapy, *J. Contr. Release* 328 (2020) 395–406. <https://doi.org/10.1016/j.jconrel.2020.08.021>.
- [82] X. Tao, X. Wang, B. Liu, J. Liu, Conjugation of antibodies and aptamers on nanozymes for developing biosensors, *Biosens. Bioelectron.* 168 (2020) 112537. <https://doi.org/10.1016/j.bios.2020.112537>.
- [83] I. Tzouvaraki, J. Zapatero-Rodríguez, S. Naus, G. de Micheli, R. O'Kennedy, S. Carrara, Memristive biosensors based on full-size antibodies and antibody fragments, *Sensor. Actuator. B Chem.* 286 (2019) 346–352. <https://doi.org/10.1016/j.snb.2019.02.001>.
- [84] K. Tsumoto, K. Shinoki, H. Kondo, M. Uchikawa, T. Juji, I. Kumagai, Highly efficient recovery of functional single-chain Fv fragments from inclusion bodies overexpressed in *Escherichia coli* by controlled introduction of oxidizing reagent - application to a human single-chain Fv fragment, *J. Immunol. Methods* 219 (1998) 119–129. [https://doi.org/10.1016/S0022-1759\(98\)00127-6](https://doi.org/10.1016/S0022-1759(98)00127-6).
- [85] K. Siuzdak, P. Niedziałkowski, M. Sobaszek, T. Łęga, M. Sawczak, E. Czarczyk, K. Dziąbowska, T. Ossowski, D. Nidzworski, R. Bogdanowicz, Biomolecular influenza virus detection based on the electrochemical impedance spectroscopy using the nanocrystalline boron-doped diamond electrodes with covalently bound antibodies, *Sensor. Actuator. B Chem.* 280 (2019) 263–271. <https://doi.org/10.1016/j.snb.2018.10.005>.
- [86] L.A. Layqah, S. Eissa, An electrochemical immunosensor for the corona virus associated with the Middle East respiratory syndrome using an array of gold nanoparticle-modified carbon electrodes, *Microchimica Acta* 186 (2019) 2–11. <https://doi.org/10.1007/s00604-019-3345-5>.
- [87] U. Jarocka, R. Sawicka, A. Góra-Sochacka, A. Sirko, W. Dehaen, J. Radecki, H. Radecka, An electrochemical immunosensor based on a 4,4'-thio-bisbenzenethiol self-assembled monolayer for the detection of hemagglutinin from avian influenza virus H5N1, *Sensor. Actuator. B Chem.* 228 (2016) 25–30. <https://doi.org/10.1016/j.snb.2016.01.001>.
- [88] R. Singh, S. Hong, J. Jang, Label-free detection of influenza viruses using a reduced graphene oxide-based electrochemical immunosensor integrated with a microfluidic platform, *Sci. Rep.* 7 (2017) 1–11. <https://doi.org/10.1038/srep42771>.
- [89] J. Huang, Z. Xie, Z. Xie, S. Luo, L. Xie, L. Huang, Q. Fan, Y. Zhang, S. Wang, T. Zeng, Silver nanoparticles coated graphene electrochemical sensor for the ultrasensitive analysis of avian influenza virus H7, *Anal. Chim. Acta* 913 (2016) 121–127. <https://doi.org/10.1016/j.aca.2016.01.050>.
- [90] S. Devarakonda, R. Singh, J. Bhardwaj, J. Jang, Cost-effective and handmade paper-based immunosensing device for electrochemical detection of influenza virus, *Sensors* (2017) 17. <https://doi.org/10.3390/s17112597>.
- [91] D. Nidzworski, K. Siuzdak, P. Niedziałkowski, R. Bogdanowicz, M. Sobaszek, J. Ryl, P. Weiher, M. Sawczak, E. Wnuk, W.A. Goddard, A. Jaramillo-Butero, T. Ossowski, A rapid-response ultrasensitive biosensor for influenza virus detection using antibody modified boron-doped diamond, *Sci. Rep.* 7 (2017) 1–10. <https://doi.org/10.1038/s41598-017-15806-7>.
- [92] M. Veerapandian, R. Hunter, S. Neethirajan, Dual immunosensor based on methylene blue-electroadsorbed graphene oxide for rapid detection of the influenza A virus antigen, *Talanta* 155 (2016) 250–257. <https://doi.org/10.1016/j.talanta.2016.04.047>.
- [93] S.R. Joshi, A. Sharma, G.H. Kim, J. Jang, Low cost synthesis of reduced graphene oxide using biopolymer for influenza virus sensor, *Mater. Sci. Eng. C* 108 (2020) 110465. <https://doi.org/10.1016/j.msec.2019.110465>.
- [94] M. Sayhi, O. Ouerghi, K. Belgacem, M. Arbi, Y. Tepeli, A. Ghrum, Ü. Anik, L. Österlund, D. Laouini, M.F. Diouani, Electrochemical detection of influenza virus H9N2 based on both immunomagnetic extraction and gold catalysis using an immobilization-free screen printed carbon microelectrode, *Biosens. Bioelectron.* 107 (2018) 170–177. <https://doi.org/10.1016/j.bios.2018.02.018>.
- [95] J.H. Han, D. Lee, C.H.C. Chew, T. Kim, J.J. Pak, A multi-virus detectable microfluidic electrochemical immunosensor for simultaneous detection of H1N1, H5N1, and H7N9 virus using ZnO nanorods for sensitivity enhancement, *Sensor. Actuator. B Chem.* 228 (2016) 36–42. <https://doi.org/10.1016/j.snb.2015.07.068>.
- [96] Q. Zhang, L. Li, Z. Qiao, C. Lei, Y. Fu, Q. Xie, S. Yao, Y. Li, Y. Ying, Electrochemical conversion of Fe₃O₄ magnetic nanoparticles to electroactive prussian blue analogues for self-sacrificial label biosensing of avian influenza virus H5N1, *Anal. Chem.* 89 (2017) 12145–12151. <https://doi.org/10.1021/acs.analchem.7b02784>.
- [97] J. Bhardwaj, A. Sharma, J. Jang, Vertical flow-based paper immunosensor for rapid electrochemical and colorimetric detection of influenza virus using a different pore size sample pad, *Biosens. Bioelectron.* 126 (2019) 36–43. <https://doi.org/10.1016/j.bios.2018.10.008>.
- [98] C.H. Zhou, Z. Wu, J.J. Chen, C. Xiong, Z. Chen, D.W. Pang, Z.L. Zhang, Biometallization-based electrochemical magnetosensing strategy for avian influenza A (H7N9) virus particle detection, *Chem. Asian J.* 10 (2015) 1387–1393. <https://doi.org/10.1002/asia.201500105>.
- [99] Z. Wu, W.J. Guo, Y.Y. Bai, L. Zhang, J. Hu, D.W. Pang, Z.L. Zhang, Digital single virus electrochemical enzyme-linked immunoassay for ultrasensitive H7N9 avian influenza virus counting, *Anal. Chem.* 90 (2018) 1683–1690. <https://doi.org/10.1021/acs.analchem.7b03281>.
- [100] Z. Wu, C.H. Zhou, J.J. Chen, C. Xiong, Z. Chen, D.W. Pang, Z.L. Zhang, Bifunctional magnetic nanobeads for sensitive detection of avian influenza A (H7N9) virus based on immunomagnetic separation and enzyme-induced metallization, *Biosens. Bioelectron.* 68 (2015) 586–592. <https://doi.org/10.1016/j.bios.2015.01.051>.
- [101] S.K. Vashist, E.M. Schneider, A.G. Venkatesh, J.H.T. Luong, Emerging human fetuin A assays for biomedical diagnostics, *Trends Biotechnol.* 35 (2017) 407–421. <https://doi.org/10.1016/j.tibtech.2016.12.006>.
- [102] Ü. Anik, Y. Tepeli, M. Sayhi, J. Nsiri, M.F. Diouani, Towards the electrochemical diagnostic of influenza virus: development of a graphene-Au hybrid nanocomposite modified influenza virus biosensor based on neuraminidase activity, *Analyst* 143 (2018) 150–156. <https://doi.org/10.1039/c7an01537b>.
- [103] T. Matsubara, M. Ujie, T. Yamamoto, M. Akahori, Y. Einaga, T. Sato, Highly sensitive detection of influenza virus by boron-doped diamond electrode terminated with sialic acid-mimic peptide, *Proc. Natl. Acad. Sci. U.S.A.* 113 (2016) 8981–8984. <https://doi.org/10.1073/pnas.1603609113>.
- [104] T. Matsubara, A. Onishi, T. Saito, D. Yamaguchi, T. Sato, Multivalent effect in influenza hemagglutinin-binding activity of sugar-mimic peptide, *Kobunshi Ronbunshu* 73 (2016) 62–68. <https://doi.org/10.1295/KORON.2015-0052>.
- [105] T. Matsubara, M. Ujie, T. Yamamoto, Y. Einaga, T. Daidoji, T. Nakaya, T. Sato, Avian influenza virus detection by optimized peptide termination on a boron-doped diamond electrode, *ACS Sens.* 5 (2020) 431–439. <https://doi.org/10.1021/acssensors.9b02126>.
- [106] K.C. Tara Bahadur, S. Tada, L. Zhu, T. Uzawa, N. Minagawa, S.C. Luo, H. Zhao, H.H. Yu, T. Aigaki, Y. Ito, *In vitro* selection of electrochemical peptide probes using bioorthogonal tRNA for influenza virus detection, *Chem. Commun.* 54 (2018) 5201–5204. <https://doi.org/10.1039/c8cc01775a>.
- [107] A. Rogowski, J.A. Briggs, J.C. Mortimer, T. Tryfona, N. Terrapon, E.C. Lowe, A. Baslé, C. Morland, A.M. Day, H. Zheng, T.E. Rogers, P. Thompson, A.R. Hawkins, M.P. Yadav, B. Henrissat, E.C. Martens, P. Dupree, H.J. Gilbert, D.N. Bolam, Glycan complexity dictates microbial resource allocation in the large intestine, *Nat. Commun.* 6 (2015).
- [108] S. Boyoglu-Barnum, G.B. Hutchinson, J.C. Boyington, S.M. Moin, R.A. Gillespie, Y. Tsybovsky, T. Stephens, J.R. Vaile, J. Lederhofer, K.S. Corbett, B.E. Fisher, H.M. Yassine, S.F. Andrews, M.C. Crank, A.B. McDermott, J.R. Mascola, B.S. Graham, M. Kanekiyo, Glycan repositioning of influenza hemagglutinin stem facilitates the elicitation of protective cross-group antibody responses, *Nat. Commun.* 11 (2020) 1–12. <https://doi.org/10.1038/s41467-020-14579-4>.
- [109] A. Hushegyi, D. Pihiková, T. Bertok, V. Adam, R. Kizek, J. Tkac, Ultrasensitive detection of influenza viruses with a glycan-based impedimetric biosensor, *Biosens. Bioelectron.* 79 (2016) 644–649. <https://doi.org/10.1016/j.bios.2015.12.102>.