

Article

Simultaneous Profiling of DNA Mutation and Methylation by Melting Analysis Using Magnetoresistive Biosensor Array

Giovanni Rizzi, Jung-Rok Lee, Christina Dahl, Per Guldberg,
Martin Dufva, Shan X. Wang, and Mikkel Fougth Hansen

ACS Nano, **Just Accepted Manuscript** • DOI: 10.1021/acsnano.7b03053 • Publication Date (Web): 23 Aug 2017

Downloaded from <http://pubs.acs.org> on August 23, 2017

Just Accepted

"Just Accepted" manuscripts have been peer-reviewed and accepted for publication. They are posted online prior to technical editing, formatting for publication and author proofing. The American Chemical Society provides "Just Accepted" as a free service to the research community to expedite the dissemination of scientific material as soon as possible after acceptance. "Just Accepted" manuscripts appear in full in PDF format accompanied by an HTML abstract. "Just Accepted" manuscripts have been fully peer reviewed, but should not be considered the official version of record. They are accessible to all readers and citable by the Digital Object Identifier (DOI®). "Just Accepted" is an optional service offered to authors. Therefore, the "Just Accepted" Web site may not include all articles that will be published in the journal. After a manuscript is technically edited and formatted, it will be removed from the "Just Accepted" Web site and published as an ASAP article. Note that technical editing may introduce minor changes to the manuscript text and/or graphics which could affect content, and all legal disclaimers and ethical guidelines that apply to the journal pertain. ACS cannot be held responsible for errors or consequences arising from the use of information contained in these "Just Accepted" manuscripts.



ACS Publications

ACS Nano is published by the American Chemical Society, 1155 Sixteenth Street N.W., Washington, DC 20036

Published by American Chemical Society. Copyright © American Chemical Society. However, no copyright claim is made to original U.S. Government works, or works produced by employees of any Commonwealth realm Crown government in the course of their duties.

Simultaneous Profiling of DNA Mutation and Methylation by Melting Analysis Using Magnetoresistive Biosensor Array

*Giovanni Rizzi[†], Jung-Rok Lee^{‡,§}, Christina Dahl[⊥], Per Guldberg[⊥], Martin Dufva[†],
Shan X. Wang^{*§,||}, Mikkel F. Hansen^{*†}*

[†]Department of Micro- and Nanotechnology DTU Nanotech, Building 345B, Technical University of Denmark, Kongens Lyngby, DK 2800, Denmark

[‡]Division of Mechanical and Biomedical Engineering, ELTEC College of Engineering, Ewha Womans University, Seoul 03760, South Korea

[§]Department of Materials Science and Engineering, Stanford University, Stanford, CA 93405, USA

[⊥]Danish Cancer Society Research Center, Copenhagen, DK 2100, Denmark

^{||}Department of Electrical Engineering, Stanford University, Stanford, CA 93405, USA

ABSTRACT

Epigenetic modifications, in particular DNA methylation, are gaining increasing interest as complementary information to DNA mutations for cancer diagnostics and prognostics. We introduce a method to simultaneously profile DNA mutation and methylation events for an array of sites with single site specificity. Genomic (mutation) or bisulphite-treated (methylation) DNA is amplified using non-discriminatory primers, and the amplicons are then hybridized to a giant magnetoresistive (GMR) biosensor array followed by melting curve measurements. The GMR biosensor platform offers scalable multiplexed detection of DNA hybridization, which is insensitive to temperature. The melting curve approach further enhances the assay specificity and tolerance to variations in probe length. We demonstrate the utility of this method by simultaneously profiling five mutation and four methylation sites in human melanoma cell lines. The method correctly identified all mutation and methylation events and further provided quantitative assessment of methylation density validated by bisulphite pyrosequencing.

KEYWORDS: methylation, mutation, DNA array, GMR biosensor, melanoma, melting curve.

Cancer is a cellular disease caused by the stepwise accumulation of genetic and epigenetic alterations.^{1,2} Extensive sequencing efforts have identified recurrent genetic mutations that are useful as genetic biomarkers for assessing risk of developing cancer, classifying disease subtypes, predicting response to treatment, and monitoring efficacy of treatment. DNA methylation causes epigenetic silencing of tumor suppressor genes and is studied for both its direct implication in oncogenesis and for its utility as cancer biomarker.^{3–5} In bladder and colon cancer, the combination of genetic and epigenetic analyses has been proven to have a higher diagnostic value than either of the two approaches applied separately.^{6,7} However, compared to mutation genotyping, methylation profiling is not a yes-no result. Many promoter regions contain CpG islands, *i.e.* clusters of multiple CpG dinucleotides (the most common methylation site), and gene silencing mechanisms driven by promoter methylation are generally sensitive to the overall density of methylated sites. Finally, the methylation density may vary between alleles and cells within a single tumor, resulting in a heterogeneous pattern.^{4,8,9}

A variety of techniques have been developed to detect single point mutations in DNA based on amplification,^{10,11} probe hybridization,¹² enzymatic digestion,¹¹ gel electrophoresis,¹³ or sequencing.¹⁴ DNA methylation information is lost during polymerase chain reaction (PCR) amplification, and DNA hybridization is insensitive to the methylation status of the target region. Therefore, a methylation sensitive pretreatment of the DNA has to be employed. The main DNA methylation analysis techniques are based on methylation sensitive enzymatic digestion, affinity enrichment using antibodies specific for methylated cytosine or bisulphite conversion of unmethylated cytosine into uracil (reviewed by Laird⁹ and Plongthongkum *et al.*⁸). Bisulphite conversion is most widely used since a methylation event is converted into a single base alteration (C/T) that can be detected with techniques derived from mutation detection including

sequencing array hybridization, methylation sensitive PCR,⁹ and methylation sensitive melting curve analysis.^{15,16} Sequencing of bisulphite-converted DNA quantifies the methylation status and allows for comparison of data from different sequencing runs and batches, but it is costly and time consuming.⁸ Amplification-based techniques are limited in multiplexing by the number of available fluorophores and the number of channels in the thermocycler. Similarly, melting-based techniques as high resolution melting (HRM) have limited multiplexing capabilities and are only weakly sensitive to A-T and C-G single base substitutions. Array-based methods, such as the Illumina BeadChip (Illumina Inc., San Diego, CA), offer a highly multiplexed site-specific assay. However, after bisulphite conversion and amplification of the template, DNA products comprise mostly three bases (guanine, adenine, and thymine plus residues of methylated cytosine). This reduced sequence complexity makes design of probes for end-point detection complicated and the decreased sequence variation reduces specificity.⁹

Our work aims to perform methylation and mutation profiling simultaneously in a scalable chip platform that offers highly specific and quantitative DNA methylation and mutation data on a compact, easy-to-use, and potentially low-cost platform. Our approach is based on hybridization of magnetically labelled target DNA to DNA probes tethered to the surface of a GMR biosensor array (Figure 1). The magnetic detection approach has been previously demonstrated for DNA detection^{17–19} as well as immunoassays.²⁰ To increase the specificity of the DNA hybridization assay, we employed melting curve measurements of the surface-tethered DNA hybrids.²¹ This avoids conventional assay condition optimization since the target-probe hybrids are exposed to continuously increasing stringency during melting curve measurement.²² Melting curves for surface-tethered DNA probes have been also measured using fluorescence²³ and surface plasmon resonance.^{24,25} Compared to these methods, the GMR biosensors offer high

sensitivity, no dependence on temperature and are insensitive to the sample matrix as virtually all biological material is non-magnetic.^{20,26}

Using our technique, we analyzed melanoma cell lines with known genetic and epigenetic alterations.^{27,28} More specifically, we investigated mutations in *NRAS* and *BRAF*, which are found in 15-25% and 50-70% of melanomas, respectively.²⁹ In addition, we quantitatively measured the methylation status of the promoter regions of the genes encoding retinoic acid receptor β (*RARB*) and the receptor tyrosine kinase KIT (*KIT*), which are targeted by hypermethylation in 20-70% and 25-40% of melanomas, respectively.^{28,30,31}

RESULTS AND DISCUSSION

DNA Mutation analysis

To detect DNA mutations, we PCR amplified the genomic regions of interest using non-discriminatory primers. The PCR products were then magnetically labeled using biotinylated primers and streptavidin-coated magnetic nanoparticles (MNPs). After magnetic column separation and denaturation of the double-stranded PCR products, ssDNA conjugated to MNPs (MNP-ssDNA) was introduced to the GMR biosensor array where multiple DNA probes were separately tethered to the surface of each sensor.¹⁹ Upon hybridization of the injected MNP-ssDNA to surface-tethered complementary probes, GMR biosensors produced changes in sensor magnetoresistive ratio (ΔMR) in proportion to the bound MNPs.³² To genotype a mutation, we employed a set of two probes complementary to the wild type (WT) and mutant type (MT) sequences of the sample (supplementary information Tabel S1). During hybridization at low stringency, amplicons hybridized to both WT and MT probes with similar affinity. To obtain single base specificity, stringent washing is typically used after hybridization in DNA microarray. To achieve a more flexible system for detection of single-base mutations, we

1
2
3 challenged the hybrids by increasing the temperature and continuously measuring DNA melting
4
5 simultaneously for all probes on the GMR biosensor array (Figure 1).
6
7

8 Figure 2a shows real-time monitoring of ΔMR during hybridization (1 hour at 37 °C) of a
9
10 known WT sample to probes with a perfect match (WT) or a single-base mismatch (MT) for the
11
12 *BRAF* c.1391G>A mutation. In addition, a biotinylated DNA probe was used as positive
13
14 reference and a DNA probe with an unspecific sequence was used as negative reference. After 1
15
16 hour of hybridization, the MT probe gave a slightly higher ΔMR signal than the WT probe,
17
18 indicating that low-stringency hybridization was insufficient to genotype the WT sample. The
19
20 positive reference signal reached a higher level than the hybridization probes due to the faster
21
22 kinetics of the biotin-streptavidin interaction compared to hybridization. After hybridization, the
23
24 unbound sample was removed by a low-temperature wash at low stringency. Then, the
25
26 temperature of the sensor was ramped at constant rate from 20 °C until all DNA hybrids melted
27
28 at 65 °C (Figure 2a). The signal (ΔMR) from GMR biosensors was corrected for its temperature
29
30 dependence during ramping using the sensor resistance (R), which is linearly related to the
31
32 sensor temperature.
33
34
35
36
37

38 Figure 2b shows the melting curve of WT *BRAF* amplicons hybridized to WT and MT probes
39
40 for the c.1391G>A mutation. Here, the ΔMR signal was normalized by the initial signal at $T=20$
41
42 °C. We defined the melting temperature T_m as the temperature at which the signal (ΔMR)
43
44 dropped to the half of its initial signal (at 20 °C). Each melting experiment was repeated with
45
46 two identical GMR biosensor chips. Three sensors were functionalized with each probe, thus
47
48 generating up to six identical melting curves for each probe. The obtained melting curves were
49
50 found to be highly reproducible – both from sensor to sensor and from chip to chip. The hybrids
51
52 of the target DNA with WT and MT probes in Figure 2b showed melting temperatures of
53
54
55
56
57
58
59
60

$T_m(\text{WT})=43.0(7)^\circ\text{C}$ and $T_m(\text{MT})=38.9(7)^\circ\text{C}$, respectively, where the numbers in parentheses are standard deviations of T_m on the last digit ($n\geq 4$). We defined the melting temperature difference, ΔT_m , as the difference between the melting temperature from the MT probe and that from the WT probe, $\Delta T_m = T_m(\text{MT}) - T_m(\text{WT})$. Thus, $\Delta T_m < 0$ indicates a higher complementarity of the target to the WT probe than the MT probe, and hence that the target is WT. The obtained value $\Delta T_m = -4.0(3)^\circ\text{C}$ is in agreement with the expectation for a single base mismatch between the WT target and MT probe using a nearest neighbor calculation.³³ We also note that the lower standard deviation of ΔT_m compared to T_m indicates that differences in melting temperatures were more reproducible than their absolute values.

Figure 2c shows melting curves measured for a cell line heterozygous for the *BRAF* c.1391G>A mutation.²⁷ The melting curves from WT and MT probes were found to overlap each other, resulting in $\Delta T_m = -0.6(4)^\circ\text{C}$ because the heterozygous sample contains both MT and WT targets, which hybridize to both WT and MT probes. The resulting melting curves from WT and MT probes were both given by the contribution of low- T_m and high- T_m DNA hybrids. Therefore, the melting curves overlapped and presented a lower slope.

DNA methylation analysis

We applied a similar detection scheme to analyze the methylation state of specific regions of the target. We employed bisulphite treatment of the genomic DNA (Figure 3a) to convert a methylation event into a single base substitution (C>T). After bisulphite conversion, we amplified the gene promoter region of interest by non-discriminatory PCR.

Figures 3b and c show melting curves to estimate methylation status of the *KIT* promoter (site p1) of hypermethylated (EST045) and wild-type (EST164) cell line.²⁸ The amplicons were

1
2
3 hybridized to probes complementary to unmethylated (U) or methylated (M) target DNA.
4
5 Melting curves were measured as described previously. Here, ΔT_m was defined as the melting
6
7 temperature of the M probe minus that of the U probe, $\Delta T_m = T_m(M) - T_m(U)$. Thus, a negative
8
9 ΔT_m indicates a higher complementarity of the target to the U probe and a lower degree of
10
11 methylation. The ~20 bp region of the *KIT* promoter investigated includes three CpG sites that
12
13 can be methylated (sequences in supplementary information Table S1), and thus we expect
14
15 higher ΔT_m than for single base substitution. For the hyper-methylated cell line in Figure 3b, we
16
17 found $\Delta T_m = 8.1(1)^\circ\text{C}$, confirming the hyper-methylation status of the *KIT* promoter, whereas we
18
19 found $\Delta T_m = -11.7(7)^\circ\text{C}$ for the WT cell line in Figure 3c, indicating the unmethylated status.
20
21
22
23
24
25
26

27 **Multiplex DNA profiling of melanoma cell lines**

28
29 The GMR biosensor array comprises of 64 individual sensors that can be individually
30
31 functionalized with amino-modified DNA probes. Using the mutation and methylation detection
32
33 techniques described above, we simultaneously probed three mutation sites in *BRAF*, two
34
35 mutation sites in *NRAS*, two methylation sites in the *KIT* promoter, and two methylation sites in
36
37 the *RARB* promoter in triplicate. We performed mutation and methylation profiling of seven
38
39 melanoma cell lines. For each cell line, the targeted regions of *BRAF* and of *NRAS* were
40
41 amplified by non- discriminatory PCR. Also, the promoter regions of *KIT* and *RARB* were
42
43 amplified by non- discriminatory PCR after bisulphite conversion. After magnetic labeling, a
44
45 mixture of all amplicons from a cell line was injected over the sensor surface. For each cell line,
46
47 melting curve profiling was repeated with two nominally identical GMR biosensor arrays. The
48
49 melting curves were analyzed in terms of melting temperatures, and we determined ΔT_m for all
50
51 investigated mutations and methylation.
52
53
54
55
56
57
58
59
60

Figure 4a shows the ΔT_m values measured for the *BRAF* c.1391G>A mutation for all cell lines (other mutation sites are shown in supplementary Figure S1). Six cell lines showed ΔT_m values around $\Delta T_m = -4$ °C, indicating a homozygous WT sequence. EST164 is known to be the only cell line with a heterozygous mutation in this site,²⁷ showing $\Delta T_m = -0.5(4)$ °C, which is significantly different from the other cell lines.

The ΔT_m values measured for all investigated mutations for each cell line are displayed in the heat map of Figure 4b. Classifying WT ($\Delta T_m < -2$ °C), heterozygous MT (-2 °C $< \Delta T_m < 2$ °C), and homozygous MT ($\Delta T_m > 2$ °C) resulted in the mutation map presented in Figure 4c. All mutations identified in the cell lines were consistent with previous genotyping data.²⁷ For the *NRAS* c.182 A>T mutation in the cell line EST045, we measured $\Delta T_m = -0.2(4)$ °C, genotyping the cell line as heterozygous for this mutation; however, the cell line is known to be heterozygous for an A>G substitution in that location (see supplementary Figure S2). As an MT probe targeting an A>T mutation was employed, both the WT and MT probes were similarly mismatched to the target, resulting in ΔT_m close to zero. The absolute values of T_m in supplementary Figure S2 were comparable to the other investigated mismatched probes confirming the mismatch of the target to both the WT and MT probes. Therefore, an unknown mutation can be detected by a lower T_m from the WT probe, but probes targeting all possible mutations should be included in the assay to perform accurate genotyping.

DNA methylation density

Methylation profiling differs substantially from genotyping since the methylation status of each CpG site in the promoter region varies between alleles and within a cell population. Therefore, it requires a different data analysis in terms of methylated fraction of the sample

1
2
3 DNA. We measured melting curves using surface-tethered probes targeting two locations of the
4
5 *KIT* promoter and two locations of the *RARB* promoter. The targeted sequences contain one to
6
7 four CpG sites. Combining multiple investigated sites with the intrinsic variation of the
8
9 methylation pattern, we obtained a continuous variation of ΔT_m for the analyzed cell lines. Figure
10
11 5a shows the measured ΔT_m values for all cell lines. Here, higher ΔT_m in yellow or red indicates
12
13 higher affinity of the sample to the M probe, *i.e.*, a hypermethylation event. The complex ΔT_m
14
15 pattern is a direct consequence of the intrinsic methylation variation.
16
17
18

19
20 The methylation density was assessed independently by pyrosequencing of the bisulphite-
21
22 converted DNA. To each target sequence corresponding to the probes, we calculated methylation
23
24 density depending on both the fraction of methylated sample and the number (1 to 4) of
25
26 methylated CpG sites in the region targeted by the probe. Figures 5b and 5c show ΔT_m measured
27
28 using the GMR biosensor *versus* the methylation density obtained by pyrosequencing for the *KIT*
29
30 p1 and p2 probe locations and the *RARB* p1 and p2 probe locations, respectively. In these plots,
31
32 each point corresponds to one of the measured cell lines. There is an evident linear correlation
33
34 between ΔT_m and the methylation density ($R^2 > 0.94$ for all probe locations, results of linear
35
36 regression are given in supplementary Table S4). For the *KIT* p1 and p2 probes, the slopes are
37
38 comparable (~ 0.22 °C/%, Figure 5b), whereas for the *RARB* probes, the slopes for the p1 and p2
39
40 probes differ significantly (p1: $0.076(5)$ °C/%, p2: $0.22(2)$ °C/%). The slope for the p2 probe was
41
42 three times that for the p1 probe because the p2 probe covers three CpG sites whereas the p1
43
44 probe only covers one CpG site. Nevertheless, three methylation sites allowed for a more
45
46 complex pattern of methylation sites and thus the *RARB* p2 probe showed a broader spread of
47
48 data around the best linear fit. The probes can be tailored to sacrifice linearity to favor higher
49
50 values of ΔT_m .
51
52
53
54
55
56
57
58
59
60

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60

These results demonstrate the application of temperature melting on a GMR biosensor as a semi-quantitative method for profiling methylation density. High-throughput profiling of genome wide methylation can be performed with single-base resolution using array-based methods like Illumina BeadChips but the specificity of such arrays is limited by lower sequence variability of bisulphite converted DNA.^{8,9} A quantification of the overall methylation density of a gene promoter can be obtained with methylation-specific melting curve analysis.^{15,16} Here, we combined the throughput and scalability of arrays with the specificity and flexibility of melting curve analysis. The obtained quantitative profiling was equivalent to the results of pyrosequencing.

CONCLUSION

We have presented an approach for simultaneous DNA mutation and methylation profiling. Our method combines the widespread DNA microarray techniques for both mutation and methylation analysis in a single platform. Melting curves measurements are used to increase the specificity of mutation detection. For methylation detection, the melting curve quantifies the methylation state at a level equivalent to pyrosequencing. The same technique could potentially be employed on a variety of other platforms capable of masuring DNA hybridization vs. temperature.

The GMR biosensor platform has a low cross-sensitivity to temperature and provides a sensitive readout. Although it does not offer the extreme throughput as advanced bead microarray systems (*e.g.*: Illumina), in its present format, the GMR biosensor platform can be used for the simultaneous triplicate investigation of about 20 mutation and methylation sites.

Nevertheless, the GMR biosensor array has a modular design that can be scaled to include up to thousands of biosensors.³²

METHODS

Cells and reagents. Melanoma cell lines for this study were obtained from The European Searchable Tumour Line Database (ESTDAB: <http://www.ebi.ac.uk/ipd/estdab>) and were maintained in RPMI-1640 medium containing 10% FBS and antibiotics at 37 °C and 5% CO₂. The PCR primers for this study have been modified from Dahl *et al.*^{27,28} and were obtained from DNA Technology A/S, Denmark. The sequences can be found in the supplementary material Table S2. The amine modified DNA probes (sequences in supplementary material Table S1) were matched for melting temperature calculated with nearest-neighbor model.³³ The probes were obtained from DNA Technology A/S. The other reagents: poly(ethylene-alt-maleic anhydride) (Sigma Aldrich), poly(allylamine hydrochloride) (Polyscience), distilled water (Invitrogen), 1-Ethyl-3-(3-dimethylaminopropyl)-carbodiimide EDC (Sigma Aldrich), N-hydroxysuccinimide NHS (Sigma Aldrich) 1% bovine serum albumin BSA (Sigma Aldrich), phosphate buffered saline PBS (Gibco), distilled water (Invitrogen), Tween 20 (Sigma Aldrich), Urea (Fisher Scientific), 20× saline sodium citrate SSC (Invitrogen), mineral oil (Sigma Aldrich), MNPs Streptavidin MicroBeads (Miltenyi), magnetic separation columns μ Columns (Miltenyi).

DNA extraction and bisulphite treatment. Genomic DNA was isolated using the Qiagen AllPrep DNA/RNA/Protein Mini kit (Qiagen GmbH, Hilden, Germany) and quantified using a NanoDrop ND-1000 spectrophotometer (NanoDrop Technologies, Wilmington, DE). Bisulphite

conversion of DNA (500 ng) was carried out using the EZ DNA Methylation-Gold™ Kit (Zymo Research, Irvine, CA) according to the manufacturer’s protocol.

PCR amplification. Prior to the GMR biosensor assay, PCR was performed using a Veriti™ 96-Well Thermal Cycler (Applied Biosystems) and TEMPase Hot Start Polymerase (VWR). All amplifications were initiated with enzyme activation and DNA denaturation at 95 °C for 15 minutes, 40 cycles of 95 °C for 30 seconds, 56 °C for 30 seconds and 72 °C for 30 seconds, followed by a final incubation at 72 °C for 10 minutes. All primer sequences used are listed in the supplementary material Table S2.

Magnetic labelling. Products from PCR amplification of each cell line were processed as described by Rizzi *et al.*¹⁹ to obtain ss-DNA target conjugated with MNPs. Briefly, for each amplified region, 10 µL of PCR products were mixed with 10 µL of stock solution of streptavidin-coated MNPs (MACS Streptavidin Microbeads, cat: 130-048-102, Miltenyi Biotec Norden AB, Lund, Sweden) and incubated for 30 min at 37 °C. A magnetic separation column (µ column, Miltenyi Biotec Norden AB, Lund, Sweden) was prepared by washing with 1 mL of 1% Tween 20 and 1 mL of 0.1% BSA containing 0.05% Tween 20, sequentially. After the conjugation with magnetic particles, the five PCR products were mixed and added to the column under an applied magnetic field for separation. While the target DNA-bead complexes were trapped in the column, the reverse strands were denatured and removed by adding 2 mL of 6 M Urea solution at 75 °C. Then, the applied magnetic field was removed, and the conjugated complexes were eluted with 100 µL of 2× SSC buffer.

Sensor preparation. The GMR biosensor chip with an array of 8 × 8 sensors was fabricated as previously described.³⁴ The chip surface was chemically activated following Kim *et al.*³⁵ Briefly,

the chip was sequentially washed with acetone, methanol, and isopropanol. After cleaned with oxygen plasma, the chip was treated with poly(ethylene-alt-maleic anhydride) for 5 min. Then, the chip was washed with distilled water, and baked at 110 °C for 1 hour using a hot plate. After treatment with poly(allylamine hydrochloride) for 5 min, the chip was washed with the distilled water, and activated with a mixture of NHS and EDC for 1 hour. After the chip was washed again with distilled water, a robotic arrayer (sciFlexarrayer, Scienion) was used to print the amino-modified DNA probes on different sensors (Supplementary Table S2). Each DNA probe was dissolved in 3× SSC buffer at 20 μM, and was used for printing in triplicate. Four sensors on the same chip were functionalized with biotinylated DNAs as positive references, and another set of four sensors was functionalized with DNA non-complementary to any of the PCR amplified regions as negative references. The chip was stored at room temperature until use.

Data acquisition for temperature calibration. Prior to the assay, the thermal resistivity of each GMR sensor was characterized for temperature calibration. The temperature coefficient for each chip was obtained by linear fitting to resistance measurements at 20, 30, 40, 50 and 60 °C. The temperature coefficient was then used to trace the instantaneous temperature of each sensor.

GMR biosensor assay. The measurement setup described previously^{26,36} was employed to measure the response of GMR sensor to MNPs. The temperature of the GMR chips was controlled by means of a Peltier element coupled to the chip. The Peltier element was driven by a LFI3751 control unit (Wavelength Electronics, USA) with a Pt1000 thermoresistor (Supplementary Information Figure S4). First, the chip was washed with 3 mL of 0.1% bovine serum albumin (BSA, Sigma Aldrich) and 0.05% Tween20 (Sigma Aldrich). The chip was then blocked with 100 μL of 1% BSA for 1 h in a shaker. After blocking, the chip was washed with 3 mL of 0.1% BSA and 0.05% Tween20, followed by washing with 3 mL of distilled water. Prior

to sample injection, a base line signal measurement was performed for 2 min at 37 °C. After sample injection, the DNA hybridization signals from different sensors were measured for 1 h at 37 °C. Then, the temperature was lowered to 20 °C to inhibit further binding of the sample and stabilize DNA hybrids. The chip was washed five times with 100 μL of 0.05× SSC to remove unbound sample. 150 μL of 0.05× SSC were left in the reaction well and covered with 50 μL of mineral oil to prevent evaporation. Melting curves for all the probes were measured while temperature was ramped from 20 °C to 65 °C at 0.05 °C/s. The temperature was then swept back to 20 °C to obtain temperature reference signals.

Data processing. The MNPs are detected as a variation of the magnetoresistivity (ΔMR) of the GMR sensors. The temperature coefficients of ΔMR were calculated for each sensor as in Hall *et al.*²⁶ using the temperature reference signals. It was necessary to use a 5th order polynomial to account for non-linear temperature dependency at high temperature (Supplementary Information Figure S5) . After temperature correction, the melting curves measured between 20 °C to 65 °C were normalized by their initial value at 20 °C. The melting temperature T_m was defined as the temperature at which the normalized signal is 0.5. To calculate T_m , a first order polynomial was fitted to the melting curve in the region of interest. Each mutation (methylation) site was genotyped using two probes. ΔT_m was defined as the difference between T_m measured for the probe complementary to the mutated (methylated) sequence minus T_m measured for the probe complementary to the Wild Type (un-methylated) sequence.

Pyrosequencing. The methylation status of the RARB and KIT promoter regions was analyzed by pyrosequencing using the PyroMark Q24 platform (Qiagen) and subsequent data analysis using the PyroMark Q24 software. Primer sequences are listed in Supplementary Table S3. DNA enzymatically methylated *in vitro* (CpGenome Universal Methylated DNA; Millipore) and

1
2
3 unmethylated DNA prepared by whole genome amplification (WGA; GenomePlex, Sigma-
4
5 Aldrich) was used as methylation-positive and -negative controls, respectively.
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60

FIGURES

Figure 1. Schematic protocol for the detection of magnetically labelled DNA using GMR biosensors: **(I)** After denaturation of the reverse strand and labeling, PCR products are injected into the reaction well over the chip. The DNA labeled with MNPs hybridizes to complementary surface-tethered probes for 1 hour at 37 °C. **(II)** The DNA is hybridized to the sensor surface and unbound sample is removed by washing. **(III)** The temperature is swept from 20 °C to 65 °C to measure the melting temperature, T_m .

Figure 2. (a) Time series of the signal (ΔMR) from GMR biosensors functionalized with positive and negative references, wild type (WT) and mutant type (MT) probes for the *BRAF* c.1391G>A mutation. Each line corresponds to up to three sensors functionalized with the same probe. The measurement was performed with PCR products from EST045 cell line that is WT for the investigated mutation. The sample was injected at $t = 2$ min. After 60 min hybridization at 37 °C, the chip was washed and the temperature was ramped from 20 °C to 65 °C while measuring the melting curves. **(b)** and **(c)** show melting curves from WT (green) and mutant type MT (red) probes targeting *BRAF* c.1391G>A mutation obtained for the indicated cell lines, where the EST045 and EST164 cell lines were wild type and heterozygous mutant, respectively. Signals were normalized by the initial signal at $T=20$ °C. The melting temperature T_m is defined as the temperatures at which the normalized curves cross 0.5. ΔT_m is the difference in melting temperature between the MT and WT probes. The numbers in parentheses are standard deviations on the last significant digit ($n = 4-6$). Graphs of the first derivative of the melting curves are available in Supplementary Information Figure S3.

Figure 3. (a) Schematic of the bisulphite conversion process: Upon bisulphite treatment, unmethylated cytosines are converted to uracil whereas 5-methylcytosines are retained. In the subsequent PCR, uracil is substituted by thymine. Thus, the methylated cytosines are mapped to single base alterations (C>T) of the amplicons. **(b)** and **(c)** show melting curves from methylated (M, red) and unmethylated (U, green) probes targeting *KIT* methylation (site p1). The melting curves were measured for **(b)** the hypermethylated cell line EST045 and **(c)** the unmethylated cell line EST164. Graphs of the first derivative of the melting curves are available in Supplementary Information Figure S3.

Figure 4: (a) Mutation profiling of melanoma cell lines. ΔT_m measured for *BRAF* c.1391G>A mutation for the seven investigated EST cell lines. Error bars are one standard deviation ($n = 4-6$). The horizontal lines are threshold values used for genotyping: $\Delta T_m < -2$ °C (green) WT, -2 °C $< \Delta T_m < 2$ °C (yellow) heterozygous MT, $\Delta T_m > 2$ °C (red) homozygous MT. **(b)** Heat map of ΔT_m measured for the mutation and for the investigated EST cell lines. **(c)** Heat map of measured ΔT_m with applied threshold to genotype mutations: WT in green, heterozygous MT in yellow, homozygous MT in red.

Figure 5. Mutation and methylation profiling of melanoma cell lines. **(a)** Heat map of ΔT_m measured for *KIT* and *RARB* methylation probes for the seven investigated EST cell lines. Calculation of ΔT_m for EST007 *KIT* was not possible due to low binding signal. **(b)** ΔT_m values measured for *KIT* p1 (blue squares) and p2 (orange circles) methylation probe locations vs. methylation density measured by pyrosequencing. **(b)** ΔT_m measured for *RARB* p1 (blue squares) and p2 (orange circles) methylation probe locations vs. methylation level measured by pyrosequencing. Error bars are one standard deviation ($n = 4-6$).

ASSOCIATED CONTENT

Supporting Information. **Table S1**, DNA probe sequences. **Table S2**, PCR primer sequences. **Table S3**, pyrosequencing primers sequences. **Figure S1**, ΔT_m measured for all probes and for all cell-lines. **Figure S2**, absolute T_m values measured for *NRAS* c.182A>T and denaturing gradient gel electrophoresis of *NRAS* exon2 PCR products. **Figure S3**, first temperature derivative of melting curves shown in Figures 2-3. **Figure S4**, pictures of temperature control unit with mounted chip and its mounting in Helmholtz coils for measurements. **Figure S5**, temperature dependence of center and side tones of GRM sensor. This material is available free of charge *via* the Internet at <http://pubs.acs.org>.

AUTHOR INFORMATION

Corresponding Author

* e-mail: sxwang@stanford.edu

*e-mail: Mikkel.Hansen@nanotech.dtu.dk

Author Contributions

The manuscript was written through contributions of all authors. All authors have given approval to the final version of the manuscript.

Funding Sources

This work was supported by the Danish Council for Independent Research (Postdoc project, DFF-4005-00116), the Autoimmunity Center of Excellence (U19AI110491), and the Center for Cancer Nanotechnology Excellence (U54CA151459 and U54CA199075).

S.X.W. has related patents or patent applications assigned to the Stanford University and out-licensed for potential commercialization. S.X.W. has stock or stock options in MagArray, Inc., which has licensed relevant patents from the Stanford University for commercialization of GMR biosensor chips.

REFERENCES

(1) Hanahan, D.; Weinberg, R. A. The Hallmarks of Cancer. *Cell* **2000**, *100*, 57–70.

(2) Hanahan, D.; Weinberg, R. A. Hallmarks of Cancer: The Next Generation. *Cell* **2011**, *144*, 646–674.

(3) Esteller, M.; Corn, P. G.; Baylin, S. B.; Herman, J. G. A Gene Hypermethylation Profile of Human Cancer 1. *Cancer* **2001**, *3*, 3225–3229.

(4) Baylin, S. B. DNA Methylation and Gene Silencing in Cancer. *Nat. Clin. Pract. Oncol.* **2005**, *2*, S4–S11.

(5) Herman, J. G.; Baylin, S. B. Gene Silencing in Cancer in Association with Promoter Hypermethylation. *N. Engl. J. Med.* **2003**, *349*, 2042–2054.

(6) Serizawa, R. R.; Ralfkiær, U.; Steven, K.; Lam, G. W.; Schmiedel, S.; Schüz, J.; Hansen, A. B.; Horn, T.; Guldberg, P. Integrated Genetic and Epigenetic Analysis of Bladder Cancer Reveals an Additive Diagnostic Value of FGFR3 Mutations and Hypermethylation Events. *Int. J. Cancer* **2011**, *129*, 78–87.

(7) Shen, L.; Toyota, M.; Kondo, Y.; Lin, E.; Zhang, L.; Guo, Y.; Hernandez, N. S.; Chen, X.;

- Ahmed, S.; Konishi, K.; *Hamilton, S. R.; Issa, J. J.* Integrated Genetic and Epigenetic Analysis Identifies Three Different Subclasses of Colon Cancer. *Proc. Natl. Acad. Sci. U. S. A.* **2007**, *104*, 18654–18659.
- (8) Plongthongkum, N.; Diep, D. H.; Zhang, K. Advances in the Profiling of DNA Modifications: Cytosine Methylation and beyond. *Nat. Rev. Genet.* **2014**, *15*, 647–661.
- (9) Laird, P. W. Principles and Challenges of Genome-Wide DNA Methylation Analysis. *Nat. Rev. Genet.* **2010**, *11*, 191–203.
- (10) Orita, M.; Suzuki, Y.; Sekiya, T.; Hayashi, K. Rapid and Sensitive Detection of Point Mutations and DNA Polymorphisms Using the Polymerase Chain Reaction. *Genomics* **1989**, *5*, 874–879.
- (11) Saiki, R. K.; Scharf, S.; Faloona, F.; Mullis, K. B.; Horn, G. T.; Erlich, H. A.; Arnheim, N. Enzymatic Amplification of Beta-Globin Genomic Sequences and Restriction Site Analysis for Diagnosis of Sickle Cell Anemia. *Science.* **1985**, *230*, 1350–1354.
- (12) Wang, D. G.; Fan, J.-B.; Siao, C.-J.; Berno, A.; Young, P.; Sapolsky, R.; Ghandour, G.; Perkins, N.; Winchester, E.; Spencer, J.; Kruglyak, L.; Stein, L.; Hsie, L.; Topaloglou, T.; Hubbell, E.; Robinson, E.; Mittmann, M.; S. Morris, M.; Shen, N.; Kilburn, D.; *et al.* Large-Scale Identification, Mapping, and Genotyping of Single-Nucleotide Polymorphisms in the Human Genome. *Science* **1998**, *280*, 1077–1082.
- (13) Orita, M.; Iwahana, H.; Kanazawat, H.; Hayashi, K.; Sekiya, T. Detection of Polymorphisms of Human DNA by Gel Electrophoresis as Single-Strand Conformation Polymorphisms (Mobility Shift of Separated Strands/point Mutation/riction Franment Length Polymorphism). *Genetics* **1989**, *86*, 2766–2770.
- (14) Metzker, M. L. Sequencing Technologies — the next Generation. *Nat. Rev. Genet.* **2010**,

11, 31–46.

(15) Worm, J.; Aggerholm, A.; Guldberg, P. In-Tube DNA Methylation Profiling by Fluorescence Melting Curve Analysis. *Clin. Chem.* **2001**, *47*, 1183–1189.

(16) Wojdacz, T. K.; Dobrovic, A. Methylation-Sensitive High Resolution Melting (MS-HRM): A New Approach for Sensitive and High-Throughput Assessment of Methylation. *Nucleic Acids Res.* **2007**, *35*, e41–e41.

(17) Xu, L.; Yu, H.; Akhras, M. S.; Han, S. J.; Osterfeld, S.; White, R. L.; Pourmand, N.; Wang, S. X. Giant Magnetoresistive Biochip for DNA Detection and HPV Genotyping. *Biosens. Bioelectron.* **2008**, *24*, 99–103.

(18) Rizzi, G.; Østerberg, F. W.; Dufva, M.; Fougth Hansen, M. Magnetoresistive Sensor for Real-Time Single Nucleotide Polymorphism Genotyping. *Biosens. Bioelectron.* **2014**, *52*, 445–451.

(19) Rizzi, G.; Lee, J.-R.; Guldberg, P.; Dufva, M.; Wang, S. X.; Hansen, M. F. Denaturation Strategies for Detection of Double Stranded PCR Products on GMR Magnetic Biosensor Array. *Biosens. Bioelectron.* **2017**, *93*, 155–160.

(20) Gaster, R. S.; Hall, D. a; Nielsen, C. H.; Osterfeld, S. J.; Yu, H.; Mach, K. E.; Wilson, R. J.; Murmann, B.; Liao, J. C.; Gambhir, S. S.; Wang, S. X. Matrix-Insensitive Protein Assays Push the Limits of Biosensors in Medicine. *Nat. Med.* **2009**, *15*, 1327–1332.

(21) Rizzi, G.; Østerberg, F. W.; Henriksen, A. D.; Dufva, M.; Hansen, M. F. On-Chip Magnetic Bead-Based DNA Melting Curve Analysis Using a Magnetoresistive Sensor. *J. Magn. Magn. Mater.* **2015**, *380*, 215–220.

(22) Dufva, M.; Petersen, J.; Poulsen, L. Increasing the Specificity and Function of DNA Microarrays by Processing Arrays at Different Stringencies. *Anal. Bioanal. Chem.* **2009**,

- 395, 669–677.
- (23) Petersen, J.; Poulsen, L.; Petronis, S.; Birgens, H.; Dufva, M. Use of a Multi-Thermal Washer for DNA Microarrays Simplifies Probe Design and Gives Robust Genotyping Assays. *Nucleic Acids Res.* **2008**, *36*, e10.
- (24) Fiche, J. B.; Buhot, a; Calemczuk, R.; Livache, T. Temperature Effects on DNA Chip Experiments from Surface Plasmon Resonance Imaging: Isotherms and Melting Curves. *Biophys. J.* **2007**, *92*, 935–946.
- (25) Wagner, C. E.; Macedo, L. J. A.; Opdahl, A. Temperature Gradient Approach for Rapidly Assessing Sensor Binding Kinetics and Thermodynamics. *Anal. Chem.* **2015**, *87*, 7825–7832.
- (26) Hall, D. A.; Gaster, R. S.; Osterfeld, S. J.; Murmann, B.; Wang, S. X. GMR Biosensor Arrays: Correction Techniques for Reproducibility and Enhanced Sensitivity. *Biosens. Bioelectron.* **2010**, *25*, 2177–2181.
- (27) Dahl, C.; Christensen, C.; Jönsson, G.; Lorentzen, A.; Skjødt, M. L.; Borg, A.; Pawelec, G.; Guldberg, P. Mutual Exclusivity Analysis of Genetic and Epigenetic Drivers in Melanoma Identifies a Link Between p14ARF and RAR β Signaling. *Mol. Cancer Res.* **2013**, *11*, 1166–1178.
- (28) Dahl, C.; Abildgaard, C.; Riber-Hansen, R.; Steiniche, T.; Lade-Keller, J.; Guldberg, P. KIT Is a Frequent Target for Epigenetic Silencing in Cutaneous Melanoma. *J. Invest. Dermatol.* **2015**, *135*, 516–524.
- (29) DAHL, C.; GULDBERG, P. The Genome and Epigenome of Malignant Melanoma. *APMIS* **2007**, *115*, 1161–1176.
- (30) Hoon, D. S. B.; Spugnardi, M.; Kuo, C.; Huang, S. K.; Morton, D. L.; Taback, B.

- Profiling Epigenetic Inactivation of Tumor Suppressor Genes in Tumors and Plasma from Cutaneous Melanoma Patients. *Oncogene* **2004**, *23*, 4014–4022.
- (31) Furuta, J.; Umebayashi, Y.; Miyamoto, K.; Kikuchi, K.; Otsuka, F.; Sugimura, T.; Ushijima, T. Promoter Methylation Profiling of 30 Genes in Human Malignant Melanoma. *Cancer Sci.* **2004**, *95*, 962–968.
- (32) Gaster, R. S.; Xu, L.; Han, S.-J.; Wilson, R. J.; Hall, D. a; Osterfeld, S. J.; Yu, H.; Wang, S. X. Quantification of Protein Interactions and Solution Transport Using High-Density GMR Sensor Arrays. *Nat. Nanotechnol.* **2011**, *6*, 314–320.
- (33) SantaLucia, J. A Unified View of Polymer, Dumbbell, and Oligonucleotide DNA Nearest-Neighbor Thermodynamics. *Proc. Natl. Acad. Sci. U. S. A.* **1998**, *95*, 1460–1465.
- (34) Osterfeld, S. J.; Yu, H.; Gaster, R. S.; Caramuta, S.; Xu, L.; Han, S.-J.; Hall, D. a; Wilson, R. J.; Sun, S.; White, R. L.; Davis, R. W.; Pourmand, N.; Wang, S. X. Multiplex Protein Assays Based on Real-Time Magnetic Nanotag Sensing. *Proc. Natl. Acad. Sci. U. S. A.* **2008**, *105*, 20637–20640.
- (35) Kim, D.; Marchetti, F.; Chen, Z.; Zaric, S.; Wilson, R. J.; Hall, D. A.; Gaster, R. S.; Lee, J.-R.; Wang, J.; Osterfeld, S. J.; Yu, H.; White, R. M.; Blakely, W. F.; Peterson, L. E.; Bhatnagar, S.; Mannion, B.; Tseng, S.; Roth, K.; Coleman, M.; Snijders, A. M. *et al.* Nanosensor Dosimetry of Mouse Blood Proteins after Exposure to Ionizing Radiation. *Sci. Rep.* **2013**, *3*, 1–8.
- (36) Hall, D. A.; Gaster, R. S.; Lin, T.; Osterfeld, S. J.; Han, S.; Murmann, B.; Wang, S. X. GMR Biosensor Arrays: A System Perspective. *Biosens. Bioelectron.* **2010**, *25*, 2051–2057.











