



Article

Adenosine Deaminase Biosensor Combining Cationic Conjugated Polymer-based FRET with Deoxyguanosine-based Photoinduced Electron Transfer

Chun Wang, Yanli Tang, and Yang Guo

ACS Appl. Mater. Interfaces, **Just Accepted Manuscript** • DOI: 10.1021/am506832y • Publication Date (Web): 31 Oct 2014Downloaded from <http://pubs.acs.org> on November 7, 2014**Just Accepted**

"Just Accepted" manuscripts have been peer-reviewed and accepted for publication. They are posted online prior to technical editing, formatting for publication and author proofing. The American Chemical Society provides "Just Accepted" as a free service to the research community to expedite the dissemination of scientific material as soon as possible after acceptance. "Just Accepted" manuscripts appear in full in PDF format accompanied by an HTML abstract. "Just Accepted" manuscripts have been fully peer reviewed, but should not be considered the official version of record. They are accessible to all readers and citable by the Digital Object Identifier (DOI®). "Just Accepted" is an optional service offered to authors. Therefore, the "Just Accepted" Web site may not include all articles that will be published in the journal. After a manuscript is technically edited and formatted, it will be removed from the "Just Accepted" Web site and published as an ASAP article. Note that technical editing may introduce minor changes to the manuscript text and/or graphics which could affect content, and all legal disclaimers and ethical guidelines that apply to the journal pertain. ACS cannot be held responsible for errors or consequences arising from the use of information contained in these "Just Accepted" manuscripts.

**ACS Publications**
High quality. High impact.

Adenosine Deaminase Biosensor Combining Cationic Conjugated Polymer-based FRET with Deoxyguanosine-based Photoinduced Electron Transfer

Chun Wang, Yanli Tang*, and Yang Guo

Key Laboratory of Applied Surface and Colloid Chemistry, Ministry of Education, Key Laboratory of Analytical Chemistry for Life Science of Shaanxi Province, School of Chemistry and Chemical Engineering, Shaanxi Normal University, Xi'an 710062, P. R. China

KEYWORDS: Fluorescence resonance energy transfer; Photoinduced electron transfer; Adenosine deaminase; Conjugated polymers; Deoxyguanosine;

ABSTRACT: We demonstrated a sensitive and selective adenosine deaminase (ADA) detection by modulating the fluorescence resonance energy transfer (FRET) between cationic conjugated poly(9,9-bis(6'-N,N,N-trimethylammonium) hexyl)fluorine phenylene) (PFP) and the deoxyguanosine-tailored hairpin aptamer. The hairpin aptamer was labeled with a fluorophore FAM at one end and three deoxyguanosines (Gs) at the other end as a quencher. In the absence of ADA, aptamer forms hairpin-like conformation with adenosines making close affinity of Gs and FAM, which results in the weak FRET from PFP to FAM because of FAM fluorescence being quenched by Gs *via* photoinduced electron transfer (PET). After addition of ADA, adenosine was hydrolyzed by ADA, followed by the release of free aptamer. In this case, FAM being far away from Gs, the strong FRET thus was obtained due to the quenching process being

1
2
3 blocked. Therefore, the new strategy based on the FRET ratio enhancement is reasonably used to
4
5 detect the ADA sensitively, combining the fluorescence signal amplification of conjugated
6
7 polymers with the initiative signal decreasing by Gs. The detection limit of the ADA assay is 0.3
8
9 U/L in both buffer solution and human serum, which is more sensitive than most of those
10
11 previously documented methods. Importantly, the assay is rapid, homogeneous and simple
12
13 without complicated treating process. The ADA inhibitor, erythro-9-(2-hydroxy-3-nonyl)
14
15 adenine hydrochloride (EHNA) was also studied based on this assay, and the detection limit of
16
17 EHNA is 10 pM. This strategy provides a new platform for the detection of other biomolecules
18
19 and enzymes.
20
21
22
23
24

25 INTRODUCTION

26
27
28 Adenosine deaminase (ADA) is an enzyme that can catalyze the reaction of adenosine and 2-
29
30 deoxyadenosine converting to inosine and deoxyinosine, respectively. It plays an important role
31
32 in the function, maturation, and maintenance of immunological responses.¹⁻⁴ Both inherited ADA
33
34 deficiency and ADA excess may cause diseases.^{5, 6} For example, it was reported that an increase
35
36 of ADA activity in serum has been found to be closely related to esophagus tumors, pancreatic
37
38 cancer, liver cancer, breast cancer and colorectal cancer.⁷ On the contrary, one of main causes for
39
40 severe combined immunodeficiency (SCID), autism, etc is ADA deficiency.⁶ Therefore, a
41
42 sensitive detection of ADA is of great significance for clinical applications. So far, several
43
44 techniques have been employed to detect ADA activity, including fluorescent aptasensor,
45
46 electrochemical aptasensor based on magnetic beads, graphene oxide or silver-nanoclusters, etc.^{5,}
47
48
49
50
51
52
53
54
55
56
57
58
59
60
8-11 However, these reported techniques have inherent disadvantages, such as complexity and
limited sensitivity,^{11, 12} or requiring a laborious and time-consuming process.⁹ Accordingly, it is
still necessary to develop a simple and sensitive method to detect ADA.

1
2
3 Recently, much attention has been placed on aptamer due to its specific and high binding
4 affinity toward the target molecules.¹³⁻¹⁷ In addition, aptamer that consists of a fluorescent label
5 at one end and a nonfluorescent molecule (e.g., nucleobase) at the other end is able to work as a
6 molecular beacon. Importantly, this kind of aptamer avoids the limitations that molecular
7 beacons suffer in overall yield and are expensive.¹⁸ In the closed-state conformation, this kind of
8 probe is able to work as an acceptor to quench the fluorescence of the fluorophore *via*
9 photoinduced electron transfer (PET).¹⁹⁻²¹ In this regard, a number of unique (rather than
10 traditional) nonfluorescent quenchers, for example, guanosine, and deazaguanosine have already
11 been successfully introduced.²²⁻²⁵ The quenching efficiency of guanosine probably owes to its
12 low oxidation potential, hydrophobic interaction between the dyes and DNA nucleotide, as well
13 as formation of nonfluorescent ground state *via* photoinduced electron transfer due to close
14 proximity of stacked arrangement of nucleobase G.^{18-21, 26-28} Interestingly, this type of aptamer
15 provides a platform to design a novel biosensor with signal from turn-off to turn-on upon
16 recognition of targets.
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35

36
37 In recent years, cationic conjugated polymers (CCPs) have attracted more and more attention
38 owing to its fluorescence signal amplification.^{29, 30} A variety of biosensors and chemosensors
39 have been designed and developed by taking advantage of its particular optical properties.³¹⁻³⁷
40 These sensors have been successfully applied for detections of oligonucleotides, enzymes,
41 protein fibrillation, inhibitors, as well as for biological applications etc. with high sensitivity and
42 specificity.³⁸⁻⁴³ It is interesting to note that fluorescence resonance energy transfer (FRET)
43 technique plays an important role in most of these methods.^{31-35, 40, 42, 44} FRET is a popular
44 technique in fluorescent biosensors, which is a well studied physical process whereby individual
45 chromophore communicates its electronic states and provide the means for transferring
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60

excitations from a donor to an acceptor.⁴⁵ Therefore, the FRET technique based on fluorescent conjugated polymers is advantageous to enhance fluorescence of the receptors, and is attractive owing to their simplicity of operation and use of standard optical equipments. We have reported a new biosensor to detect ADA based on two-step FRET from PFP to ethidium bromide (EB).⁴⁶ However, this method requires EB as energy receptor and the complementary nucleic acids to form double-stranded DNAs, which increased the complexity of the assay.⁴⁷

Herein, we designed a hairpin-like adenosine aptamer with a fluorophore FAM at 5'-terminus and three deoxyguanosines (Gs) as a quencher at 3'-terminus because of three Gs producing the greatest quenching effect⁴⁸, which undergoes a conformational change from a hairpin structure to open-chain conformation in the presence of ADA. In our strategy, aptamer and adenosine can form a hairpin-like adenosine-aptamer complex, which leads to the labeled FAM being quenched by Gs *via* PET due to its close proximity. Upon addition of PFP, the FRET from PFP to FAM is very low when PFP is excited at 380 nm. After addition of ADA, the FRET from PFP to FAM is highly efficient because the conformational change of aptamer from hairpin to open-chain results in the quencher, deoxyguanosines being away from FAM. This new biosensor can sensitively detect ADA simply through measuring FRET ratio from PFP to FAM, which utilizes the fluorescence signal amplification of CCPs and the initiative signal decreasing by Gs. Also, the method provides a new platform for other enzymes detection and is potential for drug discovery screen.

EXPERIMENTAL SENCTION

Reagents and instruments. The oligonucleotides, adenosine deaminase, adenosine and bovine serum albumin (BSA) were purchased from Shanghai Sangon Biological Engineering

Technology & Service Co. Ltd. Erythro-9-(2-hydroxy-3-nonyl) adenine hydrochloride (EHNA), α -glucosidase, α -chymotrysin, thrombin, human serum and β -glucosidase were obtained from Sigma. Cationic poly(9,9-bis(6'-N,N,N-trimethylammonium)hexyl)fluorine phenylene) (PFP) was synthesized according to the literature.⁴⁴ The enzyme assay was carried out in KH_2PO_4 buffer (pH 7.5). The aptamer was hybridized in TE buffer containing 50 mmol/L Tris-HCl, 10 mmol/L KCl, 5 mol/L MgCl_2 and 100 mmol/L NaCl (pH 7.4). The oligonucleotide sequences used in our experiments are as follows: Ap-Gs: 5'-FAM-CAC TCT **GGG GGA GTA TTG CGG AGG AAG** AGT GGG G and C-Ap: 5'-FAM-GCA CCT **GGG GGA GTA TTG CGG AGG AAG** GTG C-3'. The sequence shown in bold is an adenosine aptamer. All solutions were prepared with ultrapure water purified by a Millipore filtration. The concentrations of all oligonucleotides were determined by measuring the absorbance at 260 nm in a 250 μL quartz cuvette. UV-vis absorption spectra were taken on a Perkin Elmer Lambda 35 spectrophotometer. The fluorescence measurements were recorded on a Hitachi F-7000 spectrophotometer equipped with a Xenon lamp excitation source.

RESULTS AND DISCUSSION

Scheme 1 shows the proposed principle of ADA sensing. The hairpin-like aptamer is labeled with fluorescein (FAM) at one end and three Gs at the other end designed as a fluorescence quencher,⁴⁸ which can form stem-loop conformation with target adenosine. This probe is simple and low-cost compared to the molecular beacons. In this stem-loop conformation, the fluorescence of FAM is quenched efficiently by Gs *vis* photoinduced electron transfer due to the close affinity between FAM and Gs. Even after the addition of PFP, the fluorescence of FAM is weak and the FRET ratio from PFP to FAM is pretty low when PFP is excited. In the presence of ADA, adenosine is converted into inosine, which leads to the aptamer transferring into free

conformation and results in the separation of FAM from Gs. Correspondingly, the FRET from PFP to FAM is distinctly high due to the electrostatic interactions between PFP and aptamer when PFP is excited at 380 nm. Notably, our previous research presented that the desirable spectral overlap between PFP emission and FAM absorption makes it possible for FRET from PFP to FAM.⁴⁶ Thus, the simple and sensitive detection of ADA is realized only by obtaining the FRET ratio of FAM to PFP. Importantly, this assay is simpler without any excess operation and less expensive than that reported previously.^{9, 11, 46} Furthermore, the conjugated polymers-based FRET technique and deoxyguanosine-based PET greatly enhances the sensitivity. Finally, the new method holds the great potential for the application in more complex condition.

The preferred position for Scheme 1

To optimize the concentrations of aptamer, the fluorescence spectra were measured after the addition of aptamer to PFP solution. As shown in the Figure S1a, aqueous of aptamer without adenosine were added into the solution of PFP (2.0×10^{-6} M in RUs), and then the fluorescence emission spectra were obtained with excitation wavelength at 380 nm. In this case, the aptamer takes the free conformation in the absence of targets, which makes the fluorescence of FAM seldom be quenched by Gs. Hence, the FRET from PFP to FAM is strong when free aptamer was added. Importantly, the FRET was improved dramatically with the increasing concentration of the aptamer. Finally, the optimized concentration of aptamer was obtained for 4.0×10^{-8} M. Furthermore, the ratios of FAM intensity to PFP fluorescence intensity at different concentrations of PFP and proportional aptamer were compared. As shown in Figure S1b, the ratio becomes the highest when PFP is 2.0×10^{-6} M and aptamer is 4.0×10^{-8} M. Hence, 2.0×10^{-6} M PFP and 4.0×10^{-8} M oligonucleotides were selected to obtain the best sensitivity.

Figure 1 shows that the FRET from PFP to FAM in the presence and absence of ADA. The hairpin structure aptamer was formed with adenosine by hybridization beforehand, and the concentration ratio between adenosine and aptamer is 2:1 as used previously.^{9, 46} PFP can emit strong blue fluorescence when PFP is excited at 380 nm. Upon the addition of aptamer with three Gs (Ap-Gs), in the absence of ADA, the FRET ratio (0.62) from PFP to FAM is pretty low because of the fluorescence of FAM is quenched by three deoxyguanosines, which indicates that the hairpin-like structure formed. In the presence of ADA, adenosine is converted into inosine, which makes the conformational change from hairpin to open-chain and then the quencher, three deoxyguanosines is separated from FAM. The FRET ratio (4.2) from PFP to FAM is thus dramatically high due to strong electrostatic interactions between PFP and free aptamer when PFP is excited at 380 nm. To compare the FRET ratio with that without Gs, the fluorescence spectrum in the presence of PFP and control aptamer (C-Ap) that is labeled with FAM at 5' end was measured. As shown in Figure 1, the FRET ratio is about 2.2, which is much higher than that with Gs. The results indicate that this novel aptamer with three deoxyguanosines can greatly decrease the initiative signal and correspondingly is favorable for enhancing detection sensitivity. Accordingly, a new method is designed to detect ADA simply by measuring FRET ratio, which takes advantage of the fluorescence signal amplification of PFP and background signal decreasing from deoxyguanosine.

The preferred position for Figure 1

To test the ADA activity, we measured the fluorescence emission spectra of PFP/Ap-Gs (with adenosine) at different concentrations of ADA. As shown in Figure 2, in the absence of ADA, aptamer forms stem-loop structure with adenosines. The fluorescence of FAM is weak when PFP is excited at 380 nm, because FAM is quenched by three guanosines *via* PET. Accordingly, the

FRET ratio is pretty low due to aptamer forming nonfluorescence ground state. After ADA (from 0.3 to 150 U/L) was added into the solution and incubated for 25min (The time has been optimized as followed) at 25 °C, the emission spectra were measured with the excitation wavelength at 380 nm. With the hydrolyzation of adenosine by ADA and the release of inosine, aptamer changed from loop-stem structure to free single-stranded conformation. Then, PET process is inhibited due to FAM being far away from Gs. It is observed that the fluorescence of FAM increases dramatically, meanwhile the fluorescence of PFP decreases, indicating the efficient FRET occurring from PFP to FAM. The FRET ratio increases with the increasing concentration of ADA. When the concentration of ADA increased to 150 U/L, the ratio is almost no longer changed, which indicates that the hydrolysis reaction reaches the plateau. Figure 2b demonstrates the hyperbola relationship between the increased FRET ratio of FAM to fluorescein ($I_{527}/I_{423} - (I_{527}/I_{423})_0$) and the concentration of ADA, which is typical fitting mode for enzyme assay. The fitting equation is $y = 5.44663x/(84.82185 + x)$ ($R^2 = 0.998$). Notably, we can observe the distinct fluorescence intensity increases at a low ADA concentration of 0.3 U/L (The analyte's signal at the detection limit (S_{dl}) is given by: $S_{dl} = S_{bl} + 3 \sigma_{bl}$, where S_{bl} is the signal for the blank, σ_{bl} is the known standard deviation for the blank's signal.). These results represent that this method is sensitive to ADA and the limit of detection (LOD) is lower than most of previous biosensors reported by Wang et al.⁴⁶ (0.5 U/L, 2014), Zhang et al.⁴⁹ (0.5U/L, 2013), Feng et al.⁵⁰ (0.05 U/mL, 2013) and by Xing et al.⁵ (12.9 U/L, 2012) because of CCPs-based FRET technique amplifying the fluorescence signal and deoxyguanosine-based PET decreasing background signal.^{8, 46, 49, 51} We also compared the sensitivity with that in the absence of PFP. The assays without PFP were carried out by simply measuring the fluorescence emission spectra of aptamer at different concentrations of ADA. As shown in Figure S2, the detection limit of 2.0 U/L was

obtained which is about seven-fold higher than that with PFP. These results further provide a proof of conjugated polymers amplifying the FRET ratio and improving greatly the detection sensitivity.

The preferred position for Figure 2

Besides, in order to optimize the ADA reaction time, we measured the fluorescence emission spectra of PFP/aptamer (with adenosine)/ADA at different reaction times. Figure S3 shows the FRET ratio of FAM to PFP as a function of reaction time. At 25 min, the maximum FRET ratio was obtained and reached the plateau no matter what the concentration of ADA was.⁴⁶ As a result, 25 min was selected as the optimized reaction time, which further demonstrates the detection of ADA is rapid and can be completed in a relatively short time. We also investigated the specificity of the proposed method by comparing the FRET ratio (I_{527}/I_{423}) after aptamer hydrolysis by ADA and a variety of other enzymes, such as alpha-glucosidase, beta-glucosidase, chymotrysin and thrombin (Figure S4). The high selectivity of this method for ADA also was observed as reported previously.^{9, 46} More importantly, the detection of ADA is carried out in a homogeneous and very simple system which only includes PFP and hairpin-like aptamer. These advantages make the method prior to other sensors that are time-consuming and/or acquire complex reaction process.

In order to study the efficiency of this biosensor in complex biological samples, we investigated the ADA activity in 10% human serum. As shown in Figure 3, the pretty similar relationship to that in KH_2PO_4 buffer was observed between the increased FRET ratio and the concentration of ADA. The FRET ratio increases fast in the range of 0~50 U/L, and then the rate falls down. When the concentration of ADA is 150 U/L, the FRET ratio is about 3.9 and reaches the plateau.

Furthermore, the obvious fluorescence intensity increase at a low ADA concentration of 0.3 U/L also was observed as in buffer solution. Therefore, this new strategy is applicable in complex biological conditions, which is superior to commonly-used biosensors that are invalid to detect proteins in natural system.⁵²

The preferred position for Figure 3

In addition, the inhibition of enzyme activity was studied by incubating ADA with different concentrations of erythro-9-(2-hydroxy-3-nonyl) adenine hydrochloride (EHNA) which is a well-known inhibitor of ADA.⁵ The concentration of ADA is fixed at 150 U/L because the activity of ADA gives the maximum response at this concentration and the incubation time is set for 25 min as mentioned above. Figure 4 shows that the fluorescence intensity of FAM decreases while the fluorescence intensity of PFP enhances with the increasing concentration of EHNA in the range of 0-600 nM. As a result, the FRET ratio of FAM to PFP decreases, which is dependent on the amount of EHNA. High concentrations of EHNA lead to FRET ratio dramatically decreasing. As shown in Figure 4b, it is observed that 600 nM EHNA can inhibit ADA completely and the detection limit of EHNA is 10 pM. These results confirm that our strategy has the potential for screening ADA inhibitors.

The preferred position for Figure 4

CONCLUSIONS

In summary, we designed a new method to sense ADA sensitively and specifically based on conjugated polymers and deoxyguanosine-modified aptamer. The detection limit of the ADA assay is determined to be 0.3 U/L not only in buffer solution but also in serum, which is lower

than most of previously reported ones. Our method is pretty rapid, and the whole assay only takes about 25 min. Overall, our method possesses three advantages: First, conjugated polymer-based FRET technique amplifies the fluorescence signal; Second, the designed aptamer with three deoxyguanosines greatly decreases the initiative signal, which is helpful for increasing detection sensitivity; Third, the good water-solubility of CCPs makes the detection homogeneous and simple, avoiding complicated isolation and washing steps. Above all, it can be extended to the detection of other enzymes in complex conditions.

FIGURES:




Figure 1. Fluorescence emission spectra of PFP, PFP/C-Ap (with adenosine), PFP/Ap-Gs (with adenosine) and PFP/Ap-Gs/ADA in KH_2PO_4 buffer solution (53.3 mM, PH = 7.5). $[\text{PFP}] = 2.0 \times 10^{-6}$ M in repeat units, $[\text{aptamer}] = 4.0 \times 10^{-8}$ M, $[\text{adenosine}] = 8.0 \times 10^{-8}$ mol/L, $[\text{ADA}] = 150$ U/L. The excitation wavelength is 380 nm.




Figure 2. (a) Fluorescence emission spectra of PFP/aptamer (with adenosine) in KH_2PO_4 buffer solution (53.3 mM, pH = 7.5) with addition of ADA; (b) Increase ratio of fluorescein intensity to PFP fluorescence intensity with the increasing addition of ADA. $(I_{527}/I_{423})_0$ means the ratio of fluorescein intensity in the absence of ADA. $[\text{PFP}] = 2.0 \times 10^{-6}$ mol/L in RUs, $[\text{aptamer}] = 4.0 \times 10^{-8}$ mol/L, $[\text{adenosine}] = 8.0 \times 10^{-8}$ mol/L. The error bars represent standard deviations three parallel measurements. The excitation wavelength is 380 nm.




Figure 3. Ratio of FAM intensity to PFP fluorescence intensity as a function of ADA concentration in 10% human serum. [PFP] = 2.0×10^{-6} mol/L in RUs, [aptamer] = 4.0×10^{-8} mol/L, [adenosine] = 8.0×10^{-8} mol/L. The error bars represent standard deviations three parallel measurements. The excitation wavelength is 380 nm.





Figure 4. (a) Fluorescence emission spectra of PFP/aptamer (with adenosine) in KH_2PO_4 buffer solution (53.3 mM, pH=7.5) with addition of EHNA; (b) Ratio of fluorescein intensity to PFP fluorescence intensity with the increasing amount of EHNA. [PFP] = 2.0×10^{-6} mol/L in RUs, [aptamer] = 4.0×10^{-8} mol/L, [adenosine] = 8.0×10^{-8} mol/L, [ADA] = 150 U/L. The error bars represent standard deviations three parallel measurements. The excitation wavelength is 380 nm.

SCHEMES



Scheme 1. Principle of Adenosine Deaminase Detection Using Conjugated Polymer-Based FRET Technique, Together with the Chemical Structures of PFP, Adenosine and Inosine Used in the Detection.

ASSOCIATED CONTENT

Supporting Information. Detailed experimental procedures and other figures as noted in text.

This material is available free of charge via the Internet at <http://pubs.acs.org>.

AUTHOR INFORMATION

Corresponding Author

*Phone: (086) 029-81530844. Fax: (086) 029-81530727. E-mail: yltang@snnu.edu.cn.

Notes

The authors declare no competing financial interest.

ACKNOWLEDGMENT

The authors are grateful for the financial support from the National Natural Science Foundation of China (Grants 21222509, 21105061), Fundamental Research Funds for the Central Universities (Grant GK201103002), the 111 Project (Grant B14041), and Program for Changjiang Scholars and Innovative Research Team in University (Grant IRT1070).

REFERENCES

- (1) Garcia-Gil, M.; Tozzi, M. G.; Allegrini, S.; Folcarelli, S.; Della Sala, G.; Voccoli, V.; Colombaioni, L.; Camici, M. Novel Metabolic Aspects Related to Adenosine Deaminase Inhibition in a Human Astrocytoma Cell Line. *Neurochem. Int.* **2012**, *60*, 523-532.
- (2) Cristalli, G.; Costanzi, S.; Lambertucci, C.; Lupidi, G.; Vittori, S.; Volpini, R.; Camaioni, E. Adenosine Deaminase: Functional Implications and Different Classes of Inhibitors. *Med. Res. Rev.* **2001**, *21*, 105-128.

- (3) Aldrich, M. B.; Blackburn, M. R.; Kellems, R. E. The Importance of Adenosine Deaminase for Lymphocyte Development and Function. *Biochem. Biophys. Res. Commun.* **2000**, 272, 311-315.
- (4) Elowe, N. H.; Nutiu, R.; Allali-Hassani, A.; Cechetto, J. D.; Hughes, D. W.; Li, Y.; Brown, E. D. Small-Molecule Screening Made Simple for a Difficult Target with a Signaling Nucleic Acid Aptamer That Reports on Deaminase Activity. *Angew. Chem. Int. Ed.* **2006**, 45, 5648-5652.
- (5) Xing, X. J.; Liu, X. G.; Yue, H.; Luo, Q. Y.; Tang, H. W.; Pang, D. W. Graphene Oxide Based Fluorescent Aptasensor for Adenosine Deaminase Detection Using Adenosine as the Substrate. *Biosens. Bioelectron.* **2012**, 37, 61-67.
- (6) Sanchez, J. J.; Monaghan, G.; Borsting, C.; Norbury, G.; Morling, N.; Gaspar, H. B. Carrier Frequency of a Nonsense Mutation in the Adenosine Deaminase (ADA) Gene Implies a High Incidence of ADA-Deficient Severe Combined Immunodeficiency (SCID) in Somalia and a Single, Common Haplotype Indicates Common Ancestry. *Ann. Hum. Genet.* **2007**, 71, 336-347.
- (7) Mehmet, i.; Seyfettin, K.; Yilmaz, F. M.; Ömer, B.; Gülsen, Y.; Osman, Y.; Emre, Y.; Zeynel Abidin, Ö. Serum Adenosine Deaminase Levels in Pancreatic Diseases. *Pancreatology* **2007**, 7, 526-530.
- (8) Zhang, M.; Guo, S. M.; Li, Y. R.; Zuo, P.; Ye, B. C. A Label-Free Fluorescent Molecular Beacon Based on DNA-Templated Silver Nanoclusters for Detection of Adenosine and Adenosine Deaminase. *Chem. Commun.* **2012**, 48, 5488-5490.

- (9) Zhang, K.; Wang, K.; Xie, M. H.; Zhu, X.; Xu, L.; Yang, R. L.; Huang, B.; Zhu, X. L. DNA-Templated Silver Nanoclusters Based Label-Free Fluorescent Molecular Beacon for the Detection of Adenosine Deaminase. *Biosens. Bioelectron.* **2014**, *52*, 124-128.
- (10) Zhao, W. A.; Chiuman, W.; Lam, J. C. F.; McManus, S. A.; Chen, W.; Cui, Y. G.; Pelton, R.; Brook, M. A.; Li, Y. F. DNA Aptamer Folding on Gold Nanoparticles: From Colloid Chemistry to Biosensors. *J. Am. Chem. Soc.* **2008**, *130*, 3610-3618.
- (11) He, F.; Yu, M. H.; Wang, S. Magnetically Assisted Fluorescence Ratiometric Assays for Adenosine Deaminase Using Water-Soluble Conjugated Polymers. *Chin. Sci. Bull.* **2009**, *54*, 1340-1344.
- (12) Zhang, K.; Zhu, X.; Wang, J.; Xu, L.; Li, G. Strategy to Fabricate an Electrochemical Aptasensor: Application to the Assay of Adenosine Deaminase Activity. *Anal. Chem.* **2010**, *82*, 3207-3211.
- (13) Tuerk, C.; Gold, L. Systematic Evolution of Ligands by Exponential Enrichment: Rna Ligands to Bacteriophage T4 DNA Polymerase. *Science* **1990**, *249*, 505-510.
- (14) Jhaveri, S.; Rajendran, M.; Ellington, A. D. In Vitro Selection of Signaling Aptamers. *Nat. Biotechnol.* **2000**, *18*, 1293-1297.
- (15) Xu, W. C.; Lu, Y. Label-Free Fluorescent Aptamer Sensor Based on Regulation of Malachite Green Fluorescence. *Anal. Chem.* **2010**, *82*, 574-578.
- (16) Wang, J.; Wang, L. H.; Liu, X. F.; Liang, Z. Q.; Song, S. P.; Li, W. X.; Li, G. X.; Fan, C. H. A Gold Nanoparticle-Based Aptamer Target Binding Readout for ATP Assay. *Adv. Mater.* **2007**, *19*, 3943-3946.

- (17) Zhang, J.; Wang, L. H.; Pan, D.; Song, S. P.; Boey, F. Y. C.; Zhang, H.; Fan, C. H. Visual Cocaine Detection with Gold Nanoparticles and Rationally Engineered Aptamer Structures. *Small* **2008**, *4*, 1196-1200.
- (18) Misra, A.; Kumar, P.; Gupta, K. C. Synthesis of Hairpin Probe Using Deoxyguanosine as a Quencher: Fluorescence and Hybridization Studies. *Anal. Biochem.* **2007**, *364*, 86-88.
- (19) Marras, S. A.; Kramer, F. R.; Tyagi, S. Efficiencies of Fluorescence Resonance Energy Transfer and Contact-Mediated Quenching in Oligonucleotide Probes. *Nucleic Acids Res.* **2002**, *30*, e122.
- (20) O'Neill, M. A.; Barton, J. K. Effects of Strand and Directional Asymmetry on Base-Base Coupling and Charge Transfer in Double-Helical DNA. *Proc. Natl. Acad. Sci. U.S.A.* **2002**, *99*, 16543-16550.
- (21) Kelley, S. O.; Barton, J. K. Electron Transfer between Bases in Double Helical DNA. *Science* **1999**, *283*, 375-381.
- (22) Johansson, M. K.; Fidler, H.; Dick, D.; Cook, R. M. Intramolecular Dimers: A New Strategy to Fluorescence Quenching in Dual-Labeled Oligonucleotide Probes. *J. Am. Chem. Soc.* **2002**, *124*, 6950-6956.
- (23) Zhang, Y. W.; Tian, J. Q.; Zhai, J. F.; Luo, Y. L.; Wang, L.; Li, H. L.; Sun, X. P. Fluorescence-Enhanced Potassium Ions Detection Based on Inherent Quenching Ability of Deoxyguanosines and K⁺-Induced Conformational Transition of G-Rich ssDNA from Duplex to G-Quadruplex Structures. *J. Fluoresc.* **2011**, *21*, 1841-1846.
- (24) Wang, W. H.; Jin, Y.; Zhao, Y. N.; Yue, X. F.; Zhang, C. X. Single-Labeled Hairpin Probe for Highly Specific and Sensitive Detection of Lead(II) Based on the Fluorescence Quenching of Deoxyguanosine and G-Quartet. *Biosens. Bioelectron.* **2013**, *41*, 137-142.

- (25) Wang, L.; Tian, J.; Li, H.; Zhang, Y.; Sun, X. A Novel Single-Labeled Fluorescent Oligonucleotide Probe for Silver(I) Ion Detection Based on the Inherent Quenching Ability of Deoxyguanosines. *Analyst* **2011**, *136*, 891-893.
- (26) Kimura, T.; Kawai, K.; Majima, T. Fluorescence Properties of 2-Aminopurine-Cytidine-7-Deazaguanine (5'-ApCdG-3') Trimer in B- and Z-DNA. *Chem. Commun.* **2004**, 268-269.
- (27) O'Neill, M. A.; Barton, J. K. 2-Aminopurine: A Probe of Structural Dynamics and Charge Transfer in DNA and DNA:RNA Hybrids. *J. Am. Chem. Soc.* **2002**, *124*, 13053-13066.
- (28) Lewis, F. D.; Wu, T.; Zhang, Y.; Letsinger, R. L.; Greenfield, S. R.; Wasielewski, M. R. Distance-Dependent Electron Transfer in DNA Hairpins. *Science* **1997**, *277*, 673-676.
- (29) Thomas, S. W., 3rd; Joly, G. D.; Swager, T. M. Chemical Sensors Based on Amplifying Fluorescent Conjugated Polymers. *Chem. Rev.* **2007**, *107*, 1339-1386.
- (30) Zhu, C.; Liu, L.; Yang, Q.; Lv, F.; Wang, S. Water-Soluble Conjugated Polymers for Imaging, Diagnosis, and Therapy. *Chem. Rev.* **2012**, *112*, 4687-4735.
- (31) Song, J.; Zhang, J.; Lv, F.; Cheng, Y.; Wang, B.; Feng, L.; Liu, L.; Wang, S. Multiplex Detection of DNA Mutations by the Fluorescence Fingerprint Spectrum Technique. *Angew. Chem. Int. Ed.* **2013**, *52*, 13020-13023.
- (32) Duan, X.; Liu, L.; Feng, F.; Wang, S. Cationic Conjugated Polymers for Optical Detection of DNA Methylation, Lesions, and Single Nucleotide Polymorphisms. *Acc. Chem. Res.* **2010**, *43*, 260-270.
- (33) Ho, H. A.; Najari, A.; Leclerc, M. Optical Detection of DNA and Proteins with Cationic Polythiophenes. *Acc. Chem. Res.* **2008**, *41*, 168-178.
- (34) Liu, B.; Bazan, G. C. Synthesis of Cationic Conjugated Polymers for Use in Label-Free DNA Microarrays. *Nat. Protoc.* **2006**, *1*, 1698-1702.

- (35) Zhang, J. Y.; Xing, B. L.; Song, J. Z.; Zhang, F.; Nie, C. Y.; Jiao, L.; Liu, L. B.; Lv, F. T.; Wang, S. Associated Analysis of DNA Methylation for Cancer Detection Using CCP-Based FRET Technique. *Anal. Chem.* **2014**, *86*, 346-350.
- (36) Wang, L. H.; Pu, K. Y.; Li, J.; Qi, X. Y.; Li, H.; Zhang, H.; Fan, C. H.; Liu, B. A Graphene-Conjugated Oligomer Hybrid Probe for Light-up Sensing of Lectin and Escherichia Coli. *Adv. Mater.* **2011**, *23*, 4386-4391.
- (37) Xu, H.; Gao, S. L.; Yang, Q.; Pan, D.; Wang, L. H.; Fan, C. H. Amplified Fluorescent Recognition of G-Quadruplex Folding with a Cationic Conjugated Polymer and DNA Intercalator. *Acs Appl. Mater. Interfaces* **2010**, *2*, 3211-3216.
- (38) Kim, I. B.; Wilson, J. N.; Bunz, U. H. Mannose-Substituted Ppes Detect Lectins: A Model for Ricin Sensing. *Chem. Commun.* **2005**, 1273-1275.
- (39) Pinto, M. R.; Schanze, K. S. Amplified Fluorescence Sensing of Protease Activity with Conjugated Polyelectrolytes. *Proc. Natl. Acad. Sci. U.S.A.* **2004**, *101*, 7505-7510.
- (40) Tang, Y. L.; Liu, Y.; Cao, A. Strategy for Sensor Based on Fluorescence Emission Red Shift of Conjugated Polymers: Applications in pH Response and Enzyme Activity Detection. *Anal. Chem.* **2013**, *85*, 825-830.
- (41) Jones, R. M.; Lu, L.; Helgeson, R.; Bergstedt, T. S.; McBranch, D. W.; Whitten, D. G. Building Highly Sensitive Dye Assemblies for Biosensing from Molecular Building Blocks. *Proc. Natl. Acad. Sci. U.S.A.* **2001**, *98*, 14769-14772.
- (42) Liu, X. F.; Fan, Q. L.; Huang, W. DNA Biosensors Based on Water-Soluble Conjugated Polymers. *Biosens. Bioelectron.* **2011**, *26*, 2154-2164.

- (43) Tang, Y.; Feng, F.; He, F.; Wang, S.; Li, Y.; Zhu, D. Direct Visualization of Enzymatic Cleavage and Oxidative Damage by Hydroxyl Radicals of Single-Stranded DNA with a Cationic Polythiophene Derivative. *J. Am. Chem. Soc.* **2006**, *128*, 14972-14976.
- (44) Liu, B.; Wang, S.; Bazan, G. C.; Mikhailovsky, A. Shape-Adaptable Water-Soluble Conjugated Polymers. *J. Am. Chem. Soc.* **2003**, *125*, 13306-13307.
- (45) Lakowicz, J. R., *Principles of Fluorescence Spectroscopy*. 3rd ed.; Kluwer Academic/Plenum Publishers: New York: 2006.
- (46) Wang, C.; Tang, Y.; Liu, Y.; Guo, Y. Water-Soluble Conjugated Polymer as a Platform for Adenosine Deaminase Sensing Based on Fluorescence Resonance Energy Transfer Technique. *Anal. Chem.* **2014**, *86*, 6433-6438.
- (47) Armand, R.; Channon, J. Y.; Kintner, J.; White, K. A.; Miselis, K. A.; Perez, R. P.; Lewis, L. D. The Effects of Ethidium Bromide Induced Loss of Mitochondrial DNA on Mitochondrial Phenotype and Ultrastructure in a Human Leukemia T-Cell Line (MOLT-4 Cells). *Toxicol. Appl. Pharmacol.* **2004**, *196*, 68-79.
- (48) Crockett, A. O.; Wittwer, C. T. Fluorescein-Labeled Oligonucleotides for Real-Time Pcr: Using the Inherent Quenching of Deoxyguanosine Nucleotides. *Anal. Biochem.* **2001**, *290*, 89-97.
- (49) Zhang, K.; Yang, Q.; Zhang, J.; Fu, L.; Zhou, Y.; Wu, B.; Xie, M.; Huang, B. An Enzyme Substrate Binding Aptamer Complex Based Time-Resolved Fluorescence Sensor for the Adenosine Deaminase Detection. *Biosens. Bioelectron.* **2013**, *42*, 87-92.
- (50) Feng, T. T.; Ma, H. M. Fluorescence Sensing of Adenosine Deaminase Based on Adenosine Induced Self-Assembly of Aptamer Structures. *Analyst* **2013**, *138*, 2438-2442.

- (51) Zhang, K.; Xie, M.; Zhou, B.; Hua, Y.; Yan, Z.; Liu, H.; Guo, L. N.; Wu, B.; Huang, B. A New Strategy Based on Aptasensor to Time-Resolved Fluorescence Assay for Adenosine Deaminase Activity. *Biosens. Bioelectron.* **2013**, *41*, 123-128.
- (52) Cash, K. J.; Ricci, F.; Plaxco, K. W. A General Electrochemical Method for Label-Free Screening of Protein-Small Molecule Interactions. *Chem. Commun.* **2009**, 6222-6224.

Graphic for TOC

