

High-throughput analysis and protein engineering using microcapillary arrays

Bob Chen¹, Sungwon Lim^{1,7}, Arvind Kannan^{2,7}, Spencer C Alford^{1,7}, Fanny Sunden³, Daniel Herschlag^{3,4}, Ivan K Dimov⁵, Thomas M Baer^{6*} & Jennifer R Cochran^{1,2*}

We describe a multipurpose technology platform, termed μ SCALE (microcapillary single-cell analysis and laser extraction), that enables massively parallel, quantitative biochemical and biophysical measurements on millions of protein variants expressed from yeast or bacteria. μ SCALE spatially segregates single cells within a microcapillary array, enabling repeated imaging, cell growth and protein expression. We performed high-throughput analysis of cells and their protein products using a range of fluorescent assays, including binding-affinity measurements and dynamic enzymatic assays. A precise laser-based extraction method allows rapid recovery of live clones and their genetic material from microcapillaries for further study. With μ SCALE, we discovered a new antibody against a clinical cancer target, evolved a fluorescent protein biosensor and engineered an enzyme to reduce its sensitivity to its inhibitor. These protein analysis and engineering applications each have unique assay requirements and different host organisms, highlighting the flexibility and technical capabilities of the μ SCALE platform.

Over the past decade, high-throughput technologies have allowed researchers to gain unprecedented insights into intrinsically complex and interconnected biological systems. As examples, whole-genome sequencing has enabled the identification of crucial genes and mutations underlying disease pathophysiology^{1,2}, DNA microarrays have been used to elucidate transcription patterns involved in healthy and diseased states³, and large-scale proteomics methods have helped map the connectivity of cell signaling networks that orchestrate responses to growth factors and other external stimuli⁴. In contrast, analogously powerful approaches for rapidly and deeply interrogating the sequence-structure-activity relationship of proteins, with functional read-outs that span a range of biophysical and biochemical measurements, have lagged because of technical challenges. Here we describe a new technology platform that addresses this need and demonstrate its capabilities and breadth through applications on three distinct protein classes: antibody therapeutics, fluorescent protein biosensors and enzymes.

Protein engineers rely heavily on directed evolution, a powerful method that iterates rounds of library mutagenesis and screening⁵. In a directed-evolution experiment, randomly generated protein libraries are mined for variants with desirable characteristics, such as high-affinity binding to a target of interest⁶, stability⁷, fluorescence⁸ or enzymatic activity⁹. Maintaining a genotype-to-phenotype linkage is a fundamental requirement for any directed-evolution effort, allowing protein variants to be identified through their corresponding DNA sequence following a screen. Genotype-to-phenotype linkages are most easily preserved in screens that probe for protein binding partners. For example, genetic fusion of protein variants to microbes, phage or protein translation machinery enables rapid identification of target binders from large protein libraries (10^7 – 10^{14} variants) using fluorescence-activated cell sorting (FACS) or panning methods^{10,11}.

For protein-engineering applications that extend beyond binding interactions, spatial segregation establishes the genotype-to-phenotype linkage. For this purpose, researchers express and assay individual protein variants in microwell plates through phenotypic

screens based on colony selection in Petri dishes. Although liquid-handling robots have eased the labor required, such engineering endeavors are generally limited in throughput to 10^3 – 10^5 variants in a typical screen¹². These library sizes are relatively small and restrictive in comparison to the vast amino acid search space available for a typical protein¹³. Oil-water emulsion droplets generated in bulk or in microfluidic chips have enabled high-throughput enzyme engineering applications with libraries of 10^7 – 10^8 variants^{14–18}. Additionally, two past efforts used miniaturized microwell arrays to increase throughput for enzyme engineering^{19,20}. Notably, one effort described an enzyme-screening platform centered on reusable plates with 10^6 wells, achieving throughputs of up to 10^7 assays per day¹⁹. Improvements in handling or cell retrieval methods would facilitate widespread application of these technologies to protein characterization and engineering efforts.

Here, we describe μ SCALE (microcapillary single-cell analysis and laser extraction), a multipurpose platform technology capable of interrogating a dense array of millions of spatially segregated single cells or their protein products within minutes. A key feature of μ SCALE is the ability to isolate target cells after analysis from the microcapillary array using a precise laser-based extraction technique. Previous studies demonstrate the potential utility of microchamber, microcapillary or microwell arrays for single enzyme characterization^{21,22}, digital ELISA²³, selection of antibody-producing cells^{24,25}, and isolation of mammalian cells²⁶ or circulating tumor cells²⁷. Whereas these applications focused mainly on time-resolved kinetic analysis or cell sorting, we used μ SCALE to achieve both capabilities simultaneously. To highlight the flexibility and technical capabilities of our technology, we describe three distinct protein-analysis and engineering applications, using libraries expressed in yeast or bacteria.

RESULTS

Microcapillary single-cell analysis and laser extraction

We describe the overall concept and workflow of the μ SCALE platform in **Figure 1a,b** and **Supplementary Figure 1**. One core

¹Department of Bioengineering, Stanford University, Stanford, California, USA. ²Department of Chemical Engineering, Stanford University, Stanford, California, USA. ³Department of Biochemistry, Stanford University, Stanford, California, USA. ⁴Department of Chemistry, Stanford University, Stanford, California, USA. ⁵Institute for Stem Cell Biology and Regenerative Medicine, Stanford University, Stanford, California, USA. ⁶Stanford Photonics Research Center, Stanford University, Stanford, California, USA. ⁷These authors contributed equally to this work. *e-mail: tmbaer@stanford.edu, jennifer.cochran@stanford.edu

component of the technology is a dense glass-substrate array of millions of spatially segregated, high-aspect ratio microcapillaries (1-mm thick, 10 μm or 20 μm in diameter) (Fig. 1c). We apply a cell suspension mixed with magnetic microparticles to the microcapillary array by pipetting. The array is unsealed on the bottom, yet the liquid sample is held in place by surface tension. While the cells randomly distribute into the array following a Poisson distribution, the expected and observed means differ by two- to three-fold, probably because of the high aspect ratio of the microcapillaries (Supplementary Fig. 2), such that some cells might not travel within the imaging plane. The passive nature of the filling process results in a uniform meniscus level across the entire array. This uniformity, coupled with gravitational sedimentation of the loaded cells, simplifies establishment of an imaging focus plane without the need for autofocus.

We have demonstrated a variety of biochemical assays using the μSCALE platform (Fig. 1d). In each case, we first load bacteria or yeast cells expressing a protein of interest into a microcapillary array. The humid environment of the array permits time-resolved experimental measurements or cell growth over a period of days, if desired. We then image the microcapillaries using a high-numerical aperture microscope objective illuminated using a band-pass-filtered high-intensity arc lamp in an epi-illumination configuration. A high-sensitivity, cooled charge-coupled device (CCD) camera detects the fluorescence emission through a band-pass filter. Following established guidelines for quantitative wide-field microscopy, we measured the intercapillary variability in fluorescence signals detected from the array and found it to be comparable to other high-throughput methods²⁸ (Supplementary Fig. 3). We routinely image the arrays with multiple emission wavelengths at a rate of approximately 10,000 microcapillaries per second and use standard image-processing functions in MATLAB for analysis. The high resolution of the images enable us to distinguish promising microcapillary contents from cell debris and other false fluorescence signals via quantified metrics (size, roundness and fluorescence intensity) or by visual inspection.

After analysis, we recover the contents of individual microcapillaries of interest by laser extraction. Alternatively, we can designate multiple microcapillaries to be extracted (termed pooled extraction) at a rate of approximately two microcapillaries per second. A pulsed ultraviolet (UV) laser interacts with the magnetic particles to disrupt the surface tension within a single microcapillary, extracting its contents without affecting its neighbors (Fig. 1b,e).

We eject the contents of the microcapillary onto a capture surface below the array; this enables us to perform imaging to confirm extraction and to carry out additional morphological studies. To refine this technique, we varied the laser intensity and titrated the magnetic bead




Figure 1 | μSCALE overview. (a) Platform workflow. (1) A library of protein variants is mixed with opaque microbeads. (2) The mixture is pipetted into the array at a concentration that results in a mean single-cell occupancy. A biochemical assay is performed using fluorescence as a read-out. (3) The array is imaged via fluorescence microscopy. (4) The fluorescence intensity of each microcapillary is quantified, and desired clones are isolated with a laser-based extraction method. (5) The extracted cells are cultured. (6) Cells are lysed, and the plasmid is recovered. (b) Laser-based extraction method. A UV laser is positioned and pulsed, disrupting the surface tension of the microcapillary and ejecting its contents onto a capture surface below the array. (c) Microcapillary array, with penny for scale. (d) Three assays done with the μSCALE platform. Left, binding interactions of a yeast surface-displayed scFv library pre-stained with fluorescent antibodies. Middle, fluorescent protein variants expressed in *E. coli* clones cultured within the microcapillaries. Right, real-time kinetic measurements on a library of yeast surface-displayed enzyme variants that convert a substrate to a fluorescent product. (e) Extraction of two neighboring microcapillaries loaded with fluorescent particles. Left, the first (red circle) and second (yellow circle) capillaries to be extracted; middle left, the first extraction (red arrow), which leaves neighboring microcapillaries undisturbed; middle right, the isolated particle on the capture surface below the array; right, the second successful microcapillary extraction (yellow arrow).

concentrations to establish a range of conditions that achieved complete and specific microcapillary extraction (Supplementary Fig. 4).

Analysis and screening of protein binding interactions

As an initial demonstration, we used μSCALE to measure known protein-protein binding interactions (Fig. 2a,b) and subsequently

to rapidly screen and identify antibody binders against a clinical target (Fig. 2c,d). Growth arrest–specific-6 (Gas6) binds to the Axl receptor tyrosine kinase and mediates immune function, blood coagulation and tumor cell invasion and migration^{29,30}. We first evaluated the ability of μ SCALE to distinguish a wild-type (WT) Gas6 binding variant of the Axl receptor and a nonbinding variant on the surface of yeast cells. We used two-color fluorescence imaging to quantify yeast expression levels (assessed by fluorescence-labeled antibodies to a C-terminal c-myc epitope tag) and ligand binding (assessed with an antibody to a His₆ tag on the ligand). We observed a clear distinction between the WT and nonbinding Axl variant yeast populations (Fig. 2a,b). Next, to characterize the ability of our approach to isolate rare clones of interest, we screened mock libraries of yeast-displayed proteins in ligand-binding/nonbinding clone ratios spanning 1:10 to 1:100,000 (Fig. 2b, Supplementary Fig. 5 and Supplementary Table 1). Individual yeast cells were extracted from the array, cultured and evaluated by DNA recovery and sequencing. Across the six mock libraries, we correctly identified 37 out of 39 extracted clones as binding variants, achieving a single-pass enrichment of 100,000-fold in the most stringent condition. Additionally, we correctly confirmed all 19 extracted clones selected as nonbinders. These mock library screens validated the ability of μ SCALE to differentiate rare functional clones within a large population of nonfunctional variants.

We next applied μ SCALE to screen a large, naive library of yeast-displayed single-chain variable fragments (scFvs) to identify a variant that binds to Gas6 with high affinity. We first cleared the library of scFv binders against screening reagents using magnetic-activated cell sorting (MACS), followed by a round of MACS using Gas6-coated magnetic beads; this approach reduced the theoretical library size to 2.8×10^6 clones. We then enriched the resulting pool of scFv variants for binders to Gas6 using two rounds of screening on the μ SCALE platform (Fig. 2c). In the first round, we added a relatively low concentration of Gas6 (33 nM) to the scFv library and allowed it to reach binding equilibrium. Using pooled extraction, we isolated approximately 150 microcapillaries containing yeast with the highest Gas6-binding levels, normalized to the amount of scFv expression (Fig. 2c). We cultured this extracted yeast pool and induced scFv expression, and carried out a second μ SCALE screen under more stringent conditions (10 nM Gas6). In the second screen, we isolated 15 individual microcapillaries and sequenced the extracted contents (Fig. 2c and Supplementary Fig. 6e). Flow cytometry validated the enrichment achieved with each screening step (Supplementary Fig. 6). Retrospective examination of the μ SCALE screen revealed that the five microcapillaries with the highest observed Gas6-binding levels (normalized to scFv expression) contained the same clone (Fig. 2c). This resulting scFv variant bound to Gas6 with a K_D of 130 ± 30 nM (Fig. 2d), comparable to other binders mined from naive scFv libraries³¹.

Cell growth and screening of fluorescent proteins

Next, we demonstrated the use of μ SCALE to engineer a fluorescent protein variant, expanding the color palette of available biosensors and highlighting the compatibility of the platform




Figure 2 | High-throughput screening of binding proteins using μ SCALE. (a) Two-parameter scatter plots (left), where each dot represents a microcapillary ($n = 5,000$ microcapillaries per condition) and microscope images (right) of the WT Axl (red) and the nonbinding Axl variant (blue). (b) Scatter plot from a representative mock library screen (ratio of WT/nonbinding variant = 1:10,000, $n = 147,849$ microcapillaries). (c) Two rounds (sorts 1 and 2) of μ SCALE screening of a yeast surface-displayed naïve scFv library for Gas6 binders. Sort 1, automated pooled extraction of 143 capillaries with a user-drawn gate (red outline, left); sort 2, manual single-cell extraction of 15 capillaries (right). Red dots indicate extracted clones; red dots with black outline indicate clones with highest binding to Gas6. (d) Gas6 binding curve of the yeast-displayed scFv identified by μ SCALE, represented as the fraction of Gas6 bound versus the concentration added. Error bars indicate s.d. of three technical replicates.

with *Escherichia coli* libraries. Traditionally, fluorescent proteins are engineered by screening individual *E. coli* colonies grown on agar plates to identify desired color hues and/or fluorescence brightness³². Although straightforward, this low-throughput method relies on basic optics and laborious colony picking, which is time consuming and subjective, and hampers recovery of the most desirable variants. We first established the ability to grow spatially segregated cultures from individual *E. coli* cells in the microcapillary array. We loaded *E. coli* suspensions expressing GFP into the array at a concentration approximating 1 cell per capillary on average, incubated the array at 37 °C and imaged at regular time intervals. The resulting time course showed robust clonal expansion and fluorescent protein production, with *E. coli* growth confined to create a dense array of segregated cultures (Fig. 3a).

We then applied μ SCALE to generate a hue-shifted, new color variant of a dimerization-dependent fluorescent protein (ddFP) (Supplementary Fig. 7). We used a red dimerization-dependent fluorescent protein (ddRFP) as a template and introduced a blue-shifting and chromophore-disrupting variation (M66T) that eliminates fluorescence^{8,33} (Fig. 3b). To rescue and enhance fluorescence, we performed three sequential rounds of directed evolution, using μ SCALE to screen libraries created by error-prone PCR mutagenesis and expressed in *E. coli*. For directed evolution rounds 1 and 2, we isolated approximately 12 clones showing weak fluorescence and used them to generate a gene template mix for a subsequent round of mutagenesis. By round 3, a fraction of the library showed fluorescence levels substantially higher than the library mean. We extracted, cultured and characterized *E. coli* from the ten microcapillaries with the brightest fluorescence (Fig. 3c). The brightest

fluorescent protein variant from round 3, designated dimerization-dependent orange fluorescent protein (ddOFP), acquired five mutations in addition to the original M66T mutation (Supplementary Fig. 8a) that conferred a substantial increase in fluorescence intensity over the best mutant from round 2 and the 'parent' M66T clone (Fig. 3d). ddOFP has absorption and emission maxima at 535 nm and 565 nm, respectively (Supplementary Fig. 8b), filling the spectral gap between ddRFP and ddYFP variants (Supplementary Fig. 9). In addition, ddOFP showed comparable brightness and fluorogenic properties to other ddFPs^{8,34} (Supplementary Table 2), making it a suitable module for biosensing applications such as fluorescent protein exchange³⁵.

Improved enzyme catalysts

We next used μ SCALE to measure enzyme kinetics and identify an improved biocatalyst as a final demonstration of its capabilities and versatility. Decades of published biochemical insights for alkaline phosphatase qualified this enzyme as a model system for these studies³⁶. We measured catalytic activity using the substrate 9H-(1,3-dichloro-9,9-dimethylacridin-2-one-7-yl) phosphate (DDAOP), which yields a fluorescent product upon hydrolysis (Supplementary Fig. 10–12). The Michaelis–Menten kinetics and phosphate-inhibition profiles of yeast-displayed WT alkaline phosphatase yielded values within two-fold of those measured from purified enzyme harvested from *E. coli* (Supplementary Fig. 13). Moreover, yeast-displayed constructs preserved the kinetic differences between WT alkaline phosphatase and a previously characterized R166S mutant, which exhibits a significant reduction in the apparent second-order rate constant (k_{cat}/K_M) for the DDAOP substrate³⁷ (Supplementary Table 3).

As further validation, we used μ SCALE to measure single-cell reaction time courses for yeast-displayed WT alkaline phosphatase and the R166S mutant, demonstrating that the platform can provide time-resolved, kinetic information on the activity of enzyme variants. The microcapillaries maintained robust spatial segregation of yeast-displayed enzymes and accumulating product 7-hydroxy-9H-(1,3-dichloro-9,9-dimethylacridin-2-one) (DDAO) throughout each reaction time course in the arrays, enabling facile and quantitative kinetic discrimination of microcapillaries containing highly active yeast-displayed enzymes from less-active neighbors or empty microcapillaries (Fig. 4a and Supplementary Fig. 14). We quantified approximately 5,000 single-cell traces for WT alkaline phosphatase or the R166S variant over 20-min or 3-h reactions, respectively. Histograms of the time-course slopes are shown in Figure 4b and define the activity distribution of each mutant in the microcapillary array. Despite cell-to-cell variability in enzymatic activity within each clonal population, which is probably due to heterogeneity in protein expression levels, the μ SCALE platform clearly delineated the R166S mutant from the WT enzyme on the basis of reaction kinetics.

We next demonstrated the ability of μ SCALE to identify improved enzyme variants. We created an alkaline phosphatase library by error-prone PCR mutagenesis and screened it using DDAOP substrate at a concentration of 1 μ M (four-fold lower than the apparent




Figure 3 | Engineering an orange-hued fluorescent protein with μ SCALE. (a) Growth of spatially segregated *E. coli* cultures expressing GFP. Images show a portion of a 20- μ m array, which was loaded with, on average, 1 bacterial cell per microcapillary, cultured at 37 °C for the indicated times and serially imaged. (b) Engineering strategy to generate ddOFP from the ddRFP template. A M66T mutation was introduced into the A copy of ddRFP, generating a largely nonfluorescent hue-shifted variant. Three rounds of mutagenesis and μ SCALE screening of protein libraries expressed in *E. coli* resulted in ddOFP, which has five mutations in addition to M66T. (c) Quantification of μ SCALE screen of the third *E. coli* library. The ten brightest microcapillaries (red) from the distribution of observed fluorescence intensities of the library variants were extracted from the array ($n = 305,279$ microcapillaries). (d) Iterative directed evolution results in variants with increased fluorescence intensity. Fluorescence emission spectra of *E. coli* crude lysates expressing M66T, a variant from round 2, and ddOFP from round 3. Inset, the increase in color hue of the cell pellets of *E. coli* expressing FP variants. a.u., arbitrary units.

K_M for WT alkaline phosphatase) in the presence of 15 μ M inorganic phosphate (four-fold higher than the measured K_i for WT alkaline phosphatase) (see Online Methods and Supplementary Tables 3 and 4). These conditions were expected to enrich for enzyme variants with high tolerance to inhibitor, a common objective in many diverse subfields of enzyme engineering^{38–40}. Using μ SCALE, we rapidly analyzed the catalytic activity of 3×10^5 enzyme variants. By single-cell extraction, we isolated 15 putative hits with enhanced activity relative to WT alkaline phosphatase (Fig. 4c). We then cultured and rescreened these isolated variants in bulk under the screening conditions described above. We selected three variants for sequencing and additional characterization on the basis of their significant rate accelerations relative to WT alkaline phosphatase in bulk biochemical assays (Fig. 4d). One variant, which contained a single substitution, D101G, showed substantially reduced sensitivity to inorganic phosphate, with a measured K_i ten-fold higher than that of the WT enzyme (Fig. 4e and Supplementary Fig. 15). The D101G mutant maintained its improved performance relative to the WT enzyme upon soluble expression and purification from *E. coli* (Supplementary Table 5). The higher basal levels of activity of the two other yeast-displayed variants compared to WT alkaline phosphatase (Supplementary Fig. 16) can be explained in part by their increased yeast surface-expression levels (Supplementary Fig. 17). However, both variants were inactive when purified from *E. coli* (Supplementary Tables 5 and 6), suggesting that the beneficial mutations acquired by these clones are specific to the yeast surface environment and incompatible with recombinant expression in bacteria.




Figure 4 | High-throughput kinetic measurements and enzyme library screening with μ SCALE.

(a) Representative images from a time-resolved enzymatic assay. Four microcapillaries harboring yeast cells displaying active enzyme (white numbers) and a nearby empty capillary (dotted circle) are shown for comparison. Image contrast artificially enhanced for clarity. (b) Single-cell kinetics profiles of yeast-displayed WT alkaline phosphatase and an R166S mutant with reduced catalytic efficiency. Reactions were carried out with 10 μ M DDAOP substrate and were recorded for up to 3 h. Slopes of the linear portion of each time course were binned to yield the distribution of rates. (WT, $n = 618$ traces; R166S, $n = 812$ traces). (c) Quantification of μ SCALE screen of randomly mutated alkaline phosphatase library. The distribution of observed enzymatic activity during the screen is shown in blue ($n = 309,895$ microcapillaries). The 15 most active cells (red vertical lines) were extracted. (d) Relative rates of WT alkaline phosphatase and the three most active isolated variants. Rates were measured with yeast-displayed constructs in bulk biochemical assays under identical conditions to those used in the screen and were normalized to the WT value. (e) Phosphate inhibition curves for WT alkaline phosphatase and the three variants isolated. Reactions were carried out with substrate concentrations at least ten-fold lower than the apparent K_M of each mutant. P_i , inorganic phosphate. Fits to a competitive inhibition model are overlaid (solid lines). Error bars indicate s.d. of three technical replicates (d,e).

DISCUSSION

Here we demonstrate three disparate protein-engineering applications using μ SCALE: isolation of a novel protein binder from a naïve antibody library, creation of a new fluorescent protein biosensor and identification of an improved enzyme variant. In all three examples, the μ SCALE platform afforded robust spatial segregation of microcapillary contents and the ability to recover single cells with a laser-based extraction method.

μ SCALE offers clear advantages over microwell plates and Petri dishes traditionally used to analyze the functional activity of protein variants within a library screen. In addition to greater sample throughput, the microcapillary arrays can contain up to 4 million separate *E. coli* cultures per square inch, five orders of magnitude more than in a standard Petri dish or microwell plate. As we demonstrate in our fluorescent protein-engineering example, clonal expansion of *E. coli* from a single cell into a culture provides signal amplification and the ability to measure the population mean, thereby reducing variability. Additionally, the μ SCALE technology platform enables real-time kinetic measurements, as we demonstrate in our example of alkaline phosphatase engineering. This capability for repeated imaging helps distinguish enzyme variants by their reaction kinetics rather than by fluorescence intensity at a single time point, potentially reducing false positives. However, the requirement for a fluorescent readout in μ SCALE screens necessitates the use of a surrogate fluorogenic substrate (such as DDAOP) in enzymatic assays, which may poorly approximate the desired substrate for downstream engineering applications.

In contrast, microwell plate-based assays support a broader range of possible surrogate substrates due to the choice of absorbance, fluorescence, or chemiluminescence readouts. Direct chromatographic detection of the desired product is also possible from a microwell plate-based screen, albeit with greatly reduced throughput⁴¹.

μ SCALE affords several additional distinctions compared to FACS-based library-screening methods, including the capability for direct cell imaging and temporal decoupling of cell analysis and sorting. Before the user sets the sorting parameters, μ SCALE allows imaging and analysis of the entire population of microcapillaries. In addition, μ SCALE enables robust isolation of rare clones with high precision and enrichment of up to 100,000-fold (Supplementary Table 1), in contrast to the modest enrichment ratios supported by FACS⁶. As such, efficient library screening and clone recovery is achievable in only 1 or 2 rounds of μ SCALE screening, as evidenced by our rapid isolation of an antibody fragment to the Gas6 ligand.

Whereas other microchamber technologies primarily rely on micromanipulators^{19,20}, air²⁵ or optical tweezers²⁶ to recover samples, we present an alternative laser-based mechanism of sample recovery. Laser-induced cavitation is a well-studied phenomenon frequently used by researchers to disrupt mammalian cell membranes for delivery or specific cell targeting^{42,43}. On the basis of these studies, in our application it is likely that the laser energy rapidly heats the surface of the magnetic microparticles, forming a cavitation bubble that disrupts the microcapillary meniscus

and empties the contents onto the extraction surface. Because most of the laser energy is absorbed by the magnetic beads rather than the cells, we observe high viability in cell lines commonly used for protein engineering (up to 80% and 100% survival for extracted yeast (*Saccharomyces cerevisiae*) and bacteria (*E. coli* and *Bacillus subtilis*), respectively; Supplementary Table 7). Given that optical cavitation has been used for membrane disruption and targeted cell killing, the extension of this cell recovery mechanism to more delicate cell types, such as mammalian cells, requires additional study.

In summary, μ SCALE provides a new approach for quantifying the functional activity of millions of protein variants and isolating desirable mutants from library screens. In addition to protein characterization and engineering, μ SCALE has broad applicability to high-throughput analysis of single cells and cell cultures. For example, μ SCALE has potential uses for probing cell-to-cell interactions, single-cell growth and signaling dynamics, gene-expression differences, recombinant host-cell expression levels or other applications requiring massively parallel biological measurements.

Received 24 August 2015; accepted 26 October 2015; published online 7 December 2015

METHODS

Methods and any associated references are available in the online version of the paper.

Accession codes. GenBank: Protein and DNA sequences of anti-Gas6 scFv, ddOFP, alkaline phosphatase D101G, alkaline phosphatase I16V-N145I-D294G and alkaline phosphatase N197S-T148P-S175R have been deposited under accession codes [KT454031](#), [KT454032](#), [KT454033](#), [KT454034](#) and [KT454035](#), respectively.

References

- Lander, E.S. Initial impact of the sequencing of the human genome. *Nature* **470**, 187–197 (2011).
- Rahman, N. Realizing the promise of cancer predisposition genes. *Nature* **505**, 302–308 (2014).
- Rhodes, D.R. & Chinnaiyan, A.M. Integrative analysis of the cancer transcriptome. *Nat. Genet.* **37** (suppl.): S31–S37 (2005).
- Zhang, Y. *et al.* Time-resolved mass spectrometry of tyrosine phosphorylation sites in the epidermal growth factor receptor signaling network reveals dynamic modules. *Mol. Cell. Proteomics* **4**, 1240–1250 (2005).
- Romero, P.A. & Arnold, F.H. Exploring protein fitness landscapes by directed evolution. *Nat. Rev. Mol. Cell Biol.* **10**, 866–876 (2009).
- Kariolis, M.S. *et al.* An engineered Axl 'decoy receptor' effectively silences the Gas6-Axl signaling axis. *Nat. Chem. Biol.* **10**, 977–983 (2014).
- Wu, I. & Arnold, F.H. Engineered thermostable fungal Cel6A and Cel7A cellobiohydrolases hydrolyze cellulose efficiently at elevated temperatures. *Biotechnol. Bioeng.* **110**, 1874–1883 (2013).
- Alford, S.C., Ding, Y., Simmen, T. & Campbell, R.E. Dimerization-dependent green and yellow fluorescent proteins. *ACS Synth. Biol.* **1**, 569–575 (2012).
- Chen, I., Dorr, B.M. & Liu, D.R. A general strategy for the evolution of bond-forming enzymes using yeast display. *Proc. Natl. Acad. Sci. USA* **108**, 11399–11404 (2011).
- Gai, S.A. & Wittrup, K.D. Yeast surface display for protein engineering and characterization. *Curr. Opin. Struct. Biol.* **17**, 467–473 (2007).
- Lee, C.V. *et al.* High-affinity human antibodies from phage-displayed synthetic Fab libraries with a single framework scaffold. *J. Mol. Biol.* **340**, 1073–1093 (2004).
- Martis, E. High-throughput screening: the hits and leads of drug discovery. *J. Appl. Pharm. Sci.* **01**, 02–10 (2011).
- Povolotskaya, I.S. & Kondrashov, F.A. Sequence space and the ongoing expansion of the protein universe. *Nature* **465**, 922–926 (2010).
- Agresti, J.J. *et al.* Ultrahigh-throughput screening in drop-based microfluidics for directed evolution. *Proc. Natl. Acad. Sci. USA* **107**, 4004–4009 (2010).
- Ostafe, R., Prodanovic, R., Nazor, J. & Fischer, R. Ultra-high-throughput screening method for the directed evolution of glucose oxidase. *Chem. Biol.* **21**, 414–421 (2014).
- Zinchenko, A. *et al.* One in a million: flow cytometric sorting of single cell-lysate assays in monodisperse picoliter double emulsion droplets for directed evolution. *Anal. Chem.* **86**, 2526–2533 (2014).
- Fischlechner, M. *et al.* Evolution of enzyme catalysts caged in biomimetic gel-shell beads. *Nat. Chem.* **6**, 791–796 (2014).
- Romero, P.A., Tran, T.M. & Abate, A.R. Dissecting enzyme function with microfluidic-based deep mutational scanning. *Proc. Natl. Acad. Sci. USA* **112**, 7159–7164 (2015).
- Lafferty, M. & Dycaco, M.J. GigaMatrix: an ultra high-throughput tool for accessing biodiversity. *J. Lab. Autom.* **9**, 200–208 (2004).
- Fukuda, T., Shiraga, S., Kato, M. & Yamamura, S. Construction of novel single-cell screening system using a yeast cell chip for nano-sized modified-protein-displaying libraries. *NanoBiotechnology* **1**, 105–111 (2005).
- Gorris, H.H. & Walt, D.R. Mechanistic aspects of horseradish peroxidase elucidated through single-molecule studies. *J. Am. Chem. Soc.* **131**, 6277–6282 (2009).
- Liebherr, R.B., Renner, M. & Gorris, H.H. A single molecule perspective on the functional diversity of *in vitro* evolved β -glucuronidase. *J. Am. Chem. Soc.* **136**, 5949–5955 (2014).
- Rissin, D.M. *et al.* Single-molecule enzyme-linked immunosorbent assay detects serum proteins at subfemtomolar concentrations. *Nat. Biotechnol.* **28**, 595–599 (2010).
- Love, J.C., Ronan, J.L., Grotenbreg, G.M., van der Veen, A.G. & Ploegh, H.L. A microengraving method for rapid selection of single cells producing antigen-specific antibodies. *Nat. Biotechnol.* **24**, 703–707 (2006).
- Fitzgerald, V. *et al.* Exploiting highly ordered subnanoliter volume microcapillaries as microtools for the analysis of antibody producing cells. *Anal. Chem.* **87**, 997–103 (2015).
- Kovac, J.R. & Voldman, J. Intuitive, image-based cell sorting using optofluidic cell sorting. *Anal. Chem.* **79**, 9321–9330 (2007).
- Gach, P.C., Attayek, P.J., Whittlesey, R.L., Yeh, J.J. & Allbritton, N.L. Micropallet arrays for the capture, isolation and culture of circulating tumor cells from whole blood of mice engrafted with primary human pancreatic adenocarcinoma. *Biosens. Bioelectron.* **54**, 476–483 (2014).
- Waters, J.C. Accuracy and precision in quantitative fluorescence microscopy. *J. Cell Biol.* **185**, 1135–1148 (2009).
- Lemke, G. & Rothlin, C.V. Immunobiology of the TAM receptors. *Nat. Rev. Immunol.* **8**, 327–336 (2008).
- Graham, D.K., DeRyckere, D., Davies, K.D. & Earp, H.S. The TAM family: phosphatidylinositol-sensing receptor tyrosine kinases gone awry in cancer. *Nat. Rev. Cancer* **14**, 769–785 (2014).
- Sidhu, S.S. *et al.* Phage-displayed antibody libraries of synthetic heavy chain complementarity determining regions. *J. Mol. Biol.* **338**, 299–310 (2004).
- Ai, H.-W., Baird, M.A., Shen, Y., Davidson, M.W. & Campbell, R.E. Engineering and characterizing monomeric fluorescent proteins for live-cell imaging applications. *Nat. Protoc.* **9**, 910–928 (2014).
- Shaner, N.C. *et al.* Improved monomeric red, orange and yellow fluorescent proteins derived from *Discosoma* sp. red fluorescent protein. *Nat. Biotechnol.* **22**, 1567–1572 (2004).
- Alford, S.C., Abdelfattah, A.S., Ding, Y. & Campbell, R.E. A fluorogenic red fluorescent protein heterodimer. *Chem. Biol.* **19**, 353–360 (2012).
- Ding, Y. *et al.* Ratiometric biosensors based on dimerization-dependent fluorescent protein exchange. *Nat. Methods* **12**, 195–198 (2015).
- Coleman, J. Structure and mechanism of alkaline phosphatase. *Annu. Rev. Biophys. Biomol. Struct.* **21**, 441–483 (1992).
- O'Brien, P.J. & Herschlag, D. Functional interrelationships in the alkaline phosphatase superfamily: phosphodiesterase activity of *Escherichia coli* alkaline phosphatase. *Biochemistry* **40**, 5691–5699 (2001).
- Alberstein, M., Eisenstein, M. & Abeliovich, H. Removing allosteric feedback inhibition of tomato 4-coumarate:CoA ligase by directed evolution. *Plant J.* **69**, 57–69 (2012).
- Yang, J.-S., Seo, S.W., Jang, S., Jung, G.Y. & Kim, S. Rational engineering of enzyme allosteric regulation through sequence evolution analysis. *PLoS Comput. Biol.* **8**, e1002612 (2012).
- Hu, X., Robin, S., O'Connell, S., Walsh, G. & Wall, J.G. Engineering of a fungal β -galactosidase to remove product inhibition by galactose. *Appl. Microbiol. Biotechnol.* **87**, 1773–1782 (2010).
- Coelho, P.S., Brustad, E.M., Kannan, A. & Arnold, F.H. Olefin cyclopropanation via carbene transfer catalyzed by engineered cytochrome P450 enzymes. *Science* **339**, 307–310 (2013).
- Pitsillides, C.M., Joe, E.K., Wei, X., Anderson, R.R. & Lin, C.P. Selective cell targeting with light-absorbing microparticles and nanoparticles. *Biophys. J.* **84**, 4023–4032 (2003).
- Wu, Y.-C. *et al.* Massively parallel delivery of large cargo into mammalian cells with light pulses. *Nat. Methods* **12**, 439–444 (2015).

Acknowledgments

The authors thank K.D. Wittrup and J. Van Deventer (MIT) and S. Sidhu (University of Toronto) for providing the yeast-displayed scFv library. This project was funded in part by the Stanford–Wallace H. Coulter Translational Partnership Award Program, the Siebel Stem Cell Institute and the Thomas and Stacey Siebel Foundation (to I.K.D.), the Stanford Photonics Research Center, a Hitachi America Faculty Scholar Award (to J.R.C.) and the US National Institutes of Health (GM49243 to D.H.). We acknowledge support from the National Science Foundation Graduate Fellowship Program (B.C. and A.K.), the Howard Hughes Medical Institute International Student Research Program (S.L.), a Fannie and John Hertz Foundation Graduate Fellowship (A.K.), the Stanford Bio-X Fellowship Program (S.L.), the Stanford Graduate Fellowship Program (B.C.) and the Stanford Dean's Fellowship Program (S.C.A.).

Author contributions

All authors conceived and designed experiments and analyzed data. T.M.B., I.K.D. and B.C. designed and established the μ SCALE platform. B.C., S.L. and I.K.D. established the μ SCALE workflow. S.L. and B.C. performed protein engineering for binding interactions. S.C.A. and B.C. performed protein engineering for fluorescent proteins. A.K. and B.C. performed enzyme characterization and engineering with guidance from F.S. and D.H. The manuscript was prepared by B.C., S.L., A.K., S.C.A. and J.R.C. with input from all coauthors.

Competing financial interests

The authors declare competing financial interests: details accompany the [online version of the paper](#).

Additional information

Any supplementary information, chemical compound information and source data are available in the [online version of the paper](#). Reprints and permissions information is available online at <http://www.nature.com/reprints/index.html>. Correspondence and requests for materials should be addressed to T.M.B. or J.R.C.

ONLINE METHODS

μSCALE instrumentation and software. μSCALE experiments were performed with a Veritas laser-capture microdissection (LCM) system (Arcturus), adapted with hardware and software modifications to enable microcapillary screening applications. The system contains an inverted fluorescence microscope with motorized stage and the Triton UV laser (diode-pumped Q-switched Nd:YLF laser, Spectra-Physics) (**Supplementary Fig. 1**). The laser system uses a wavelength of 349 nm, a focused beam diameter of 5 μm, and a pulse duration of 20 ns. For fluorescence imaging, an X-cite 120 illumination system (EXFO Photonic Solutions Inc.) was used along with the XF410 QMAX FITC and the XF406 QMAX red filter set (Omega Optical). All images were acquired with a 20× Plan Fluorite objective (numerical aperture: 0.45, CFI, WD: 7.4, Nikon) by an ORCA-ER cooled CCD camera (Hamamatsu).

Custom software, written in MATLAB code, was used to control the LCM microscope and laser. The software suite was used to automatically move the stage and objective to acquire multicolor fluorescent and bright-field images of every microcapillary. To quantify each microcapillary, image segmentation was performed using Otsu's method to threshold the chosen color image (frequently the bright-field image), creating a binary mask⁴⁴. This binary mask was used to quantify the fluorescence intensity at each region of interest. The quantified data were then filtered by size, roundness (eccentricity), and minimal fluorescence intensity with user-defined parameters to minimize false positives, such as cell debris or other fluorescent debris. These image analysis steps were performed with standard functions in the MATLAB Image Processing Toolbox (available upon request). The software was also used to retrieve cells from desired microcapillaries in two different modalities: single-capillary extraction and pooled extraction. In single-capillary extraction, the software was used to move the stage to center the selected microcapillary, allowing further examination and extraction of the desired individual cell onto a sterile capture surface. In pooled extraction, selection gates were established, and the software was used to sequentially move the stage and fire the laser to extract the contents of all specified microcapillaries onto a single capture surface, resulting in a population of enriched cells.

Microcapillary array preparation, loading and extraction. Microcapillary arrays (10 and 20 μm diameter, 1 mm thick; INCOM, Inc.) were sterilized in ethanol and dried. The loading side of the arrays was treated using a corona wand (BD-20AC Electro-Technic Products) to generate a hydrophilic surface, which facilitates loading. To achieve the maximum fraction of wells containing single cells, cell suspensions were diluted prior to loading according to Poisson statistics (~3,200 cells/μl in 20-μm arrays and ~12,800 cells/μl in 10-μm arrays). Specific loading conditions for each protein-engineering application are described in subsequent sections. Cell suspensions were mixed with magnetic microbeads (Life Technologies, 37002D) to a final bead concentration of 10 mg/ml and pipetted into the arrays. The magnetic microparticles are opaque and impart a brownish color to the cell suspensions (**Supplementary Fig. 1**). The addition of these microparticles uniformly occludes the fluorescence signal to a limited degree and prevents bright-field imaging of the contents of the microcapillary (**Supplementary Fig. 18**). A 2 mm slab of 1% weight/volume agarose was overlaid on the array to help prevent evaporation.

The Triton UV laser in the LCM system was used to extract the contents of desired capillaries. As described in **Supplementary Figure 4**, the laser power was adjusted so that the extraction efficiency is close to 100%. With our optimized extraction parameters, the laser operates for 18 ± 2 ms ($n = 5$ measurements), delivering a train of pulses at 2.5 kHz with a total energy of approximately 100 μJ. The microcapillary contents were extracted onto a glass cover slip, which was then placed in yeast or bacterial growth medium (liquid medium or agar plates) to propagate the extracted cells. After each experiment, the arrays were cleaned by removing their contents with a strong stream of distilled water, followed by brief sonication in 1 M NaOH, and storage in 100% ethanol prior to reuse. In our hands, the microcapillary arrays can be reused at least 10 times without observable deterioration.

Preparation of Axl protein constructs and libraries. DNA encoding human Axl Ig1 (amino acids Ala19–Pro131) and a nonbinding Axl variant (E59R, T77R) were cloned into the pCT yeast display plasmid between NheI and BamHI restriction sites⁶. Plasmid DNA was transformed into the *S. cerevisiae* strain EBY100 by electroporation for yeast surface-display studies. Soluble Gas6 was recombinantly expressed in human embryonic kidney (HEK) cells

using the FreeStyle Max 293 Expression System (Invitrogen) and purified as previously described⁶. The yeast surface-displayed naïve scFv library (~7 × 10⁸ variants) was generously provided by J. Van Deventer and K.D. Wittrup (Massachusetts Institute of Technology)⁴⁵ and S. Sidhu (University of Toronto). All yeast cells harboring the pCT display plasmid were grown in selective medium and induced as previously described⁴⁶.

Axl Ig1 mock library screening. For the Axl Ig1 mock libraries, induced yeast cells displaying wild-type (WT) and the nonbinding Axl variant were mixed at defined cell ratios of WT/nonbinding of 1:10, 1:100, 1:1,000, 1:10,000, and 1:100,000. Approximately 3–6 × 10⁶ yeast cells (depending on the mock library) were incubated at room temperature for 6 h with 1 nM purified Gas6 (ref. 6) in PBS with 1 mg/ml BSA (PBSA). Yeast were then washed and resuspended in PBSA containing a 1:250 dilution of chicken anti-c-Myc antibody (Life Technologies, A21281) for 45 min at 4 °C. To detect Gas6 binding, yeast were washed and resuspended in PBSA containing a 1:100 dilution of mouse anti-His IgG Hilyte Fluor 555 (Anaspec, 61250-H555) for 45 min at 4 °C. To enhance the binding signal, yeast were washed and resuspended in PBSA, and a 1:100 dilution of goat anti-mouse IgG Hilyte Fluor 555 (Anaspec, AS-28175-05-H555) was subsequently added for 30 min at 4 °C. Finally, to detect c-Myc tag expression, yeast were washed and resuspended in PBSA containing a 1:100 dilution of goat anti-chicken IgG Alexa Fluor 488 (Life Technologies, A11039) for 30 min at 4 °C. Labeled yeast were diluted to ~12,800 cells/μl, loaded on a 10 μm array, and analyzed for Gas6 binding with 475/40 nm excitation/510 nm long-pass emission filters, and for c-Myc cell surface expression with 525/45 nm excitation/565 nm long pass emission filters. Yeast cells predicted to express WT or the nonbinding Axl variant were extracted via single-cell sorting and cultured on agar plates containing selective yeast medium. To identify the extracted cells, plasmid DNA was recovered using a Zymoprep kit (Zymo Research Corporation), amplified by PCR and sequenced using the Sanger method (Sequetech).

scFv naïve library screening. The yeast-displayed scFv library was grown and induced for expression as previously described⁴⁵. The library was subjected to two rounds of magnetic-activated cell sorting (MACS) to reduce the diversity prior to μSCALE screening: a negative selection for cells that did not bind His-tag Isolation beads (Dynabeads, 10103D, Life Technologies) and a positive selection for cells that bind Gas6-coated magnetic heads. Both sorting rounds were performed according to the manufacturer's instructions (Dynabeads His-Tag isolation and pulldown protocol) and modeled after procedures previously described⁴⁶. Gas6-coated magnetic beads (4 mg) were prepared by incubating Dynabeads with a saturating amount of His-tagged Gas6 on a rotator for 2 h at 4 °C. These beads were washed using a magnetic holder (MPC-S, Dynal) with Binding/Wash Buffer (50 mM sodium phosphate, pH 8.0, 300 mM NaCl, and 0.01% Tween-20) to removed unconjugated Gas6. For sort 1, 8 × 10⁹ yeast cells were washed with pulldown buffer (3.25 mM sodium phosphate, pH 7.4, 70 mM NaCl, and 0.01% Tween-20), and incubated with non-coated His-tag Dynabeads in pulldown buffer for 2 h at 4 °C. After incubation, cells and beads were placed in a magnetic holder, and unbound cells were collected; these nonbinding yeast were transferred to fresh tubes and incubated with the prepared Gas6-coated beads for 2 h at 4 °C. Using the magnetic holder, the yeast and beads were washed several times with Binding/Wash Buffer. Gas6-bound yeast cells were eluted from the magnetic beads with His-elution buffer (150 mM imidazole, 25 mM sodium phosphate, pH 8.0, 150 mM NaCl, and 0.005% Tween-20) and grown in selective yeast medium. Using this protocol, MACS reduced the library to ~2 × 10⁶ variants, which corresponds to 0.4% of the original library size.

μSCALE was used for two sequential screening rounds of the MACS-reduced scFv library. Yeast were incubated at room temperature for 2 h in PBSA containing 33 nM Gas6 (sort 1) or 10 nM Gas6 (sort 2). Following incubation with Gas6, cells were treated with PBSA containing a 1:250 dilution of chicken anti-c-Myc antibody (Life Technologies, A21281) for 30 min at 4 °C. Cells were then incubated with PBSA containing a 1:100 dilution of mouse anti-His Tag IgG Hilyte Fluor 488 (Anaspec, 61250-H488) for 30 min at 4 °C. Finally, cells were incubated in PBSA containing a 1:100 dilution of both goat anti-mouse IgG Alexa Fluor 488 (Life Technologies, A11001) and goat anti-chicken IgG Alexa Fluor 555 (Life Technologies, A21437) for 30 min at 4 °C. Labeled yeast were diluted to ~12,800 cells/μl and loaded on a 10-μm array and analyzed for Gas6 binding using excitation-emission parameters described above.

For μ SCALE round 1, 143 capillaries with the highest Gas6 binding/c-Myc expression ratio were auto-extracted within a single pool, resulting in a population enriched for binding affinity to Gas6. In μ SCALE round 2, the top 15 capillaries were extracted via single-cell sorting. After round 2, plasmid DNA was recovered using a Zymoprep kit (Zymo Research Corporation), PCR amplified, and sequenced (Sequetech).

The Gas6 binding affinity of the best scFv clone was measured by incubating 10^5 induced yeast cells with varying concentrations of Gas6 in PBSA₂₀₀ (0.1% BSA + 200 mM NaCl in PBS) for 2 h at room temperature. Cells were incubated with PBSA₂₀₀ containing a 1:250 dilution of chicken anti-c-Myc antibody (Life Technologies, A21281) for 15 min at 4 °C. Secondary antibody labeling was carried out in PBSA₂₀₀ containing a 1:100 dilution of mouse anti-His Tag IgG Hilyte Fluor 488 (Anaspec, 61250-H488) and goat anti-chicken IgY PE (Santa Cruz Biotechnology, sc-3730) for 30 min at 4 °C. Binding and expression signals of the labeled cells were measured by flow cytometry (FACS Calibur, BD Biosciences). The error for the reported K_D corresponds to a $1 - \sigma$ (68%) confidence interval for the estimated binding affinity, and was calculated with the parametric bootstrap method using data from three technical replicates of each fit point⁴⁷.

Fluorescent protein library construction and screening. ddFP technology involves the reversible association of two dark (i.e., nonfluorescent) FP monomers to form a fluorescent heterodimer (**Supplementary Fig. 7**). The convention for naming the constituent FP monomers is ddFP-A copy and ddFP-B copy. The A copy monomer possesses the quenched chromophore, whereas the B copy monomer lacks a chromophore. We designate the noncovalent complex as ddFP (e.g., ddOFF) and the genetically fused tandem dimer as tdFP. tdRFP was cloned into the pBAD expression plasmid between XhoI and EcoRI restriction sites (where ddRFP-B was the 5' partner in the tandem gene fusion) and used as a template for protein-engineering experiments. First, M66T was introduced into the ddRFP-A copy by site-directed mutagenesis. A library was then created by randomly mutagenizing the ddRFP-A copy (M66T) gene using the following error-prone PCR conditions: 0.15 or 0.075 mM MnCl₂, 10 ng template, 200 μ M dATP, 200 μ M dGTP, 1 mM dTTP, 1 mM dCTP, and 4 mM MgCl₂. Forty cycles of PCR amplification were carried out for each reaction using Taq polymerase with an extension temperature of 68 °C. Mutant gene libraries were cloned 3' to the ddRFP-B copy, to generate a library of tandem ddFP variants, and transformed into DH10B electrocompetent *E. coli* for growth and screening within the μ SCALE platform.

To create a highly fluorescent ddOFF, μ SCALE was used to screen three successive *E. coli* libraries, with error-prone PCR performed on the isolated DNA between each round of screening. In each library cycle, an overnight outgrowth of cells in LB medium at 37 °C with shaking was performed after transformation to allow *E. coli* to recover. Cultures were then diluted to ~12,800 cells/ μ l in expression medium (LB supplemented with ampicillin (100 μ g/ml) and arabinose (0.1%)), loaded on a 10- μ m array, and incubated overnight at 37 °C in a sealed Petri dish with a moistened Kimwipe. After overnight growth, arrays containing 3×10^5 clones each were analyzed for fluorescence intensity with 525/45 nm excitation/565 nm long-pass emission filters. For μ SCALE round 1, 143 capillaries were extracted manually following visual inspection, and the contents were used to create the next library via error-prone PCR as described above. For μ SCALE round 2, 12 capillaries were extracted via single-cell sorting. The four clones with the highest fluorescence signal were used to create the third generation error-prone library with the same conditions above. For the third and final μ SCALE round, the top 10 capillaries were extracted via single-cell sorting. To evaluate protein characteristics after each library screen, extracted cells were cultured overnight at 37 °C with shaking in expression medium. Fluorescent proteins were extracted using B-Per (Pierce), and spectra were collected using a Biotek Synergy H4 plate reader.

ddOFF spectral properties. Prior to spectral analysis, all proteins were purified using metal chelating Ni-NTA chromatography and dialyzed into PBS. Emission spectra were recorded with a Biotek Synergy H4 plate reader. Absorbance measurements were made on a Varian Cary 50 UV/Vis Spectrophotometer using a 1-cm quartz microcell cuvette. The alkaline chromophore denaturation method was used to determine ϵ values. mCitrine ($\Phi = 0.76$) was used as the reference for quantum yield determination for ddOFF. pH titrations were performed by incubating purified proteins in buffers of desired pH and acquiring emission spectra with a 96-well Biotek Synergy H4 plate reader. A 1 μ M solution of fluorescent protein was prepared in PBS and diluted 1:10 with

a universal buffer of desired pH. This universal buffer solution was prepared by mixing equal volumes of 0.04 M H₃BO₃, 0.04 M CH₃COOH, and 0.04 M H₃PO₄. The pH was adjusted to the desired value by adding 1M NaOH to the prepared stock solution. The pK_a was determined by fitting the experimental data to the equation

$$F = A + B(1 + 10^{(pK_a - pH)n_H})^{-1}$$

where F is fluorescence, A and B are variables that define the baselines, and n_H is the Hill coefficient. Mean integrated emission peak intensities were normalized and plotted as a function of pH.

Preparation of alkaline phosphatase protein constructs. The *E. coli* WT alkaline phosphatase gene and R166S variant used for kinetic comparisons were cloned into the pCT yeast display plasmid between NheI and BamHI restriction sites. Plasmid DNA was subsequently transformed into *S. cerevisiae* EBY100 by electroporation for yeast surface-display studies. Soluble versions of WT alkaline phosphatase and the R166S mutant were also expressed and purified from *E. coli* SM547 λ DE3 cells as described previously⁴⁸. Yeast cells transformed with pCT display plasmid were grown in selective medium and induced as described above. The yeast induction medium was supplemented with 500 μ M ZnCl₂ and 500 μ M MgCl₂ to provide the required metal ions for catalysis. To quantify the level of yeast surface-displayed enzyme, antibody labeling of a C-terminal c-Myc tag fused to each enzyme was performed. Approximately 8×10^4 induced yeast cells were washed and resuspended in PBSA containing a 1:250 dilution of chicken anti-c-Myc antibody (Life Technologies, A21281) for 45 min at room temperature. After incubation, yeast were washed and resuspended in PBSA containing a 1:100 dilution of goat anti-chicken IgG Alexa Fluor 488 (Life Technologies, A11039) for 30 min at room temperature. c-Myc expression levels were then measured using a Guava easyCyte flow cytometer (Millipore).

Measurement of DDAO product pK_a . pH titrations were performed on 7-hydroxy-9H-(1,3-dichloro-9,9-dimethylacridin-2-one)(DDAO) to determine the pH dependence of product fluorescence (**Supplementary Fig. 10**). A 1 μ M stock solution of DDAO was prepared in Milli-Q water and titrated with 1M HCl and 1M NaOH to adjust the solution pH. Using a Biotek Synergy H4 plate reader, fluorescence measurements at an excitation/emission of 590/660 nm were performed on aliquots from each titration step. A fit of the data to the Henderson-Hasselbalch equation yielded a pK_a of 5.07 for the DDAO product. Thus, subsequent enzymatic assays were conducted at pH 8.0 to eliminate fluorescence quenching arising from protonation of the DDAO nitrogen.

Analysis of alkaline phosphate activity. Reaction conditions and calibration curves were established for the substrate 9H-(1,3-dichloro-9,9-dimethylacridin-2-one-7-yl) phosphate (DDAOP) and its product DDAO (**Supplementary Fig. 11**). Induced yeast cells were diluted to a final concentration of ~3,200 cells/ μ l, washed twice in AP Reaction Buffer (100 mM sodium MOPS, pH 8.0, 0.5 M NaCl, 1 mM MgCl₂, 100 μ M ZnSO₄, 0.2% BSA, 1 mM mannose)⁴⁸ and resuspended in AP Reaction Buffer containing variable concentrations of DDAOP and inorganic phosphate immediately prior to analysis. Reaction mixtures were then loaded onto either 96-well Microtiter plates or μ SCALE arrays for bulk and single-cell experiments, respectively. Bulk kinetic assays were conducted by measuring fluorescence time courses at an excitation/emission of 590 nm/660 nm on a Biotek Synergy H4 plate reader spectrophotometer, which allowed linear correlation of the fluorescent signal with DDAO concentration over a range from 0–3 μ M (**Supplementary Fig. 11a**). For kinetic measurements on the μ SCALE platform, a series of images were acquired at set intervals with 600/60 nm excitation/655 nm long-pass emission filters to generate single-cell reaction time courses for each capillary in the array. The μ SCALE filter set and camera detected a linear correlation of product formation with fluorescence over a comparable dynamic range to the plate reader (**Supplementary Fig. 11b**). In both cases, the linear regime of each kinetic time series was used to estimate reaction rates. An effective enzyme concentration of 50 pM (assuming ~ 10^4 displayed proteins per cell^{14,46}) was used for back-calculating rate constants from kinetic data.

Measurement of inorganic phosphate contamination in alkaline phosphatase reactions. Inorganic phosphate, which affects measurement of K_M and K_I values, was measured in buffer components and reagents using

fluorescence-labeled phosphate-binding protein (PBP) from *E. coli* (Life Technologies, PV4406) (**Supplementary Fig. 12** and **Supplementary Table 4**) as described previously⁴⁹. Briefly, 100 μ l of each sample was mixed with 100 μ l of 1 μ M PBP in Phosphate Sensor Detection Buffer (20 mM Tris-HCl, pH 7.6, 0.05% Triton X-100), and fluorescence of the mixture was immediately measured at an excitation/emission of 430 nm/450 nm on a Biotek Synergy H4 plate reader spectrophotometer. When quantifying phosphate content in the AP Reaction Buffer, the buffer was diluted 1:20 in water prior to incubation with the sensor in order to minimize interference from the concentrated MOPS and salt solutions on the PBP binding thermodynamics. PBP stocks were stored as 10 μ M aliquots in Phosphate Sensor Storage Buffer (10 mM Tris-HCl, pH 7.6, 50 mM NaCl) at -80°C and individually thawed prior to use. Fluorescence measurements were converted to phosphate concentrations using a standard curve (**Supplementary Fig. 12**), which was fit to a two-state binding isotherm of the form $F = F_{\min} + (F_{\max} - F_{\min}) \times (1 + K_D / [P_i])^{-1}$. The binding parameters were determined to be: $F_{\min} = 14,200 \pm 500$ RFU, $F_{\max} = 85,000 \pm 1,000$ RFU and $K_D = 0.94 \pm 0.05$ μ M. The presence of phosphate contamination in the AP Reaction Buffer (**Supplementary Table 4**) accounts for the discrepancy between kinetic parameters measured for WT alkaline phosphatase in this work and values previously reported in the literature³⁷.

Alkaline phosphatase library construction and screening. Error-prone PCR was used to randomly mutate the gene encoding WT alkaline phosphatase using the following reaction conditions: 0.1 mM MnCl_2 , 10 ng WT template, 200 μ M dATP, 200 μ M dGTP, 1 mM dTTP, 1 mM dCTP, and 4 mM MgCl_2 . PCR amplification was performed as described above. The mutated alkaline phosphatase insert DNA and *NheI*/*Bam*HI-digested pCT plasmid were electroporated in a 5:1 weight ratio into EBY100 yeast and assembled *in vivo* via homologous recombination⁴⁶. Library size was estimated to be $\sim 10^7$ by dilution plating.

μ SCALE was used to screen 3×10^5 variants from the alkaline phosphatase library to identify variants with improved activity under the assay conditions. Induced yeast cells were washed twice in AP Reaction Buffer, resuspended at $\sim 3,200$ cells/ μ l in AP Reaction Buffer containing 1 μ M DDAOP and 15 μ M inorganic phosphate and loaded into the array. After 30 min of reaction time, the fluorescent intensity of each capillary was analyzed as described in the alkaline phosphatase kinetic studies. The 15 capillaries with the highest fluorescence intensity were individually extracted and grown on agar plates containing selective yeast medium. Kinetic parameters of yeast-displayed alkaline phosphatase variants from the extracted cells were measured as described above. Plasmid DNA was recovered using a Zymoprep kit (Zymo

Research Corp.), transformed into DH10B electrocompetent *E. coli*, isolated using a plasmid miniprep kit (Thermo Scientific), and sequenced.

Assessment of metal binding in purified alkaline phosphatase constructs. Metal content in alkaline phosphatase variants purified from *E. coli* was determined by equilibrium dialysis followed by atomic emission spectroscopy, as described previously⁵⁰. Briefly, purified alkaline phosphatase mutants were incubated in AP Storage Buffer (10 mM Tris, pH 7.5, 50 mM NaCl, 100 μ M ZnCl_2 , 1 mM MgCl_2) for 4 d at room temperature. Enzyme aliquots at 10 μ M concentration were then dialyzed in storage buffer lacking zinc or magnesium using Amicon Ultra-15 centrifugal filter units (10 kDa MWCO). Dialyzed samples were diluted in 1 M sodium acetate, pH 5.5, and analyzed by atomic emission spectroscopy, using the following wavelengths for detection: 213.8 nm for Zn, 280.2 nm for Mg, 213.6 nm for P, and 180.7 nm for S. A standard curve containing control samples of dialysis buffer with known concentrations of sulfur, phosphorous, zinc, and magnesium (prepared as stock solutions in 2% nitric acid) was measured in the same manner and used to correct for background. Measured levels of phosphorous, zinc, and magnesium in each enzyme sample were normalized to protein levels (determined from sulfur content using a conversion of 12 sulfur atoms per alkaline phosphatase monomer) in order to determine the active site stoichiometry of each cofactor.

44. Otsu, N. A threshold selection method from gray-level histograms. *IEEE SMC* **9**, 62–66 (1979).
45. Van Deventer, J.A. & Wittrup, K.D. Yeast surface display for antibody isolation: library construction, library screening, and affinity maturation. *Methods Mol. Biol.* **1131**, 151–181 (2014).
46. Chao, G. *et al.* Isolating and engineering human antibodies using yeast surface display. *Nat. Protoc.* **1**, 755–768 (2006).
47. Benton, D. & Krishnamoorthy, K. Performance of the parametric bootstrap method in small sample interval estimates. *Adv. Appl. Stat.* **2**, 269–285 (2002).
48. Andrews, L.D., Zalatan, J.G. & Herschlag, D. Probing the origins of catalytic discrimination between phosphate and sulfate monoester hydrolysis: comparative analysis of alkaline phosphatase and protein tyrosine phosphatases. *Biochemistry* **53**, 6811–6819 (2014).
49. Brune, M., Hunter, J.L., Corrie, J.E.T. & Webb, M.R. Direct, real-time measurement of rapid inorganic phosphate release using a novel fluorescent probe and its application to actomyosin subfragment 1 ATPase. *Biochemistry* **33**, 8262–8271 (1994).
50. Zalatan, J.G., Fenn, T.D. & Herschlag, D. Comparative enzymology in the alkaline phosphatase superfamily to determine the catalytic role of an active-site metal ion. *J. Mol. Biol.* **384**, 1174–1189 (2008).