

CrossMark
click for updates

Cite this: DOI: 10.1039/c4an00781f

Electrochemical detection of glycan and protein epitopes of glycoproteins in serum

Alok K. Shah,^a Michelle M. Hill,^a Muhammad J. A. Shiddiky^{*b} and Matt Trau^{*b}

Aberrant protein glycosylation is associated with a range of pathological conditions including cancer and possesses diagnostic importance. Translation of glycoprotein biomarkers will be facilitated by the development of a rapid and sensitive analytical platform that simultaneously interrogates both the glycan and protein epitopes of glycoproteins in body fluids such as serum or saliva. To this end, we developed an electrochemical biosensor based on the immobilization of a lectin on the gold electrode surface to recognize/capture a target glycan epitope conjugated to glycoproteins, followed by detection of the protein epitope using a target protein-specific antibody. Electrochemical signals are generated by label-free voltammetric or impedimetric interrogation of a ferro/ferricyanide redox couple (e.g. $[\text{Fe}(\text{CN})_6]^{3-/4-}$) on the sensing surface, where the change in voltammetric current or interfacial electron transfer resistance was measured. The detection system was demonstrated using the model glycoprotein chicken ovalbumin with *Sambucus nigra* agglutinin type I (SNA lectin), and exhibits femtomolar sensitivity in the background of diluted human serum. The results obtained in this proof-of-concept study demonstrate the possibility of using electrochemical detection for developing cheap point-of-care diagnostics with high specificity and sensitivity for blood glycoprotein biomarkers.

Received 2nd May 2014
Accepted 29th August 2014

DOI: 10.1039/c4an00781f

www.rsc.org/analyst

Introduction

Differential protein glycosylation is a feature of diverse pathological conditions such as atherosclerosis, ulcerative colitis, rheumatoid arthritis, microbial infection, Alzheimer's disease and cancer.¹ Monitoring specific glycosylation changes for glycoprotein biomarker candidates such as fucosylated haptoglobin (pancreatic cancer) and sialylated prostate-specific antigen (prostate cancer) provides higher diagnostic power compared to changes in total glycoprotein levels.² In fact, measurement of fucosylated α -fetoprotein in blood (AFP-L3 test) has been approved by the FDA for early detection of hepatocellular carcinoma.^{3,4} Over the past few years, various glycoprotein enrichment platforms have been coupled with mass spectrometric and nuclear magnetic resonance spectroscopic techniques to uncover disease specific glycosylation changes for candidate glycoprotein biomarkers. These include single, serial and multi-lectin affinity chromatography,^{5,6} lectin magnetic bead array (LeMBA),^{7,8} boronic acid⁹ and hydrazide chemistry¹⁰ based extraction methods. While all these methods have excellent analytical performance in detecting candidate cancer glyco-biomarkers, they are poorly suited for routine clinical use due to high maintenance/running cost, requirement of

technical expertise, long assay time and complicated data analysis procedures. The development of detection methods that are user-friendly, robust, sensitive, quick and cheaper than currently available methods is warranted to fulfill important needs for developing future diagnostics.

Electrochemical (EC) detection methods offer elegant ways for interfacing bio-recognition and transduction events and represent substantial drivers to achieve rapid, cost-effective, sensitive, selective and accurate quantification of biomolecules.^{11–15} Among many EC methods, faradaic electrochemical impedance spectroscopy (F-EIS) is one of the most effective methods for the label-free detection of biomolecules and for probing the build-up of the biomaterial sensing film on the electrodes.^{16–18} In F-EIS, the binding of a target protein with its ligand (*i.e.*, a specific antibody) on the electrode surface can be detected, where the change in impedance of the electrode surface and its interface to the electrolyte solution containing a redox probe (*e.g.* $[\text{Fe}(\text{CN})_6]^{3-/4-}$) is measured in the form of its electron transfer resistance (R_{ct}).^{19–21} This interfacial electron transfer reaction of the redox process can also be measured *via* a voltammetric technique, where the presence or absence of the target proteins will alter the voltammetric current of the redox process at the sensing surface. More recently, the differential pulse voltammetric (DPV) interrogation of the interfacial electron transfer reaction of the $[\text{Fe}(\text{CN})_6]^{3-/4-}$ process generated upon protein binding has been used as an effective label-free tool for protein detection.^{22,23}

^aThe University of Queensland Diamantina Institute, The University of Queensland, Translational Research Institute, QLD 4102, Australia. E-mail: m.shiddiky@uq.edu.au; m.trau@uq.edu.au; Fax: +61-7-33463973; Tel: +61-7-33464178

^bAustralian Institute for Bioengineering and Nanotechnology (AIBN), Corner College and Cooper Roads (Bldg 75), The University of Queensland, QLD 4072, Australia

In the past few years, much attention has been focused on development of label-free electrochemical biosensors to monitor glycosylation changes (*e.g.* mannosylation, galactosylation, sialylation and fucosylation) in complex biological fluids^{24–27} using naturally occurring lectin as a glycan recognition element. These biosensors detect overall changes in the glycan profile without monitoring a specific glycoprotein to which glycan is attached. To increase the specificity and diagnostic value, analytical methods are required to detect not only the overall status of glycosylation in the sample but also aberrant glycosylation for a specific glycoprotein. To overcome this limitation, we investigated label-free electrochemical detection methods for the simultaneous interrogation of specific glycan on a target glycoprotein.

Here we describe the development of a simple method that can efficiently and reproducibly detect both glycan and protein epitopes of a glycoprotein in the background of a diluted serum sample at a femtomolar concentration. This method consists of a gold macrodisk electrode that is biochemically functionalized with *Sambucus nigra* agglutinin type I lectin (SNA lectin) [specific to recognize terminal sialic acid attached to galactose through α -(2-6) linkage].²⁸ The attachment of the target glycoprotein chicken egg albumin (ovalbumin) on the SNA lectin functionalized electrode and complexation with a polyclonal anti-ovalbumin antibody (Fig. 1 scheme) were followed by the F-EIS and DPV measurements. Each bimolecular layer on the sensor surface acts as a barrier for the interfacial electron transfer reaction of the $[\text{Fe}(\text{CN})_6]^{3-/4-}$ process, resulting in an increase in R_{ct} or a decrease in DPV current response. The presence of target glycoprotein ovalbumin and subsequent complexation with the polyclonal anti-ovalbumin antibody appear to further block the $[\text{Fe}(\text{CN})_6]^{3-/4-}$ process from accessing the electrode surface effectively. When we monitored R_{ct} or DPV current responses generated from the $[\text{Fe}(\text{CN})_6]^{3-/4-}$ process before and after ovalbumin capture, there was a clear correlation between the presence of the target ovalbumin and

changes in R_{ct} or DPV current response. To mimic clinical scenario, we spiked ovalbumin into serum and showed linear changes in current readout with a concentration range from 10 pg mL^{-1} to 500 pg mL^{-1} . To the best of our knowledge, this is the first electrochemical method that simultaneously interrogates specific glycans and the target protein on which the glycan is attached. Moreover, we believe that the simplicity of this technology could facilitate the translation of the current glycoprotein biomarker research.

Materials and methods

Materials

Biotin labeled *Sambucus nigra* agglutinin type I (SNA-I) lectin was purchased from Vector Laboratories (USA). Lyophilized chicken egg albumin (ovalbumin) (#A5503), polyclonal anti-chicken egg albumin antibody produced in rabbit (#C6534), potassium ferrocyanide, potassium ferricyanide, and KCl were purchased from Sigma (Australia). Biotinylated BSA was purchased from Thermo Scientific (USA) while multivalent streptavidin was purchased from Invitrogen (USA). Human blood sample was collected from a healthy volunteer with consent and ethics approved by the University of Queensland Human Ethics Committee.

Cleaning gold disk electrodes

Gold disk working electrodes (diameter = 3 mm) were purchased from CH Instruments (Austin, USA). The electrodes were reused after cleaning and regeneration. The electrodes were cleaned using piranha ($\text{H}_2\text{SO}_4 : \text{H}_2\text{O}_2 = 3 : 1$) on a sonicator water-bath for 30 s to 1 min (note: piranha is a highly toxic and hazardous chemical. It has to be handled in a fume-hood with adequate personal protective equipment. Universal safety guidelines should be followed for its disposal). The electrodes were then physically polished with 1 micron alumina and subsequently with a 0.05 micron alumina slurry. Prior to electrochemical cleaning, electrodes were sonicated in acetone for 20 min. Electrochemical cleaning was performed in 0.5 M H_2SO_4 until reproducible characteristic gold electrode profiles were achieved.

Construction of biosensor

Immediately before functionalization, thoroughly cleaned gold disk electrodes were dried under a flow of nitrogen gas. The electrodes were incubated with 500 $\mu\text{g mL}^{-1}$ biotin-BSA in 1 $\times$ PBS (137 mM NaCl, 2 mM KCl, 10 mM phosphate buffer, pH 7.4) on a thermo shaker set at 25 $^\circ\text{C}$, 300 rpm for 45 min. The electrodes were washed with 1 $\times$ PBS after each incubation step 3 times. The electrodes were then incubated with multivalent streptavidin (500 $\mu\text{g mL}^{-1}$ in 1 $\times$ PBS, at 25 $^\circ\text{C}$, 300 rpm for 45 min), followed by biotinylated SNA lectin (500 $\mu\text{g mL}^{-1}$ in 150 mM NaCl containing 0.1 mM Ca^{2+} , at 25 $^\circ\text{C}$, 300 rpm for 45 min) to form the bio-recognition layer of lectin on the electrode surface. Ovalbumin stock solution was made at a concentration of 2 mg mL^{-1} and stored at $-30\text{ }^\circ\text{C}$ in aliquots. The designated concentration of ovalbumin was freshly made (in 150 mM NaCl




Fig. 1 (A) Schematic illustration of the preparation of an SNA lectin immunosensing layer. (B) Schematic view of glycoprotein ovalbumin capture and label-free detection using an anti-ovalbumin antibody which results in increased charge transfer resistance (R_{ct}) and a corresponding decrease in DPV peak current.

solution containing 0.1 mM Ca^{2+}) prior to use and protein capture was performed for 45 min by gentle agitation using an Intelli-Mixer (PCOD Scientific). 1 in 1000 dilution of the polyclonal anti-ovalbumin antibody in $1\times$ PBS for 15 min was used for detection of captured chicken egg albumin. Designated concentrations of ovalbumin were spiked in diluted serum (1 in 1000 dilution of serum in 150 mM NaCl solution containing 0.1 mM Ca^{2+}).

After modification of the sensor surfaces (up to SNA lectin attachment), the actual detection steps include capturing the ovalbumin antigen followed by anti-ovalbumin detection. Both these steps are directly followed by the label-free impedance (or DPV) technique. Thus once the electrode is modified with the SNA lectin, the overall detection process takes approximately 2 h.

Electrochemical measurement

All electrochemical experiments were conducted at room temperature ($25 \pm 1^\circ\text{C}$) in a standard three-electrode electrochemical cell arrangement using an electrochemical workstation CHI 650D, CH Instruments (Austin, USA). The electrochemical cell consisted of a gold disk electrode sensor as a working electrode, a platinum wire counter electrode and a Ag/AgCl (3 M KCl) reference electrode. Electrochemical signals were measured in $1\times$ PBS buffer (pH 7.4) containing 2.5 mM $[\text{Fe}(\text{CN})_6]^{3-}/[\text{Fe}(\text{CN})_6]^{4-}$ (1 : 1) and 0.1 M KCl. DP signals were obtained with a potential step of 4 mV, a pulse amplitude of 50 mV, a pulse width of 200 ms, a sampling width of 16.7 ms and a pulse period of 500 ms. The F-EIS spectra were recorded in $1\times$ PBS buffer (pH 7.4) containing 2.5 mM $[\text{Fe}(\text{CN})_6]^{3-}/[\text{Fe}(\text{CN})_6]^{4-}$ (1 : 1) and 0.1 M KCl using an alternating current voltage of 5 mV, with the frequency range of 0.1 (or 1) Hz to 100 kHz. During the run, the bias DC current was applied below a frequency of 100 Hz. The faradaic current generated by the $\text{K}_3[\text{Fe}(\text{CN})_6]/\text{K}_4[\text{Fe}(\text{CN})_6]$ probe accounts for the presence of a biomolecule. The current changes corresponding to ovalbumin detection were calculated as follows:

$$\% \text{ Decrease of DPV peak current} = (I_{\text{before}} - I_{\text{after}}) \times 100 / I_{\text{before}} \quad (1)$$

where I_{before} = peak current at the ovalbumin capture step [e.g., current recorded at biotin-BSA/multivalent streptavidin/biotinylated SNA lectin/designated concentration of ovalbumin (either in buffer or spiked into 1 in 1000 serum)] and I_{after} = peak current at the detection step [e.g., current recorded at biotin-BSA/multivalent streptavidin/biotinylated SNA lectin/designated concentration of ovalbumin (either in buffer or spiked into 1 in 1000 serum)/1 in 1000 anti-ovalbumin polyclonal antibody].

For each electrode, the peak current was normalized with a DPV response obtained at the initial step when electrodes were clean.

Results and Discussion

Construction of biosensor

Fig. 1 describes an assembly scheme of the biosensor. As reported previously,¹⁸ biotin labeled BSA is used for coating the

gold electrode surface. Subsequently, multivalent streptavidin was used as a linker to immobilize the biotin labeled SNA lectin to form a bio-recognition layer on the electrode surface. The formation of the bio-recognition layer on the electrode surface affects analytical performance of the biosensor and in our experiments this formation is controlled by attachment of biotin-BSA, streptavidin and biotinylated SNA lectin. To achieve maximal analytical performance, we optimized binding conditions for all three biomolecules one after another by incubating electrodes with three different concentrations of biotin-BSA, multivalent streptavidin and biotinylated SNA lectin.

To determine the optimal biotin-BSA concentration at first, we incubated thoroughly cleaned gold electrodes with $100 \mu\text{g mL}^{-1}$, $500 \mu\text{g mL}^{-1}$ and $1000 \mu\text{g mL}^{-1}$ solution of biotin-BSA for 45 min. The electrodes were washed three times with $1\times$ PBS, followed by F-EIS measurements to determine the optimal biotin-BSA concentration. As compared to $100 \mu\text{g mL}^{-1}$, incubation of electrodes with $500 \mu\text{g mL}^{-1}$ concentration of biotin-BSA showed an increase in size of the semicircle on the Nyquist plot (Fig. 2A, i vs. ii) suggesting an increased electron transfer resistance (R_{ct}). There was no further increase in R_{ct} at $1000 \mu\text{g mL}^{-1}$ biotin-BSA concentration as compared to $500 \mu\text{g mL}^{-1}$




Fig. 2 Optimization of biosensor construction. Nyquist plots of a gold electrode modified with (A) biotin-BSA, (C) biotin-BSA/streptavidin, and (E) biotin-BSA/streptavidin/SNA lectin in $1\times$ PBS buffer containing 2.5 mM $\text{K}_3[\text{Fe}(\text{CN})_6]$, 2.5 mM $\text{K}_4[\text{Fe}(\text{CN})_6] \cdot 3\text{H}_2\text{O}$ and 0.1 M KCl. (B, D, and F) Corresponding DPV responses obtained at a gold electrode modified with (B) biotin-BSA, (D) biotin-BSA/streptavidin, and (F) biotin-BSA/streptavidin/SNA lectin, respectively. Concentrations of biotin-BSA in Fig. A & B, multivalent streptavidin in Fig. C & D (following incubation with $500 \mu\text{g mL}^{-1}$ of biotin-BSA), and SNA lectin in Fig. E & F (following incubation with $500 \mu\text{g mL}^{-1}$ biotin-BSA and $500 \mu\text{g mL}^{-1}$ streptavidin each) were (i) $100 \mu\text{g mL}^{-1}$, (ii) $500 \mu\text{g mL}^{-1}$, and (iii) $1000 \mu\text{g mL}^{-1}$.

(Fig. 2A, ii vs. iii). As $500\ \mu\text{g mL}^{-1}$ biotin-BSA showed best F-EIS *i.e.* maximum R_{ct} response, the same binding condition was used for subsequent experiments. Next, F-EIS for three different concentrations $100\ \mu\text{g mL}^{-1}$, $500\ \mu\text{g mL}^{-1}$ and $1000\ \mu\text{g mL}^{-1}$ of multivalent streptavidin under optimal biotin-BSA binding conditions was obtained. As shown in Fig. 2C, incubation of the electrode with a $500\ \mu\text{g mL}^{-1}$ concentration of multivalent streptavidin showed a maximum R_{ct} under optimal biotin-BSA binding conditions. Similarly, as shown in Figure 2E, $500\ \mu\text{g mL}^{-1}$ of biotinylated SNA lectin showed the best F-EIS response under optimal biotin-BSA and multivalent streptavidin binding conditions. To confirm F-EIS results we conducted parallel DPV measurements and as shown in Fig. 2B, D and F, there was a decrease in peak DPV current corresponding to an increase in R_{ct} suggesting reduction in actual electron transfer between the electrode/redox electrolyte double layer with increased resistance. For all three biomolecules, the maximum R_{ct} and minimal DPV peak current was observed at $500\ \mu\text{g mL}^{-1}$ concentration. Moreover, there was no further increase in R_{ct} at a higher concentration *i.e.* $1000\ \mu\text{g mL}^{-1}$. This suggests that the electrode surface is getting saturated when incubated with $500\ \mu\text{g mL}^{-1}$ concentration of biomolecules and there is no non-specific binding of the protein directly on the electrode surface at a higher concentration. As the protein binding to the electrode surface is showing saturation without any non-specific binding, it is anticipated that there will be minimal or no non-specific binding of biomolecule of interest directly to the electrode surface during further stages of the experiment.

Glycoprotein capture and detection

After optimizing conditions for SNA lectin immobilization on the electrode surface, next we captured glycoprotein ovalbumin in $150\ \text{mM NaCl}$ solution containing $0.1\ \text{mM Ca}^{2+}$. The formation of bio-recognition layers and capture of ovalbumin was followed by electrochemical detection. As shown in Fig. 3A, the biotin-BSA coated electrode gave rise to the smallest semicircle (Fig. 3Ai), followed by streptavidin (Fig. 3Aii), SNA lectin (Fig. 3Aiii) and ovalbumin (Fig. 3Aiv) which showed a gradual increase in semicircle indicating an increase in electron transfer resistance. In line with changes observed in impedance, we observed a corresponding decrease in peak DPV current (Fig. 3B). These results confirm the results shown in a previous section for successful formation of bio-recognition layers on the electrode surface. Next, it demonstrates successful capture of target glycoprotein ovalbumin on a SNA biosensor with electrochemical detection.

Lectins generally recognize glycan structures with low affinity but with high avidity mainly through hydrogen bonding, hydrophobic interactions and van der Waals forces.²⁹ In fact, glycan structure specific interactions of a glycoprotein with lectins are very well demonstrated in the recent past mainly using sialic acid binding lectin biosensors.^{25–27,30} Incubation of the SNA lectin biosensor with asialofetuin, a non-sialic acid expressing glycoform variant of glycoprotein fetuin, showed very minimal change in the baseline impedance.^{25–27,30} In contrast, a linear increase from baseline impedance was




Fig. 3 (A) Nyquist plots and (B) differential pulse voltammetric responses for monitoring (i) biotin-BSA (ii) biotin-BSA/multivalent streptavidin (iii) biotin-BSA/multivalent streptavidin/SNA lectin (iv) biotin-BSA/multivalent streptavidin/SNA lectin/1 ng mL^{-1} ovalbumin and (v) biotin-BSA/multivalent streptavidin/SNA lectin/1 ng mL^{-1} ovalbumin/anti-ovalbumin antibody in $1\times$ PBS buffer containing $2.5\ \text{mM K}_3[\text{Fe}(\text{CN})_6]$, $2.5\ \text{mM K}_4[\text{Fe}(\text{CN})_6]\cdot 3\text{H}_2\text{O}$ and $0.1\ \text{M KCl}$. (C) Differential pulse voltammetric responses to detect any non-specific binding of anti-ovalbumin antibody with SNA lectin (i) biotin-BSA/multivalent streptavidin/SNA lectin and (ii) biotin-BSA/multivalent streptavidin/SNA lectin/anti-ovalbumin antibody.

observed when the SNA lectin biosensor was incubated with an increasing concentration of fetuin expressing a correct glycan epitope.^{25–27,30} Apart from showing very high binding specificity towards the glycan epitope, these SNA lectin biosensors have demonstrated sensitivity for glycoprotein detection (*e.g.* fetuin) up to attomolar concentration.^{25–27,30} SNA along with other electrochemical lectin biosensors developed so far has been applied to glycan profiling of healthy and rheumatoid arthritis human serum samples,²⁷ monitor glycosylation changes between normal and cancerous pancreatic cell extracts,²⁵ dengue diagnosis³¹ and microorganism recognition,³² without detecting specific glycoproteins. Detection of both glycan and protein epitopes of a glycoprotein is required to provide the specificity to measure the glycosylation-specific biomarker, for example fucosylated alpha-fetoprotein. Therefore, following the capture of ovalbumin glycoprotein using a SNA lectin biosensor, we interrogated the ovalbumin protein backbone using a polyclonal anti-ovalbumin antibody. Upon capturing ovalbumin on a SNA biosensor, electrodes were incubated with 1 in 1000 dilution of anti-ovalbumin antibody, followed by F-EIS and DPV measurements. As compared to the ovalbumin capture step (Fig. 3Aiv) incubation of the electrode with an anti-ovalbumin antibody (Fig. 3Av) showed an increase in impedance, with a corresponding decrease in DPV peak current (Fig. 3B). These results indicate stepwise construction of the bio-recognition layer, capture of ovalbumin on the SNA biosensor and detection of ovalbumin using a polyclonal anti-ovalbumin antibody. As F-EIS and DPV techniques showed an excellent agreement between the results obtained for stepwise binding of biomolecules on the sensor surface, we presented only DPV responses for remaining all experiments.

As antibodies are glycoproteins themselves, they may directly interact with lectins which may lead to false positive results. Therefore, contribution of a false positive response due to direct interaction between the SNA lectin and a polyclonal anti-ovalbumin antibody was tested. The gold electrode modified with biotin-BSA/streptavidin/SNA lectin was directly incubated with an anti-ovalbumin antibody to allow any possible direct interaction between them. As shown in Fig. 3C we did not observe any change in the DPV response between the biotin-BSA/streptavidin/SNA lectin and the biotin-BSA/streptavidin/SNA lectin incubated with an anti-ovalbumin antibody. This finding indicates that there was no direct interaction between the ovalbumin antibody with the SNA lectin.

Thus, we demonstrate stepwise formation of a bio-recognition layer; capture of glycoprotein ovalbumin on the SNA biosensor surface followed by detection of ovalbumin using an anti-ovalbumin antibody and label-free electrochemical detection using F-EIS and DPV measurements. Moreover, we did not observe any direct interaction between the lectin and the antibody which excludes possibility for false positive results. To the best of our knowledge, this is the first electrochemical biosensor which interrogates both the glycan epitope of a glycoprotein using a lectin and protein epitope using an anti-protein antibody with label-free electrochemical monitoring.

Detection of ovalbumin in serum

To evaluate the biosensor performance under more physiological conditions, we conducted a series of experiments in which ovalbumin was spiked into diluted human serum collected from a healthy donor. To mimic biomarker detection conditions, ovalbumin at a final concentration of 1 ng mL⁻¹ was spiked into the 1 : 1000 serum sample. Similar dilution of the serum sample has been recommended to study the overall changes in glycan profiles using electrochemical detection.²⁷ Blood is very heterogeneous in terms of its protein composition, consisting of proteins with abundance varying by several orders of magnitude. Top six abundant plasma proteins account for 85% of the total plasma proteome.³³ Out of them, five are found to be glycosylated (transferrin, haptoglobin, α 1-antitrypsin, IgG, and IgA) and show varying degree of binding to the SNA lectin. This means that one would expect a significant amount of glycoprotein capture when the electrode immobilized with the SNA lectin is incubated with the serum/plasma sample. As can be seen in Fig. 4A, the DPV peak current was significantly reduced when the SNA lectin modified electrode was incubated with the serum sample (Fig. 4A, i vs. ii). However, there was no further decrease in the DPV response with anti-ovalbumin antibody incubation suggesting that the antibody did not bind non-specifically with any SNA lectin-bound human serum proteins (Fig. 4A, ii vs. iii).

When ovalbumin was spiked into serum, incubation of the electrode with the polyclonal anti-ovalbumin antibody (Fig. 4Bii) showed significant reduction in the peak DPV response as compared to the DPV peak current resulted at the ovalbumin capture step in serum (Fig. 4Bi). These data clearly demonstrate that our method is highly effective for specific detection of ovalbumin glycoprotein in the presence of large excess of several other SNA lectin binding proteins in serum.

To assess the dynamic range and lower limit of detection of our method, a dilution series of ovalbumin protein (1 pg mL⁻¹ to 10 ng mL⁻¹) was spiked in diluted serum. For each ovalbumin concentration, the electrochemical response was measured at the glycoprotein capture step and detection step using an anti-ovalbumin antibody. The percentage change in the peak current between these two measurements is attributed to the amount of ovalbumin present in the sample. As shown in Fig. 4C, there was a gradual increase in the percentage change in DPV peak current with increasing concentrations of ovalbumin spiked in serum ranging from 10 pg mL⁻¹ to 500 pg mL⁻¹.

Beyond 500 pg mL⁻¹, there was no further increase in DPV response suggesting saturation of the glycoprotein capture on the electrode surface. The lower limit of detection of 10 pg mL⁻¹ described here is comparable with previous electrochemical glycan profiling studies.^{25,27} This 10 pg mL⁻¹ lower limit of detection in a 1000 fold diluted serum sample will actually translate into 10 ng mL⁻¹ in the actual undiluted sample. Previously published lectin electrochemical biosensors showed femtomolar sensitivity and outperformed conventional lectin-ELISA and lectin-microarrays by several fold in terms of lower limit of detection.^{25,27} Unlike previous studies which only




Fig. 4 Differential pulse voltammetric responses for (A) detection of SNA binding proteins present in human serum (i) biotin-BSA/multivalent streptavidin/SNA lectin (ii) capturing SNA lectin binding glycoproteins present in human serum (1 in 1000 dilution) and (iii) detection of any non-specific recognition of SNA bound serum glycoprotein using an anti-ovalbumin antibody. (B) Detection of ovalbumin spiked in human serum (i) capturing 1 ng mL⁻¹ ovalbumin in human serum (1 in 1000 dilution) using SNA lectin and (ii) its detection using an anti-ovalbumin antibody. (C) Calibration plot for % change in the differential pulse voltammetric response for detection of a designated concentration of ovalbumin spiked in human serum (1 in 1000 dilution) using an anti-ovalbumin antibody. Each data point represents average $\pm$ SD (standard deviation) of 3 independent experiments.

interrogated glycan, here we interrogated both glycan and protein epitopes of a glycoprotein in the background of diluted serum with a detection limit of 10 pg mL⁻¹. The detection limit of our assay could be further improved *via* incorporation and optimization of the device geometry and experimental parameters.³⁴ Moreover, designing a multiplexed electrochemical biosensor^{25,34} would help to monitor the panel of clinically relevant glycoprotein biomarkers in parallel.

Conclusions

In this proof-of-concept study we demonstrated interrogation of glycan along with a protein epitope of a glycoprotein ovalbumin using SNA lectin and an anti-ovalbumin antibody respectively followed by label-free electrochemical detection in the background of diluted serum. Our lectin-antibody biosensor detects specific glycans and the glycoprotein to which the glycan is attached, hence provides increased specificity over existing lectin biosensors. Furthermore, with a detection limit of 10 pg mL⁻¹, the lectin-antibody electrochemical biosensor described here could be developed further to achieve point-of-care diagnosis for clinically relevant glycoprotein biomarkers.

Conflict of interest

The authors declare no competing financial interest.

Acknowledgements

AKS is supported by International Postgraduate Research Scholarship (IPRS) and UQ Centennial (UQ Cent) Scholarship. This work was supported by the ARC DECRA (DE120102503) to MJAS. We also acknowledge funding received by our laboratories from the National Breast Cancer Foundation of Australia (CG-12-07) to MT, NC-14-022 to MH and MT. These grants have significantly contributed to the environment to stimulate the research described here.

References

- 1 A. Varki and H. H. Freeze, in *Essentials of Glycobiology*, ed. A. Varki, R. D. Cummings, J. D. Esko, H. H. Freeze, P. Stanley, C. R. Bertozzi, G. W. Hart and M. E. Etzler, Cold Spring Harbor (NY), 2nd edn, 2009.
- 2 B. Adamczyk, T. Tharmalingam and P. M. Rudd, *Biochim. Biophys. Acta*, 2012, **1820**, 1347–1353.
- 3 K. Oda, A. Ido, T. Tamai, M. Matsushita, K. Kumagai, S. Mawatari, A. Saishoji, T. Kure, K. Ohno, E. Toyokura, D. Imanaka, A. Moriuchi, H. Uto, M. Oketani, T. Hashiguchi and H. Tsubouchi, *Oncol. Rep.*, 2011, **26**, 1227–1233.
- 4 C. Kagebayashi, I. Yamaguchi, A. Akinaga, H. Kitano, K. Yokoyama, M. Satomura, T. Kurosawa, M. Watanabe, T. Kawabata, W. Chang, C. Li, L. Bousse, H. G. Wada and S. Satomura, *Anal. Biochem.*, 2009, **388**, 306–311.
- 5 R. D. Cummings and S. Kornfeld, *J. Biol. Chem.*, 1982, **257**, 11235–11240.
- 6 Z. Yang and W. S. Hancock, *J. Chromatogr. A*, 2004, **1053**, 79–88.
- 7 D. Loo, A. Jones and M. M. Hill, *J. Proteome Res.*, 2010, **9**, 5496–5500.
- 8 E. Choi, D. Loo, J. W. Dennis, C. A. O'Leary and M. M. Hill, *Electrophoresis*, 2011, **32**, 3564–3575.
- 9 K. Sparbier, T. Wenzel and M. Kostrzewa, *J. Chromatogr. B: Anal. Technol. Biomed. Life Sci.*, 2006, **840**, 29–36.

- 10 H. Zhang, X. J. Li, D. B. Martin and R. Aebbersold, *Nat. Biotechnol.*, 2003, **21**, 660–666.
- 11 K. Chuah, L. M. Lai, I. Y. Goon, S. G. Parker, R. Amal and J. Justin Gooding, *Chem. Commun.*, 2012, **48**, 3503–3505.
- 12 Y. Du, C. Chen, B. Li, M. Zhou, E. Wang and S. Dong, *Biosens. Bioelectron.*, 2010, **25**, 1902–1907.
- 13 J. F. Rusling, *Anal. Chem.*, 2013, **85**, 5304–5310.
- 14 M. J. Shiddiky, M. A. Rahman and Y. B. Shim, *Anal. Chem.*, 2007, **79**, 6886–6890.
- 15 J. Wang, *Biosens. Bioelectron.*, 2006, **21**, 1887–1892.
- 16 E. Katz and I. Willner, *Electroanalysis*, 2003, **15**, 913–947.
- 17 M. Zhou and S. Dong, *Acc. Chem. Res.*, 2011, **44**, 1232–1243.
- 18 Y. S. Grewal, M. J. Shiddiky, S. A. Gray, K. M. Weigel, G. A. Cangelosi and M. Trau, *Chem. Commun.*, 2013, **49**, 1551–1553.
- 19 A. J. Bard and L. R. Faulkner, *Electrochemical Methods: Fundamentals and Applications*, John Wiley & Sons, Inc, 2nd edn, 2001.
- 20 F. Lisdat and D. Schafer, *Anal. Bioanal. Chem.*, 2008, **391**, 1555–1567.
- 21 J. S. Daniels and N. Pourmand, *Electroanalysis*, 2007, **19**, 1239–1257.
- 22 J. Das and S. O. Kelley, *Anal. Chem.*, 2011, **83**, 1167–1172.
- 23 Y. S. Grewal, M. J. Shiddiky, L. J. Spadafora, G. A. Cangelosi and M. Trau, *Biosens. Bioelectron.*, 2014, **55**, 417–422.
- 24 H. Yang, Z. Li, X. Wei, R. Huang, H. Qi, Q. Gao, C. Li and C. Zhang, *Talanta*, 2013, **111**, 62–68.
- 25 V. J. Nagaraj, S. Aithal, S. Eaton, M. Bothara, P. Wiktor and S. Prasad, *Nanomedicine*, 2010, **5**, 369–378.
- 26 J. T. La Belle, J. Q. Gerlach, S. Svarovsky and L. Joshi, *Anal. Chem.*, 2007, **79**, 6959–6964.
- 27 T. Bertok, L. Klukova, A. Sediva, P. Kasak, V. Semak, M. Micusik, M. Omastova, L. Chovanova, M. Vlcek, R. Imrich, A. Vikartovska and J. Tkac, *Anal. Chem.*, 2013, **85**, 7324–7332.
- 28 N. Shibuya, I. J. Goldstein, W. F. Broekaert, M. Nsimba-Lubaki, B. Peeters and W. J. Peumans, *J. Biol. Chem.*, 1987, **262**, 1596–1601.
- 29 N. Sharon, *Trends Biochem. Sci.*, 1993, **18**, 221–226.
- 30 T. Bertok, A. Sediva, J. Katrlík, P. Gemeiner, M. Mikula, M. Nosko and J. Tkac, *Talanta*, 2013, **108**, 11–18.
- 31 M. D. L. Oliveira, M. L. Nogueira, M. T. S. Correia, L. C. B. B. Coelho and C. A. S. Andrade, *Sens. Actuators, B*, 2011, **155**, 789–795.
- 32 M. Gamella, S. Campuzano, C. Parrado, A. J. Reviejo and J. M. Pingarron, *Talanta*, 2009, **78**, 1303–1309.
- 33 I. Musselman and D. W. Speicher, *Curr Protoc Protein Sci*, 2005, Ch. 24, Unit 24.1.
- 34 E. J. Wee, S. Rauf, K. M. Koo, M. J. Shiddiky and M. Trau, *Lab Chip*, 2013, **13**, 4385–4391.